



Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN081823\  
 Data File : BN026950.D  
 Acq On : 18 Aug 2023 08:48  
 Operator : MA/JU  
 Sample : DFTPP365  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DFTPP365

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN080323.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Fri Aug 18 13:11:26 2023



AutoFind: Scans 2543, 2544, 2545; Background Corrected with Scan 2535

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 29.0         | 147073     | PASS                |
| 68             | 69              | 0.00            | 2               | 1.8          | 3188       | PASS                |
| 69             | 198             | 0.00            | 100             | 34.3         | 174173     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.5          | 956        | PASS                |
| 127            | 198             | 10              | 80              | 45.1         | 228885     | PASS                |
| 197            | 198             | 0.00            | 2               | 0.4          | 1816       | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 507904     | PASS                |
| 199            | 198             | 5               | 9               | 6.6          | 33637      | PASS                |
| 275            | 198             | 10              | 60              | 26.2         | 132843     | PASS                |
| 365            | 198             | 1               | 100             | 3.3          | 16930      | PASS                |
| 441            | 198             | 0.01            | 100             | 12.5         | 63411      | PASS                |
| 442            | 198             | 50              | 100             | 78.6         | 399253     | PASS                |
| 443            | 442             | 15              | 24              | 19.1         | 76136      | PASS                |

| m/z    | Abundance |
|--------|-----------|
| 38.90  | 260.0     |
| 40.00  | 1767.0    |
| 41.00  | 433.0     |
| 42.00  | 232.0     |
| 43.00  | 433.0     |
| 44.00  | 3176.0    |
| 44.90  | 206.0     |
| 49.00  | 367.0     |
| 50.90  | 279.0     |
| 55.10  | 337.0     |
| 57.00  | 312.0     |
| 67.00  | 164.0     |
| 70.90  | 156.0     |
| 73.00  | 620.0     |
| 74.80  | 184.0     |
| 77.00  | 287.0     |
| 78.00  | 195.0     |
| 78.90  | 219.0     |
| 83.90  | 506.0     |
| 85.90  | 356.0     |
| 91.00  | 294.0     |
| 95.90  | 200.0     |
| 114.90 | 192.0     |
| 133.00 | 363.0     |
| 147.00 | 254.0     |
| 191.00 | 363.0     |
| 192.90 | 278.0     |
| 207.00 | 2928.0    |
| 208.00 | 705.0     |
| 209.00 | 295.0     |
| 281.00 | 677.0     |
| 282.00 | 231.0     |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP365

| m/z    | Abundance |
|--------|-----------|
| 36.00  | 309.0     |
| 37.00  | 869.0     |
| 38.00  | 2682.0    |
| 39.00  | 13807.0   |
| 39.90  | 1964.0    |
| 41.00  | 710.0     |
| 41.90  | 283.0     |
| 43.00  | 483.0     |
| 44.00  | 3050.0    |
| 45.00  | 510.0     |
| 46.90  | 209.0     |
| 47.90  | 172.0     |
| 49.00  | 1955.0    |
| 50.00  | 50424.0   |
| 51.00  | 190080.0  |
| 52.00  | 10782.0   |
| 53.00  | 385.0     |
| 55.00  | 1420.0    |
| 56.00  | 6315.0    |
| 57.00  | 13120.0   |
| 58.00  | 648.0     |
| 58.90  | 186.0     |
| 59.80  | 396.0     |
| 61.00  | 2975.0    |
| 62.00  | 3141.0    |
| 63.00  | 8103.0    |
| 64.10  | 1122.0    |
| 65.00  | 4331.0    |
| 66.00  | 276.0     |
| 66.90  | 508.0     |
| 68.00  | 3895.0    |
| 68.90  | 223744.0  |
| 70.00  | 1338.0    |
| 71.00  | 367.0     |
| 73.00  | 2519.0    |
| 74.00  | 25352.0   |
| 75.00  | 38480.0   |
| 76.00  | 13015.0   |
| 77.00  | 269952.0  |
| 78.00  | 19184.0   |
| 79.00  | 17712.0   |
| 80.00  | 12618.0   |
| 81.00  | 18184.0   |
| 82.00  | 4188.0    |
| 83.00  | 4371.0    |
| 83.90  | 1209.0    |
| 85.00  | 3308.0    |
| 86.00  | 5874.0    |
| 87.00  | 2299.0    |
| 88.00  | 979.0     |
| 88.90  | 502.0     |
| 91.00  | 4021.0    |
| 92.00  | 4913.0    |
| 93.00  | 31344.0   |
| 94.00  | 2025.0    |
| 95.00  | 523.0     |
| 96.00  | 1710.0    |
| 96.90  | 564.0     |
| 98.00  | 25048.0   |
| 99.00  | 22240.0   |
| 100.00 | 1749.0    |
| 100.90 | 10124.0   |
| 101.90 | 610.0     |
| 102.90 | 2936.0    |
| 104.00 | 6092.0    |
| 105.00 | 6234.0    |
| 106.00 | 2156.0    |
| 107.00 | 79024.0   |
| 108.00 | 12803.0   |
| 109.00 | 2016.0    |
| 110.00 | 145984.0  |
| 111.00 | 23008.0   |
| 112.00 | 2843.0    |
| 113.00 | 818.0     |
| 113.80 | 285.0     |
| 114.90 | 367.0     |
| 116.00 | 4648.0    |
| 117.00 | 75152.0   |
| 118.00 | 5876.0    |
| 119.00 | 652.0     |
| 120.00 | 994.0     |
| 121.00 | 296.0     |
| 122.00 | 5780.0    |
| 123.00 | 9018.0    |
| 124.00 | 4134.0    |
| 125.00 | 3522.0    |
| 126.10 | 1136.0    |
| 127.00 | 283008.0  |
| 128.00 | 21688.0   |
| 129.00 | 107272.0  |
| 130.00 | 9089.0    |
| 130.90 | 1756.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP365

|        |          |
|--------|----------|
| 131.90 | 832.0    |
| 132.80 | 539.0    |
| 134.00 | 3157.0   |
| 135.00 | 9021.0   |
| 136.00 | 4092.0   |
| 137.00 | 5764.0   |
| 137.90 | 1053.0   |
| 139.00 | 785.0    |
| 140.00 | 1295.0   |
| 140.90 | 14016.0  |
| 142.00 | 4803.0   |
| 143.00 | 3357.0   |
| 144.00 | 938.0    |
| 145.00 | 673.0    |
| 146.00 | 2389.0   |
| 147.00 | 7617.0   |
| 147.90 | 19080.0  |
| 149.00 | 4081.0   |
| 149.90 | 1205.0   |
| 151.10 | 1648.0   |
| 151.90 | 1008.0   |
| 152.90 | 4162.0   |
| 154.00 | 3466.0   |
| 155.00 | 8228.0   |
| 156.00 | 11745.0  |
| 157.00 | 2249.0   |
| 158.00 | 2536.0   |
| 158.90 | 2016.0   |
| 160.00 | 4327.0   |
| 161.00 | 6718.0   |
| 162.00 | 1969.0   |
| 162.90 | 670.0    |
| 164.00 | 615.0    |
| 164.90 | 5544.0   |
| 166.00 | 4179.0   |
| 167.00 | 31064.0  |
| 168.00 | 23528.0  |
| 169.00 | 3013.0   |
| 169.90 | 1128.0   |
| 171.00 | 1041.0   |
| 171.90 | 2763.0   |
| 173.00 | 3636.0   |
| 173.90 | 5751.0   |
| 175.00 | 11307.0  |
| 176.00 | 3329.0   |
| 176.90 | 4774.0   |
| 177.90 | 1445.0   |
| 179.00 | 20936.0  |
| 180.00 | 16736.0  |
| 181.00 | 6992.0   |
| 182.00 | 1186.0   |
| 183.00 | 885.0    |
| 183.90 | 1929.0   |
| 185.00 | 10050.0  |
| 186.00 | 78568.0  |
| 187.00 | 22368.0  |
| 188.00 | 2565.0   |
| 189.00 | 4691.0   |
| 189.90 | 689.0    |
| 191.00 | 2793.0   |
| 192.00 | 6932.0   |
| 193.00 | 9046.0   |
| 194.00 | 1939.0   |
| 195.10 | 862.0    |
| 196.00 | 16568.0  |
| 196.90 | 2652.0   |
| 198.00 | 588416.0 |
| 199.00 | 39544.0  |
| 200.00 | 3149.0   |
| 201.30 | 2474.0   |
| 203.00 | 4435.0   |
| 204.00 | 21072.0  |
| 205.00 | 34272.0  |
| 206.00 | 155712.0 |
| 207.00 | 24408.0  |
| 208.00 | 5618.0   |
| 209.00 | 1937.0   |
| 210.00 | 2007.0   |
| 211.00 | 5991.0   |
| 212.90 | 512.0    |
| 213.80 | 229.0    |
| 215.00 | 1625.0   |
| 215.90 | 3423.0   |
| 217.00 | 40496.0  |
| 218.00 | 5113.0   |
| 219.00 | 388.0    |
| 221.00 | 24024.0  |
| 221.80 | 5432.0   |
| 223.00 | 9372.0   |
| 224.00 | 85888.0  |
| 225.00 | 21184.0  |
| 226.00 | 2755.0   |
| 227.00 | 35672.0  |
| 228.00 | 4957.0   |
| 229.00 | 7373.0   |
| 229.90 | 1010.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP365

|        |          |
|--------|----------|
| 231.00 | 3284.0   |
| 232.00 | 512.0    |
| 232.90 | 756.0    |
| 234.00 | 2214.0   |
| 235.00 | 2691.0   |
| 235.90 | 1803.0   |
| 237.00 | 3354.0   |
| 238.00 | 477.0    |
| 238.90 | 1479.0   |
| 239.90 | 1251.0   |
| 240.90 | 1959.0   |
| 242.00 | 4525.0   |
| 243.00 | 4928.0   |
| 244.00 | 70192.0  |
| 245.10 | 9074.0   |
| 246.00 | 12378.0  |
| 247.00 | 2795.0   |
| 248.00 | 685.0    |
| 248.90 | 2370.0   |
| 249.90 | 376.0    |
| 250.90 | 660.0    |
| 251.90 | 473.0    |
| 253.00 | 1245.0   |
| 254.10 | 2074.0   |
| 255.00 | 331136.0 |
| 256.00 | 50336.0  |
| 257.00 | 4067.0   |
| 258.00 | 19416.0  |
| 259.00 | 3284.0   |
| 259.90 | 508.0    |
| 260.90 | 384.0    |
| 264.10 | 597.0    |
| 265.00 | 7213.0   |
| 266.00 | 1104.0   |
| 267.00 | 234.0    |
| 269.80 | 322.0    |
| 271.00 | 901.0    |
| 272.00 | 783.0    |
| 273.00 | 9782.0   |
| 274.00 | 24816.0  |
| 275.00 | 139328.0 |
| 276.00 | 18832.0  |
| 277.00 | 11338.0  |
| 278.00 | 2186.0   |
| 278.90 | 591.0    |
| 281.00 | 833.0    |
| 282.00 | 549.0    |
| 283.00 | 1339.0   |
| 284.00 | 980.0    |
| 285.00 | 2302.0   |
| 286.10 | 437.0    |
| 288.90 | 461.0    |
| 290.00 | 474.0    |
| 290.90 | 415.0    |
| 291.90 | 554.0    |
| 293.00 | 3056.0   |
| 294.00 | 869.0    |
| 294.90 | 763.0    |
| 296.00 | 40304.0  |
| 297.00 | 6121.0   |
| 297.90 | 380.0    |
| 300.90 | 549.0    |
| 302.00 | 619.0    |
| 303.00 | 4827.0   |
| 304.00 | 1259.0   |
| 305.00 | 175.0    |
| 308.00 | 544.0    |
| 308.90 | 246.0    |
| 309.90 | 619.0    |
| 310.90 | 189.0    |
| 313.00 | 388.0    |
| 314.00 | 1773.0   |
| 315.00 | 4386.0   |
| 316.00 | 2707.0   |
| 317.00 | 455.0    |
| 319.90 | 162.0    |
| 320.90 | 1147.0   |
| 322.00 | 689.0    |
| 323.00 | 13331.0  |
| 324.00 | 2420.0   |
| 325.00 | 279.0    |
| 325.80 | 302.0    |
| 326.90 | 2643.0   |
| 327.90 | 1427.0   |
| 329.00 | 238.0    |
| 332.00 | 882.0    |
| 333.00 | 1151.0   |
| 334.00 | 8817.0   |
| 335.00 | 2364.0   |
| 335.90 | 295.0    |
| 338.90 | 228.0    |
| 339.90 | 280.0    |
| 340.90 | 1507.0   |
| 342.00 | 369.0    |
| 345.90 | 2745.0   |
| 346.90 | 535.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP365

|        |          |
|--------|----------|
| 350.90 | 254.0    |
| 352.00 | 3392.0   |
| 353.00 | 2693.0   |
| 354.00 | 3581.0   |
| 355.00 | 1002.0   |
| 358.90 | 331.0    |
| 364.90 | 17096.0  |
| 366.00 | 2919.0   |
| 366.90 | 234.0    |
| 369.90 | 434.0    |
| 371.00 | 840.0    |
| 372.10 | 6321.0   |
| 373.00 | 1684.0   |
| 373.90 | 245.0    |
| 382.90 | 1771.0   |
| 384.00 | 452.0    |
| 384.90 | 264.0    |
| 390.00 | 751.0    |
| 390.90 | 638.0    |
| 392.00 | 447.0    |
| 401.00 | 364.0    |
| 401.90 | 2799.0   |
| 403.00 | 4005.0   |
| 403.90 | 1153.0   |
| 405.00 | 307.0    |
| 415.00 | 164.0    |
| 421.00 | 3209.0   |
| 422.00 | 3577.0   |
| 423.00 | 22288.0  |
| 424.00 | 6431.0   |
| 425.00 | 896.0    |
| 438.80 | 177.0    |
| 439.80 | 217.0    |
| 441.00 | 58032.0  |
| 442.00 | 366336.0 |
| 443.00 | 61720.0  |
| 444.00 | 7129.0   |
| 444.90 | 513.0    |

**Instrument :**  
BNA\_N  
**ClientSampleId :**  
DFTPP365

| m/z    | Abundance |
|--------|-----------|
| 35.00  | 231.0     |
| 37.00  | 1133.0    |
| 38.00  | 2290.0    |
| 39.00  | 11758.0   |
| 40.00  | 1694.0    |
| 41.00  | 791.0     |
| 42.00  | 332.0     |
| 43.00  | 439.0     |
| 44.00  | 3309.0    |
| 44.90  | 395.0     |
| 46.80  | 173.0     |
| 47.80  | 252.0     |
| 49.00  | 1659.0    |
| 50.00  | 45736.0   |
| 51.00  | 164800.0  |
| 52.00  | 8113.0    |
| 53.00  | 373.0     |
| 54.00  | 204.0     |
| 54.90  | 1207.0    |
| 56.00  | 4871.0    |
| 57.00  | 11186.0   |
| 58.00  | 525.0     |
| 58.90  | 227.0     |
| 59.90  | 287.0     |
| 61.00  | 2293.0    |
| 62.00  | 2479.0    |
| 63.00  | 7212.0    |
| 64.00  | 1107.0    |
| 65.00  | 3677.0    |
| 66.00  | 327.0     |
| 67.00  | 496.0     |
| 68.00  | 3799.0    |
| 69.00  | 195776.0  |
| 69.90  | 967.0     |
| 71.00  | 282.0     |
| 73.00  | 2175.0    |
| 74.00  | 21080.0   |
| 75.00  | 32752.0   |
| 76.00  | 11609.0   |
| 77.00  | 240832.0  |
| 78.00  | 16608.0   |
| 79.00  | 15627.0   |
| 80.00  | 11643.0   |
| 81.00  | 16520.0   |
| 82.00  | 3765.0    |
| 83.00  | 3836.0    |
| 83.90  | 996.0     |
| 85.00  | 2756.0    |
| 86.00  | 5051.0    |
| 87.00  | 2354.0    |
| 87.90  | 1066.0    |
| 88.90  | 467.0     |
| 90.90  | 3448.0    |
| 92.00  | 4500.0    |
| 93.00  | 29312.0   |
| 94.00  | 1828.0    |
| 95.00  | 617.0     |
| 95.90  | 1471.0    |
| 97.00  | 749.0     |
| 98.00  | 22408.0   |
| 99.00  | 19192.0   |
| 99.90  | 1520.0    |
| 101.00 | 9295.0    |
| 101.90 | 639.0     |
| 103.00 | 3203.0    |
| 104.00 | 6325.0    |
| 105.00 | 5614.0    |
| 106.00 | 1725.0    |
| 107.00 | 73488.0   |
| 108.00 | 11204.0   |
| 108.90 | 1945.0    |
| 110.00 | 132288.0  |
| 111.00 | 20784.0   |
| 111.90 | 2450.0    |
| 113.00 | 925.0     |
| 114.00 | 215.0     |
| 115.00 | 558.0     |
| 116.00 | 3565.0    |
| 117.00 | 70160.0   |
| 118.00 | 5680.0    |
| 119.00 | 756.0     |
| 119.90 | 950.0     |
| 121.00 | 527.0     |
| 121.90 | 5193.0    |
| 123.00 | 7135.0    |
| 124.00 | 3915.0    |
| 125.00 | 3288.0    |
| 126.00 | 979.0     |
| 127.00 | 261632.0  |
| 128.00 | 19560.0   |
| 129.00 | 102336.0  |
| 129.90 | 8539.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP365

|        |          |
|--------|----------|
| 131.00 | 2007.0   |
| 131.90 | 940.0    |
| 133.00 | 646.0    |
| 133.90 | 2820.0   |
| 135.00 | 8187.0   |
| 136.00 | 3251.0   |
| 137.00 | 5696.0   |
| 138.00 | 1155.0   |
| 138.90 | 657.0    |
| 139.90 | 1138.0   |
| 141.00 | 13498.0  |
| 142.00 | 4558.0   |
| 143.00 | 2937.0   |
| 144.00 | 844.0    |
| 144.90 | 870.0    |
| 145.90 | 2488.0   |
| 147.00 | 7814.0   |
| 147.90 | 17832.0  |
| 149.00 | 3935.0   |
| 150.10 | 1072.0   |
| 151.00 | 1795.0   |
| 151.80 | 933.0    |
| 152.90 | 3915.0   |
| 154.00 | 3276.0   |
| 155.00 | 7882.0   |
| 156.00 | 11100.0  |
| 156.90 | 2045.0   |
| 158.00 | 2425.0   |
| 159.00 | 2019.0   |
| 160.00 | 3975.0   |
| 161.00 | 6378.0   |
| 161.90 | 1752.0   |
| 162.90 | 526.0    |
| 164.00 | 826.0    |
| 165.00 | 5060.0   |
| 166.00 | 4359.0   |
| 167.00 | 29536.0  |
| 168.00 | 21640.0  |
| 169.00 | 3132.0   |
| 170.00 | 1022.0   |
| 170.90 | 1319.0   |
| 171.90 | 2401.0   |
| 173.00 | 3286.0   |
| 174.00 | 6086.0   |
| 175.00 | 10959.0  |
| 176.00 | 3302.0   |
| 176.90 | 5008.0   |
| 178.00 | 1699.0   |
| 178.90 | 21256.0  |
| 180.00 | 15107.0  |
| 181.00 | 6949.0   |
| 182.00 | 1470.0   |
| 182.80 | 516.0    |
| 183.90 | 1898.0   |
| 185.00 | 11214.0  |
| 186.00 | 79448.0  |
| 187.00 | 22840.0  |
| 188.00 | 2162.0   |
| 189.00 | 5165.0   |
| 190.00 | 1034.0   |
| 191.00 | 2613.0   |
| 192.00 | 6988.0   |
| 193.00 | 7876.0   |
| 194.00 | 1738.0   |
| 195.00 | 880.0    |
| 196.00 | 16528.0  |
| 196.90 | 2797.0   |
| 197.90 | 596224.0 |
| 198.90 | 40096.0  |
| 200.00 | 3263.0   |
| 201.50 | 2678.0   |
| 203.00 | 3975.0   |
| 204.00 | 20336.0  |
| 205.00 | 37184.0  |
| 206.00 | 154560.0 |
| 207.00 | 22336.0  |
| 208.00 | 5036.0   |
| 209.00 | 2107.0   |
| 210.10 | 2863.0   |
| 211.00 | 6392.0   |
| 212.90 | 463.0    |
| 214.00 | 157.0    |
| 215.00 | 1614.0   |
| 216.00 | 3258.0   |
| 216.90 | 39784.0  |
| 218.00 | 5020.0   |
| 219.00 | 634.0    |
| 221.00 | 24520.0  |
| 221.80 | 6044.0   |
| 223.00 | 9320.0   |
| 224.00 | 87976.0  |
| 225.00 | 22272.0  |
| 226.00 | 2496.0   |
| 227.00 | 33904.0  |
| 228.00 | 5098.0   |
| 229.00 | 7078.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP365



|        |          |
|--------|----------|
| 230.00 | 1086.0   |
| 231.00 | 3280.0   |
| 231.90 | 490.0    |
| 233.00 | 559.0    |
| 234.00 | 2029.0   |
| 234.90 | 2424.0   |
| 236.00 | 1864.0   |
| 237.00 | 3214.0   |
| 238.00 | 721.0    |
| 238.90 | 1418.0   |
| 239.90 | 1002.0   |
| 241.00 | 1951.0   |
| 242.00 | 4302.0   |
| 243.00 | 5388.0   |
| 244.00 | 71864.0  |
| 245.00 | 9317.0   |
| 246.00 | 13297.0  |
| 246.90 | 2828.0   |
| 247.90 | 694.0    |
| 248.90 | 2385.0   |
| 250.00 | 410.0    |
| 250.90 | 546.0    |
| 251.90 | 593.0    |
| 253.00 | 1452.0   |
| 254.00 | 1917.0   |
| 255.00 | 345920.0 |
| 256.00 | 49832.0  |
| 257.00 | 3954.0   |
| 258.00 | 20360.0  |
| 259.00 | 3331.0   |
| 259.90 | 593.0    |
| 260.90 | 588.0    |
| 263.00 | 203.0    |
| 264.00 | 534.0    |
| 265.00 | 8658.0   |
| 265.90 | 1104.0   |
| 266.90 | 337.0    |
| 267.80 | 222.0    |
| 269.80 | 348.0    |
| 271.00 | 1000.0   |
| 272.00 | 918.0    |
| 273.00 | 11612.0  |
| 274.00 | 27944.0  |
| 275.00 | 164992.0 |
| 276.00 | 21664.0  |
| 277.00 | 13749.0  |
| 278.00 | 2372.0   |
| 278.90 | 508.0    |
| 281.10 | 769.0    |
| 281.90 | 594.0    |
| 283.00 | 1632.0   |
| 284.00 | 982.0    |
| 285.00 | 2332.0   |
| 285.90 | 577.0    |
| 287.80 | 186.0    |
| 288.90 | 479.0    |
| 290.00 | 526.0    |
| 290.90 | 315.0    |
| 292.00 | 845.0    |
| 292.90 | 2889.0   |
| 293.90 | 689.0    |
| 295.00 | 952.0    |
| 296.00 | 42856.0  |
| 297.00 | 5295.0   |
| 297.90 | 418.0    |
| 300.90 | 634.0    |
| 302.00 | 577.0    |
| 303.00 | 4933.0   |
| 304.00 | 1367.0   |
| 305.10 | 203.0    |
| 308.00 | 632.0    |
| 308.90 | 474.0    |
| 309.90 | 663.0    |
| 312.90 | 346.0    |
| 314.00 | 2184.0   |
| 314.90 | 4835.0   |
| 316.10 | 2850.0   |
| 316.80 | 476.0    |
| 319.80 | 227.0    |
| 321.00 | 1383.0   |
| 321.90 | 867.0    |
| 323.00 | 15292.0  |
| 324.00 | 2529.0   |
| 325.00 | 311.0    |
| 325.90 | 290.0    |
| 326.90 | 2947.0   |
| 328.00 | 1740.0   |
| 329.00 | 230.0    |
| 332.00 | 932.0    |
| 333.00 | 1389.0   |
| 334.00 | 9982.0   |
| 335.00 | 2922.0   |
| 336.10 | 316.0    |
| 338.90 | 254.0    |
| 340.00 | 202.0    |
| 341.00 | 1840.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP365

|        |          |
|--------|----------|
| 342.00 | 565.0    |
| 345.90 | 3855.0   |
| 346.80 | 684.0    |
| 349.80 | 165.0    |
| 351.00 | 320.0    |
| 352.00 | 4908.0   |
| 353.00 | 3243.0   |
| 354.00 | 5056.0   |
| 355.00 | 1121.0   |
| 358.90 | 380.0    |
| 364.00 | 180.0    |
| 365.00 | 20000.0  |
| 365.90 | 2856.0   |
| 366.90 | 363.0    |
| 370.00 | 461.0    |
| 371.00 | 1113.0   |
| 372.00 | 8041.0   |
| 373.00 | 2033.0   |
| 377.00 | 275.0    |
| 383.00 | 1969.0   |
| 384.00 | 667.0    |
| 384.90 | 230.0    |
| 390.10 | 984.0    |
| 391.00 | 739.0    |
| 391.90 | 617.0    |
| 401.00 | 484.0    |
| 402.00 | 2687.0   |
| 403.00 | 5116.0   |
| 404.00 | 2009.0   |
| 404.90 | 327.0    |
| 410.00 | 210.0    |
| 414.80 | 319.0    |
| 421.00 | 4208.0   |
| 422.00 | 4368.0   |
| 423.00 | 27904.0  |
| 424.00 | 8144.0   |
| 425.00 | 783.0    |
| 439.10 | 171.0    |
| 441.00 | 78240.0  |
| 442.00 | 498624.0 |
| 443.00 | 94536.0  |
| 444.00 | 9113.0   |
| 445.00 | 590.0    |

**Instrument :**  
BNA\_N  
**ClientSampleId :**  
DFTPP365

| m/z    | Abundance |
|--------|-----------|
| 36.00  | 237.0     |
| 37.00  | 538.0     |
| 38.00  | 1021.0    |
| 39.00  | 6856.0    |
| 39.90  | 1787.0    |
| 41.00  | 733.0     |
| 43.00  | 426.0     |
| 44.00  | 3090.0    |
| 44.90  | 466.0     |
| 49.00  | 988.0     |
| 50.00  | 22256.0   |
| 51.00  | 87176.0   |
| 52.00  | 4512.0    |
| 53.00  | 358.0     |
| 54.90  | 913.0     |
| 56.00  | 2911.0    |
| 57.00  | 6233.0    |
| 57.90  | 247.0     |
| 59.80  | 288.0     |
| 61.00  | 1319.0    |
| 62.00  | 1351.0    |
| 63.00  | 3723.0    |
| 63.90  | 497.0     |
| 65.00  | 1703.0    |
| 65.90  | 265.0     |
| 66.90  | 299.0     |
| 68.00  | 1869.0    |
| 68.90  | 103000.0  |
| 70.00  | 562.0     |
| 71.10  | 255.0     |
| 73.00  | 1668.0    |
| 74.00  | 11019.0   |
| 75.00  | 16944.0   |
| 76.00  | 6473.0    |
| 77.00  | 122608.0  |
| 78.00  | 8962.0    |
| 79.00  | 7916.0    |
| 79.90  | 6020.0    |
| 81.00  | 8338.0    |
| 81.90  | 2243.0    |
| 83.00  | 1997.0    |
| 83.90  | 675.0     |
| 85.00  | 1575.0    |
| 86.00  | 3052.0    |
| 87.00  | 1154.0    |
| 87.90  | 435.0     |
| 88.90  | 322.0     |
| 91.00  | 1796.0    |
| 91.90  | 2524.0    |
| 93.00  | 14909.0   |
| 94.00  | 1118.0    |
| 95.00  | 280.0     |
| 96.10  | 1072.0    |
| 97.00  | 435.0     |
| 98.00  | 12213.0   |
| 99.00  | 10943.0   |
| 99.90  | 894.0     |
| 101.00 | 5210.0    |
| 101.90 | 257.0     |
| 103.00 | 1915.0    |
| 104.00 | 2930.0    |
| 105.00 | 3112.0    |
| 106.00 | 923.0     |
| 107.00 | 38912.0   |
| 108.00 | 5753.0    |
| 108.90 | 1268.0    |
| 110.00 | 70416.0   |
| 110.90 | 11031.0   |
| 111.90 | 1251.0    |
| 113.10 | 519.0     |
| 115.00 | 352.0     |
| 116.00 | 1899.0    |
| 117.00 | 40128.0   |
| 117.90 | 2838.0    |
| 119.00 | 439.0     |
| 119.90 | 575.0     |
| 120.80 | 236.0     |
| 122.00 | 3061.0    |
| 123.00 | 4027.0    |
| 123.90 | 1988.0    |
| 125.00 | 1879.0    |
| 126.10 | 438.0     |
| 127.00 | 142016.0  |
| 128.00 | 10345.0   |
| 129.00 | 54536.0   |
| 130.00 | 4629.0    |
| 131.00 | 947.0     |
| 132.00 | 435.0     |
| 133.00 | 571.0     |
| 133.90 | 1820.0    |
| 135.00 | 4587.0    |
| 135.90 | 1769.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP365

|        |          |
|--------|----------|
| 137.00 | 3243.0   |
| 138.00 | 622.0    |
| 139.10 | 239.0    |
| 140.00 | 800.0    |
| 140.90 | 7696.0   |
| 142.00 | 2744.0   |
| 143.00 | 1717.0   |
| 144.00 | 490.0    |
| 145.00 | 514.0    |
| 145.90 | 1447.0   |
| 146.90 | 3697.0   |
| 147.90 | 9932.0   |
| 149.00 | 2413.0   |
| 150.00 | 863.0    |
| 150.90 | 1024.0   |
| 151.90 | 567.0    |
| 153.00 | 2395.0   |
| 153.90 | 1717.0   |
| 155.00 | 4149.0   |
| 156.00 | 5906.0   |
| 156.90 | 1063.0   |
| 157.90 | 1606.0   |
| 158.90 | 1174.0   |
| 160.00 | 2151.0   |
| 161.00 | 3796.0   |
| 161.90 | 1137.0   |
| 162.90 | 446.0    |
| 163.90 | 395.0    |
| 164.90 | 3122.0   |
| 166.00 | 2256.0   |
| 167.00 | 17016.0  |
| 168.00 | 12648.0  |
| 169.00 | 1938.0   |
| 170.00 | 457.0    |
| 171.00 | 695.0    |
| 171.90 | 1528.0   |
| 173.00 | 1913.0   |
| 173.90 | 3818.0   |
| 175.00 | 6623.0   |
| 176.00 | 1929.0   |
| 176.90 | 2426.0   |
| 177.90 | 1078.0   |
| 178.90 | 11653.0  |
| 180.00 | 8578.0   |
| 181.00 | 3848.0   |
| 182.00 | 611.0    |
| 182.80 | 282.0    |
| 183.90 | 1123.0   |
| 185.00 | 6096.0   |
| 186.00 | 44528.0  |
| 187.00 | 12688.0  |
| 188.00 | 1367.0   |
| 189.00 | 2981.0   |
| 189.90 | 331.0    |
| 190.90 | 1729.0   |
| 192.00 | 3965.0   |
| 193.00 | 4587.0   |
| 193.90 | 872.0    |
| 195.00 | 634.0    |
| 196.00 | 8862.0   |
| 197.90 | 339072.0 |
| 198.90 | 21272.0  |
| 200.00 | 1901.0   |
| 201.50 | 1657.0   |
| 202.90 | 2221.0   |
| 204.00 | 12272.0  |
| 205.00 | 20384.0  |
| 206.00 | 92192.0  |
| 207.00 | 14205.0  |
| 208.00 | 3438.0   |
| 209.00 | 1305.0   |
| 210.00 | 1523.0   |
| 211.00 | 3451.0   |
| 213.00 | 280.0    |
| 214.90 | 948.0    |
| 216.00 | 2131.0   |
| 216.90 | 23496.0  |
| 218.00 | 3033.0   |
| 219.00 | 400.0    |
| 220.00 | 433.0    |
| 221.00 | 13812.0  |
| 221.90 | 3160.0   |
| 223.00 | 4889.0   |
| 224.00 | 49312.0  |
| 225.00 | 13248.0  |
| 226.00 | 1456.0   |
| 227.00 | 20552.0  |
| 228.00 | 2911.0   |
| 229.00 | 4449.0   |
| 229.90 | 495.0    |
| 231.00 | 1643.0   |
| 231.90 | 278.0    |
| 232.90 | 355.0    |
| 233.90 | 1204.0   |
| 234.90 | 1412.0   |
| 236.00 | 928.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP365

|        |          |
|--------|----------|
| 237.00 | 2006.0   |
| 238.00 | 268.0    |
| 239.00 | 771.0    |
| 240.00 | 745.0    |
| 240.90 | 1199.0   |
| 242.00 | 2886.0   |
| 243.00 | 3078.0   |
| 244.00 | 42544.0  |
| 245.00 | 5284.0   |
| 246.00 | 7756.0   |
| 247.00 | 1412.0   |
| 247.90 | 386.0    |
| 249.00 | 1533.0   |
| 249.90 | 210.0    |
| 251.00 | 555.0    |
| 252.00 | 303.0    |
| 253.00 | 919.0    |
| 254.00 | 1228.0   |
| 255.00 | 200256.0 |
| 256.00 | 29504.0  |
| 256.90 | 2427.0   |
| 258.00 | 11001.0  |
| 258.90 | 1712.0   |
| 259.90 | 316.0    |
| 260.10 | 320.0    |
| 261.00 | 397.0    |
| 263.80 | 287.0    |
| 264.90 | 5079.0   |
| 265.90 | 699.0    |
| 266.90 | 238.0    |
| 270.90 | 463.0    |
| 272.00 | 580.0    |
| 273.00 | 6749.0   |
| 274.00 | 16776.0  |
| 275.00 | 94208.0  |
| 276.00 | 13187.0  |
| 277.00 | 7652.0   |
| 278.00 | 1439.0   |
| 279.00 | 279.0    |
| 281.00 | 1000.0   |
| 281.90 | 379.0    |
| 283.00 | 809.0    |
| 284.00 | 678.0    |
| 284.90 | 1590.0   |
| 285.90 | 219.0    |
| 289.00 | 403.0    |
| 289.90 | 192.0    |
| 290.90 | 284.0    |
| 292.10 | 316.0    |
| 293.00 | 1731.0   |
| 294.00 | 520.0    |
| 295.10 | 435.0    |
| 295.90 | 29728.0  |
| 297.00 | 4135.0   |
| 298.00 | 283.0    |
| 301.00 | 381.0    |
| 302.00 | 525.0    |
| 303.00 | 3032.0   |
| 303.90 | 635.0    |
| 308.00 | 457.0    |
| 308.90 | 322.0    |
| 309.90 | 436.0    |
| 312.90 | 317.0    |
| 314.00 | 1327.0   |
| 315.00 | 3208.0   |
| 316.00 | 1869.0   |
| 317.00 | 340.0    |
| 320.90 | 989.0    |
| 322.00 | 520.0    |
| 323.00 | 8888.0   |
| 324.10 | 1397.0   |
| 325.00 | 175.0    |
| 326.90 | 1871.0   |
| 327.90 | 889.0    |
| 331.90 | 540.0    |
| 332.90 | 901.0    |
| 334.00 | 6100.0   |
| 335.00 | 1619.0   |
| 336.00 | 244.0    |
| 340.90 | 1265.0   |
| 342.10 | 298.0    |
| 345.90 | 2002.0   |
| 347.00 | 399.0    |
| 350.80 | 199.0    |
| 352.00 | 3380.0   |
| 353.00 | 2542.0   |
| 354.00 | 2971.0   |
| 355.00 | 553.0    |
| 358.90 | 233.0    |
| 365.00 | 13694.0  |
| 365.90 | 2018.0   |
| 369.80 | 190.0    |
| 370.10 | 222.0    |
| 370.90 | 762.0    |
| 372.00 | 5140.0   |
| 373.00 | 1598.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP365

|        |          |
|--------|----------|
| 376.90 | 193.0    |
| 383.00 | 1362.0   |
| 384.00 | 423.0    |
| 385.00 | 175.0    |
| 389.90 | 613.0    |
| 390.90 | 462.0    |
| 392.00 | 468.0    |
| 400.80 | 386.0    |
| 402.00 | 2094.0   |
| 403.00 | 3671.0   |
| 404.00 | 1350.0   |
| 404.90 | 229.0    |
| 415.00 | 221.0    |
| 421.00 | 3025.0   |
| 422.00 | 2305.0   |
| 423.00 | 19968.0  |
| 424.00 | 4773.0   |
| 425.00 | 618.0    |
| 441.00 | 53960.0  |
| 442.00 | 332800.0 |
| 443.00 | 72152.0  |
| 444.00 | 5763.0   |
| 445.10 | 277.0    |

**Instrument :**  
BNA\_N  
**ClientSampleId :**  
DFTPP365