

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082120\
 Data File : BN011576.D
 Acq On : 22 Aug 2020 07:53
 Operator : CG/JU
 Sample : L3718-02
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB181-GW-198-200

Quant Time: Aug 24 01:34:09 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082120.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 22 04:52:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	1600	0.40	ng	0.00
7) Naphthalene-d8	10.14	136	6403	0.40	ng	-0.01
13) Acenaphthene-d10	14.03	164	4060	0.40	ng	0.00
19) Phenanthrene-d10	16.79	188	8929	0.40	ng	0.00
29) Chrysene-d12	21.01	240	8856	0.40	ng	0.00
36) Perylene-d12	23.10	264	8012	0.40	ng	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.07	112	629	0.18	ng	0.00
5) Phenol-d6	6.63	99	303	0.09	ng	0.00
8) Nitrobenzene-d5	8.54	82	1961	0.47	ng	-0.01
11) 2-Methylnaphthalene-d10	11.74	152	2870	0.25	ng	0.00
14) 2,4,6-Tribromophenol	15.54	330	778	0.38	ng	-0.01
15) 2-Fluorobiphenyl	12.64	172	9003	0.54	ng	-0.01
27) Fluoranthene-d10	18.83	212	9905	0.33	ng	0.00
31) Terphenyl-d14	19.45	244	16311	0.71	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.11	88	233	0.107	ng	# 100
9) Naphthalene	10.19	128	2718	0.146	ng	99
12) 2-Methylnaphthalene	11.81	142	652	0.051	ng	97
34) Bis(2-ethylhexyl)phthalate	20.95	149	6548	0.495	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082120\
Data File : BN011576.D
Acq On : 22 Aug 2020 07:53
Operator : CG/JU
Sample : L3718-02
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB181-GW-198-200

Quant Time: Aug 24 01:34:09 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082120.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 22 04:52:07 2020
Response via : Initial Calibration

