

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082120\
 Data File : BN011583.D
 Acq On : 22 Aug 2020 12:06
 Operator : CG/JU
 Sample : L3718-05
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB181-GW-238-240

Quant Time: Aug 24 01:58:30 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082120.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 22 04:52:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	1470	0.40	ng	0.00
7) Naphthalene-d8	10.14	136	5589	0.40	ng	-0.01
13) Acenaphthene-d10	14.03	164	3641	0.40	ng	0.00
19) Phenanthrene-d10	16.79	188	7432	0.40	ng	0.00
29) Chrysene-d12	21.01	240	6244	0.40	ng	0.00
36) Perylene-d12	23.10	264	6125	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.06	112	704	0.22	ng	0.00
5) Phenol-d6	6.63	99	325	0.10	ng	0.00
8) Nitrobenzene-d5	8.54	82	1661	0.45	ng	-0.01
11) 2-Methylnaphthalene-d10	11.74	152	2506	0.25	ng	0.00
14) 2,4,6-Tribromophenol	15.54	330	660	0.36	ng	-0.01
15) 2-Fluorobiphenyl	12.64	172	8246	0.55	ng	-0.01
27) Fluoranthene-d10	18.83	212	7300	0.29	ng	0.00
31) Terphenyl-d14	19.45	244	11819	0.73	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.11	88	578	0.288	ng	# 100
9) Naphthalene	10.19	128	411	0.025	ng	# 56
34) Bis(2-ethylhexyl)phthalate	20.95	149	3083	0.331	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082120\
Data File : BN011583.D
Acq On : 22 Aug 2020 12:06
Operator : CG/JU
Sample : L3718-05
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB181-GW-238-240

Quant Time: Aug 24 01:58:30 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082120.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 22 04:52:07 2020
Response via : Initial Calibration

