

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082120\
 Data File : BN011568.D
 Acq On : 22 Aug 2020 02:49
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 22 05:50:48 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082120.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 22 04:52:07 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	99	0.00
2	1,4-Dioxane	0.400	0.424	-6.0	105	0.00
3	n-Nitrosodimethylamine	0.400	0.338	15.5	89	0.00
4 S	2-Fluorophenol	0.400	0.374	6.5	106	0.00
5 S	Phenol-d6	0.400	0.353	11.8	115	0.00
6	bis(2-Chloroethyl)ether	0.400	0.326	18.5	105	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	103	-0.01
8 S	Nitrobenzene-d5	0.400	0.329	17.8	102	0.00
9	Naphthalene	0.400	0.373	6.8	101	-0.01
10	Hexachlorobutadiene	0.400	0.389	2.8	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.383	4.3	111	0.00
12	2-Methylnaphthalene	0.400	0.376	6.0	107	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	112	0.00
14 S	2,4,6-Tribromophenol	0.400	0.349	12.8	115	0.00
15 S	2-Fluorobiphenyl	0.400	0.383	4.3	113	-0.01
16	Acenaphthylene	0.400	0.348	13.0	108	0.00
17	Acenaphthene	0.400	0.363	9.3	109	0.00
18	Fluorene	0.400	0.367	8.3	112	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	122	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.374	6.5	133	-0.01
21	4-Bromophenyl-phenylether	0.400	0.319	20.3	83	0.00
22	Hexachlorobenzene	0.400	0.349	12.8	112	0.00
23	Atrazine	0.400	0.247	38.3#	123	0.00
24	Pentachlorophenol	0.400	0.315	21.3	121	-0.01
25	Phenanthrene	0.400	0.344	14.0	117	0.00
26	Anthracene	0.400	0.329	17.8	120	0.00
27 SURR	Fluoranthene-d10	0.400	0.322	19.5	121	0.00
28	Fluoranthene	0.400	0.325	18.8	121	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	139	0.00
30	Pyrene	0.400	0.348	13.0	121	0.00
31 S	Terphenyl-d14	0.400	0.344	14.0	127	0.00
32	Benzo(a)anthracene	0.400	0.354	11.5	146	-0.01
33	Chrysene	0.400	0.355	11.3	129	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.369	7.8	141	-0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.342	14.5	130	-0.01
36 I	Perylene-d12	0.400	0.400	0.0	130	-0.01
37	Benzo(b)fluoranthene	0.400	0.365	8.8	157	0.00
38	Benzo(k)fluoranthene	0.400	0.364	9.0	144	0.00
39 C	Benzo(a)pyrene	0.400	0.358	10.5	132	0.00
40	Dibenzo(a,h)anthracene	0.400	0.333	16.8	126	-0.01
41	Benzo(g,h,i)perylene	0.400	0.341	14.8	123	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082120\
Data File : BN011568.D
Acq On : 22 Aug 2020 02:49
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 22 05:50:48 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082120.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 22 04:52:07 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)
----------	--------	-------	------	------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0