

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082120\
 Data File : BN011584.D
 Acq On : 22 Aug 2020 12:43
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 24 01:24:06 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082120.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 22 04:52:07 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	140	0.00
2	1,4-Dioxane	0.400	0.332	17.0	116	0.00
3	n-Nitrosodimethylamine	0.400	0.295	26.3#	110	0.00
4 S	2-Fluorophenol	0.400	0.358	10.5	143	0.00
5 S	Phenol-d6	0.400	0.348	13.0	159	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.320	20.0	146	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	133	-0.01
8 S	Nitrobenzene-d5	0.400	0.410	-2.5	165	0.00
9	Naphthalene	0.400	0.375	6.3	132	-0.01
10	Hexachlorobutadiene	0.400	0.376	6.0	125	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.389	2.8	146	0.00
12	2-Methylnaphthalene	0.400	0.397	0.8	146	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	136	0.00
14 S	2,4,6-Tribromophenol	0.400	0.332	17.0	133	-0.01
15 S	2-Fluorobiphenyl	0.400	0.435	-8.7	156	-0.01
16	Acenaphthylene	0.400	0.354	11.5	134	-0.01
17	Acenaphthene	0.400	0.367	8.3	134	0.00
18	Fluorene	0.400	0.358	10.5	132	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	149	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.354	11.5	145	-0.02
21	4-Bromophenyl-phenylether	0.400	0.242	39.5#	77	0.00
22	Hexachlorobenzene	0.400	0.331	17.3	130	-0.01
23	Atrazine	0.400	0.222	44.5#	134	-0.01
24	Pentachlorophenol	0.400	0.273	31.8#	129	-0.01
25	Phenanthrene	0.400	0.354	11.5	147	-0.01
26	Anthracene	0.400	0.331	17.3	147	-0.01
27 SURR	Fluoranthene-d10	0.400	0.280	30.0#	128	0.00
28	Fluoranthene	0.400	0.279	30.3#	127	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	146	0.00
30	Pyrene	0.400	0.342	14.5	125	0.00
31 S	Terphenyl-d14	0.400	0.323	19.3	125	0.00
32	Benzo(a)anthracene	0.400	0.349	12.8	152	-0.01
33	Chrysene	0.400	0.359	10.3	137	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.327	18.3	132	-0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.449	-12.2	179	-0.01
36 I	Perylene-d12	0.400	0.400	0.0	150	-0.01
37	Benzo(b)fluoranthene	0.400	0.308	23.0	153	0.00
38	Benzo(k)fluoranthene	0.400	0.302	24.5	138	0.00
39 C	Benzo(a)pyrene	0.400	0.267	33.3#	113	0.00
40	Dibenzo(a,h)anthracene	0.400	0.424	-6.0	186	-0.01
41	Benzo(g,h,i)perylene	0.400	0.400	0.0	166	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082120\
Data File : BN011584.D
Acq On : 22 Aug 2020 12:43
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 24 01:24:06 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082120.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 22 04:52:07 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1