


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082218\
Data File : BN002448.D
Acq On : 22 Aug 2018 09:48
Operator : JU/SJ
Sample : DFTPP06
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP06

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN081318MA.M
Title : SVOA CALIBRATION
Last Update : Thu Aug 23 04:38:12 2018



AutoFind: Scans 2467, 2468, 2469; Background Corrected with Scan 2461

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.6	115816	PASS
68	69	0.00	2	1.6	1820	PASS
69	198	0.00	100	41.1	117072	PASS
70	69	0.00	2	0.5	634	PASS
127	198	10	80	48.1	137106	PASS
197	198	0.00	2	0.5	1424	PASS
198	198	100	100	100.0	284928	PASS
199	198	5	9	6.8	19467	PASS
275	198	10	60	26.8	76429	PASS
365	198	1	100	3.5	9998	PASS
441	198	0.01	100	14.3	40701	PASS
442	198	50	100	95.7	272768	PASS
443	442	15	24	19.1	51984	PASS

m/z	Abundance
37.00	371.0
38.00	1390.0
39.00	9816.0
40.00	1305.0
41.00	552.0
44.00	10511.0
44.90	308.0
49.10	993.0
50.00	30280.0
51.00	129464.0
52.00	6706.0
53.00	315.0
54.90	679.0
56.00	3616.0
57.00	9332.0
58.00	381.0
61.00	1685.0
62.00	1899.0
63.00	5484.0
63.90	724.0
65.00	2231.0
68.00	2140.0
69.00	130512.0
70.00	859.0
73.00	1331.0
74.00	12370.0
75.00	20456.0
76.00	7156.0
77.00	140160.0
78.00	9294.0
79.00	9383.0
80.00	6712.0
81.00	10052.0
81.90	2394.0
82.90	2562.0
83.90	383.0
85.00	1972.0
86.00	2934.0
86.90	1088.0
87.80	522.0
90.90	2288.0
91.90	2427.0
93.00	15246.0
93.90	1068.0
94.90	413.0
96.00	700.0
97.10	361.0
98.00	10693.0
99.00	10409.0
99.90	989.0
101.00	5331.0
103.00	1725.0
103.90	3523.0
105.00	3483.0
106.00	1326.0
107.00	40096.0
108.00	6745.0
109.00	1005.0
110.00	73984.0
111.00	11175.0
111.90	1439.0
112.90	454.0
116.00	2437.0
117.00	31128.0
117.90	2458.0
118.90	494.0
120.10	538.0
120.90	323.0
122.00	2543.0
122.90	4433.0
124.00	1920.0
125.00	1733.0
127.00	142720.0
128.00	11124.0
129.00	55800.0
130.00	4249.0
131.00	731.0
132.00	542.0
133.00	323.0
133.90	1638.0
135.00	4676.0
135.90	2019.0
137.00	2585.0
138.00	377.0
139.90	705.0
140.90	7598.0
141.90	2011.0
143.00	1750.0
144.10	551.0
144.90	397.0
145.90	1378.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06

146.90	3410.0
147.90	7696.0
149.00	1961.0
150.00	436.0
151.30	941.0
152.90	2193.0
153.90	1524.0
155.00	3516.0
156.00	6598.0
157.00	1195.0
157.90	1142.0
158.90	1008.0
160.00	2016.0
160.90	3755.0
161.90	841.0
165.00	2756.0
166.00	2026.0
167.00	13164.0
168.00	7571.0
168.90	1284.0
170.20	512.0
170.90	553.0
171.90	1182.0
173.00	1535.0
174.00	2817.0
175.00	5629.0
176.00	1959.0
176.90	2484.0
177.90	741.0
178.90	10940.0
180.00	7514.0
180.90	3508.0
182.00	785.0
183.10	323.0
184.00	761.0
185.00	4927.0
186.00	37208.0
187.00	11453.0
188.00	1257.0
188.90	2485.0
190.00	448.0
190.90	1049.0
191.90	3060.0
193.00	3587.0
194.00	968.0
195.00	389.0
196.00	7392.0
197.00	1219.0
197.90	279040.0
198.90	19776.0
199.90	1495.0
201.30	1220.0
202.90	1958.0
204.00	8772.0
205.00	16520.0
206.00	72064.0
207.00	10894.0
208.00	2367.0
208.90	1026.0
210.00	1164.0
211.00	3235.0
211.80	640.0
214.80	787.0
216.00	1465.0
216.90	18792.0
217.90	1914.0
221.00	10661.0
223.00	4412.0
224.00	42168.0
225.00	10656.0
226.00	1088.0
227.00	16544.0
228.00	1972.0
228.90	3449.0
230.00	501.0
231.00	1291.0
234.00	879.0
235.00	1053.0
235.90	665.0
236.90	1706.0
238.90	982.0
239.90	383.0
240.90	990.0
242.00	1612.0
243.00	2584.0
244.00	28936.0
245.00	4163.0
245.90	5721.0
246.90	1266.0
249.00	1180.0
252.00	354.0
253.00	600.0
255.00	157440.0
256.00	23008.0
257.00	2155.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06

258.00	7911.0
259.00	1599.0
259.90	307.0
265.00	3058.0
266.10	742.0
270.90	662.0
273.00	4033.0
274.00	12490.0
275.00	68128.0
276.00	8754.0
277.00	5105.0
278.00	1020.0
283.00	705.0
284.00	570.0
285.00	983.0
290.90	352.0
292.10	403.0
293.00	1424.0
295.00	423.0
296.00	17648.0
296.90	2632.0
301.90	344.0
303.00	2337.0
303.90	692.0
307.80	337.0
314.00	923.0
314.90	1744.0
316.00	1278.0
320.90	560.0
322.10	329.0
323.00	6030.0
324.00	1169.0
326.90	1118.0
327.90	687.0
331.90	423.0
332.90	314.0
334.00	3684.0
335.00	1031.0
341.00	700.0
345.80	737.0
352.00	1852.0
353.00	1257.0
354.00	1602.0
354.90	433.0
364.90	8445.0
365.90	1386.0
371.00	659.0
372.00	2800.0
372.90	902.0
382.90	795.0
390.00	386.0
390.80	320.0
400.90	304.0
401.90	1258.0
403.00	1450.0
404.00	590.0
421.00	1290.0
422.00	1581.0
423.00	9899.0
424.00	3005.0
425.00	419.0
441.00	29544.0
442.00	204160.0
443.00	37240.0
444.00	3437.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06

m/z	Abundance
37.00	424.0
38.00	1641.0
39.00	10357.0
40.00	1517.0
41.00	398.0
44.00	10713.0
44.90	547.0
49.00	854.0
50.00	34792.0
51.00	136832.0
52.00	7136.0
52.90	390.0
55.00	616.0
55.90	4064.0
57.00	9570.0
57.90	320.0
61.00	1607.0
62.00	1876.0
63.00	5880.0
64.00	650.0
65.00	2723.0
67.90	1810.0
69.00	136512.0
70.00	524.0
73.00	1230.0
74.00	13562.0
75.00	21336.0
76.00	7287.0
77.00	153536.0
78.00	10867.0
79.00	10338.0
79.90	7695.0
81.00	10665.0
82.00	2507.0
83.00	2848.0
83.90	327.0
84.90	1600.0
86.00	3065.0
86.90	1324.0
87.90	559.0
88.90	327.0
91.00	2410.0
92.00	2952.0
93.00	15771.0
93.90	1005.0
96.00	673.0
98.00	11900.0
99.00	10557.0
99.90	1228.0
101.00	5677.0
102.00	353.0
102.90	2095.0
103.90	3811.0
105.00	3396.0
106.00	1160.0
107.00	46344.0
108.00	7090.0
109.10	1320.0
110.00	84176.0
111.00	12296.0
112.00	1528.0
112.90	644.0
116.00	2564.0
117.00	34336.0
117.90	2841.0
119.90	507.0
122.00	2431.0
123.00	4769.0
124.00	1844.0
124.90	1760.0
127.00	161280.0
128.00	12439.0
129.00	62168.0
130.00	5095.0
131.00	1055.0
131.90	616.0
133.90	1863.0
135.00	4882.0
135.90	1831.0
137.00	3129.0
138.10	462.0
138.80	380.0
139.90	928.0
140.90	8047.0
142.00	2343.0
142.90	1914.0
143.90	567.0
144.90	731.0
146.00	1541.0
147.00	3981.0
147.90	8723.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06

149.00	2147.0
149.80	477.0
151.10	1063.0
151.70	765.0
152.90	2407.0
153.90	2007.0
155.00	5024.0
156.00	7149.0
157.10	1210.0
157.90	1645.0
158.90	1210.0
159.90	2440.0
161.00	3797.0
162.00	1160.0
163.10	352.0
163.90	454.0
164.90	2545.0
166.00	2523.0
167.00	14893.0
168.00	7343.0
169.00	1427.0
170.00	595.0
170.90	728.0
171.90	1619.0
173.00	2057.0
174.00	3471.0
174.90	6286.0
176.00	1797.0
176.90	3388.0
178.00	1021.0
178.90	12655.0
180.00	8861.0
181.00	3829.0
182.00	742.0
182.90	700.0
184.00	929.0
184.90	5714.0
186.00	46152.0
187.00	13015.0
187.90	1272.0
188.90	2701.0
189.90	598.0
190.90	1325.0
192.00	3719.0
193.00	4289.0
193.90	819.0
195.00	664.0
195.90	8750.0
197.00	1972.0
197.90	339776.0
198.90	22488.0
199.90	1959.0
201.50	1817.0
202.90	2529.0
204.00	12614.0
205.00	20640.0
206.00	88448.0
207.00	12731.0
207.90	3178.0
209.00	1021.0
210.00	1428.0
211.00	3403.0
211.70	860.0
214.80	652.0
215.90	1876.0
216.90	23904.0
217.90	3135.0
218.90	303.0
221.00	14011.0
223.00	5406.0
224.00	47464.0
225.00	12267.0
226.00	1437.0
227.00	19744.0
227.90	3249.0
228.90	4188.0
230.00	925.0
231.00	1833.0
232.00	358.0
233.00	412.0
233.90	1038.0
235.00	1406.0
235.90	781.0
237.00	1692.0
238.90	914.0
239.90	773.0
240.90	983.0
242.00	2805.0
243.00	2828.0
244.00	37152.0
245.00	5212.0
245.90	7838.0
246.90	1347.0
247.90	498.0
248.90	1327.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06

249.80	310.0
250.90	323.0
252.90	897.0
254.00	1616.0
255.00	204032.0
256.00	28360.0
257.00	2430.0
258.00	10953.0
259.00	1623.0
260.90	415.0
264.20	333.0
265.00	4269.0
266.00	446.0
270.90	477.0
272.00	649.0
273.00	6191.0
274.00	15231.0
275.00	91808.0
276.00	12158.0
276.90	6985.0
277.90	1097.0
279.00	422.0
282.90	971.0
284.00	840.0
285.00	1502.0
289.00	384.0
292.00	524.0
292.90	1849.0
294.00	642.0
294.90	577.0
296.00	24920.0
296.90	3184.0
302.00	464.0
303.00	2908.0
304.00	871.0
308.00	489.0
309.90	343.0
314.00	1096.0
315.00	2762.0
315.90	1624.0
321.00	844.0
322.00	376.0
323.00	7836.0
324.00	1052.0
326.00	321.0
326.90	1747.0
328.00	887.0
331.90	689.0
333.00	922.0
334.00	5122.0
335.00	1267.0
341.00	960.0
346.00	1733.0
347.00	383.0
351.90	2268.0
353.00	1893.0
354.00	2394.0
355.10	603.0
364.90	11537.0
365.90	1457.0
369.90	365.0
371.00	519.0
372.00	4650.0
372.90	1108.0
376.90	357.0
383.00	1356.0
389.90	659.0
391.00	396.0
400.80	329.0
402.00	1914.0
403.00	2889.0
404.00	1020.0
420.90	2351.0
422.00	2485.0
423.00	15734.0
424.00	3996.0
425.10	474.0
441.00	49880.0
442.00	320128.0
443.00	63608.0
444.00	5814.0
444.90	453.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06

m/z	Abundance
37.00	367.0
38.00	975.0
39.00	6090.0
40.00	1029.0
40.90	367.0
44.00	11544.0
45.10	357.0
48.90	691.0
50.00	19224.0
51.00	81152.0
52.00	3980.0
54.90	467.0
55.90	2345.0
57.00	6265.0
61.10	988.0
62.00	1143.0
63.00	3758.0
64.10	484.0
65.00	1784.0
68.10	1511.0
69.00	84192.0
70.00	519.0
72.90	947.0
74.00	7406.0
75.00	12850.0
76.00	4507.0
77.00	93200.0
78.10	5906.0
79.00	6114.0
80.00	4668.0
81.00	6502.0
81.90	1672.0
83.00	1890.0
85.00	1155.0
86.00	1992.0
87.00	816.0
90.90	1579.0
91.90	1566.0
93.00	10151.0
93.90	605.0
96.00	581.0
98.00	7448.0
99.00	6977.0
100.00	471.0
100.90	3382.0
102.90	1290.0
104.00	2221.0
105.00	2301.0
106.00	851.0
107.00	28920.0
108.00	4858.0
109.00	852.0
110.00	51872.0
111.00	8124.0
112.00	979.0
112.90	472.0
116.00	1654.0
117.00	22152.0
117.90	1688.0
121.90	1966.0
123.00	3174.0
123.90	1126.0
124.90	1116.0
126.10	580.0
127.00	107320.0
128.00	8735.0
128.90	41400.0
130.00	3431.0
130.90	718.0
131.90	505.0
133.90	1028.0
135.00	3017.0
136.00	1133.0
137.00	1680.0
137.90	350.0
140.00	654.0
140.90	5554.0
141.90	1748.0
143.00	1161.0
144.10	338.0
145.00	613.0
145.90	939.0
147.00	3253.0
148.00	6119.0
149.00	1114.0
149.90	407.0
151.00	790.0
151.90	460.0
152.90	1600.0
154.00	1098.0
154.90	3083.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06

156.00	4501.0
156.90	734.0
157.80	808.0
159.00	908.0
159.80	1639.0
161.00	2647.0
161.90	604.0
162.90	335.0
164.90	2191.0
165.90	2107.0
167.00	10263.0
168.00	5148.0
169.00	1037.0
169.80	437.0
170.10	445.0
170.80	378.0
172.00	1080.0
172.90	1257.0
173.90	2790.0
175.00	3838.0
175.90	931.0
177.00	2191.0
178.00	947.0
178.90	8081.0
180.00	6464.0
181.00	2719.0
182.00	313.0
183.00	389.0
184.00	765.0
185.00	4136.0
186.00	32136.0
187.00	9763.0
188.00	937.0
189.00	2211.0
190.90	939.0
191.90	2450.0
193.00	3250.0
194.10	397.0
196.00	7083.0
197.00	1081.0
197.90	235968.0
198.90	16137.0
199.90	1212.0
201.30	919.0
202.90	1473.0
204.00	7994.0
205.00	14854.0
206.00	63056.0
207.00	8581.0
208.00	2179.0
208.90	605.0
210.90	2658.0
211.80	739.0
215.00	691.0
215.90	1194.0
216.90	16150.0
217.90	1966.0
221.00	9826.0
221.70	2981.0
223.00	3907.0
224.00	33496.0
225.00	9014.0
226.00	801.0
227.00	14778.0
227.90	2324.0
229.00	3119.0
230.00	473.0
230.90	1276.0
234.00	664.0
234.90	998.0
235.80	433.0
235.90	423.0
236.90	1235.0
239.00	617.0
239.90	541.0
241.00	856.0
242.00	1836.0
243.00	2380.0
244.00	26712.0
245.00	3645.0
245.90	5528.0
246.90	1047.0
248.90	1162.0
251.80	316.0
252.90	475.0
253.90	1005.0
255.00	154432.0
256.00	20736.0
257.00	1564.0
257.90	8320.0
259.00	1095.0
264.90	3779.0
266.00	569.0
271.00	414.0
272.00	453.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06

272.90	4253.0
274.00	12106.0
275.00	69352.0
276.00	9056.0
276.90	5776.0
278.00	1025.0
283.00	682.0
283.90	564.0
285.00	1051.0
292.00	322.0
293.00	1191.0
294.00	480.0
294.80	454.0
295.90	18120.0
296.90	2801.0
297.90	445.0
302.00	367.0
303.00	2211.0
303.90	669.0
307.90	347.0
314.00	778.0
314.90	2145.0
315.90	1258.0
321.00	696.0
322.00	408.0
323.00	6330.0
324.00	1199.0
326.90	1341.0
328.10	603.0
332.00	554.0
333.00	693.0
334.00	3796.0
335.00	866.0
341.10	707.0
345.90	1555.0
351.90	2085.0
353.00	1310.0
353.90	2192.0
355.00	455.0
364.90	10012.0
365.90	1060.0
371.00	586.0
372.00	3321.0
372.90	1011.0
382.90	1177.0
384.00	328.0
390.00	542.0
390.90	309.0
400.90	332.0
401.90	1977.0
403.00	2586.0
404.00	829.0
420.90	2017.0
422.00	1825.0
423.00	14503.0
424.00	3241.0
441.00	42680.0
442.00	294016.0
443.00	55104.0
444.00	5090.0
445.00	435.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06

Scan 2461 (17.463 min): BN002448.D
DFTPP06

m/z	Abundance
39.90	872.0
44.00	10015.0
207.00	641.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06