

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082318\
 Data File : BN002481.D
 Acq On : 24 Aug 2018 00:11
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.416

Quant Time: Aug 24 00:56:24 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN082318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 24 00:52:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	1473	0.40	ng/ul	0.00
2) Naphthalene-d8	10.46	136	6150	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.31	164	3566	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.06	188	8758	0.40	ng/ul	0.00
16) Chrysene-d12	21.26	240	9264	0.40	ng/ul	0.00
20) Perylene-d12	23.48	264	8439	0.40	ng/ul	-0.01

System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.05	152	3956	0.41	ng/ul	0.00
14) Fluoranthene-d10	19.09	212	10671	0.41	ng/ul	0.00

Target Compounds				Ovalue		
3) Naphthalene	10.50	128	6339	0.400	ng/ul#	96
5) 2-Methylnaphthalene	12.13	142	4447	0.403	ng/ul	100
7) Acenaphthylene	14.03	152	6844	0.410	ng/ul	97
8) Acenaphthene	14.37	153	5181	0.407	ng/ul	94
9) Fluorene	15.37	166	5937	0.407	ng/ul	92
11) Pentachlorophenol	16.74	266	461	0.390	ng/ul	97
12) Phenanthrene	17.10	178	10534	0.407	ng/ul#	89
13) Anthracene	17.20	178	9071	0.407	ng/ul#	92
15) Fluoranthene	19.12	202	12518	0.411	ng/ul	96
17) Pyrene	19.49	202	12215	0.387	ng/ul	97
18) Benzo(a)anthracene	21.24	228	11435	0.413	ng/ul#	85
19) Chrysene	21.29	228	12641	0.404	ng/ul	97
21) Benzo(b)fluoranthene	22.81	252	12882	0.425	ng/ul	87
22) Benzo(k)fluoranthene	22.86	252	12740	0.407	ng/ul#	87
23) Benzo(a)pyrene	23.38	252	11891	0.417	ng/ul#	93
24) Indeno(1,2,3-cd)pyrene	25.70	276	11525	0.388	ng/ul#	85
25) Dibenzo(a,h)anthracene	25.71	278	9093	0.388	ng/ul#	90
26) Benzo(g,h,i)perylene	26.38	276	9550	0.375	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082318\
Data File : BN002481.D
Acq On : 24 Aug 2018 00:11
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.416

Quant Time: Aug 24 00:56:24 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN082318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 24 00:52:48 2018
Response via : Initial Calibration

