

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082319\
 Data File : BN007342.D
 Acq On : 24 Aug 2019 00:25
 Operator : HP/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: Aug 24 07:46:18 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 24 00:38:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	131	0.00
2	1,4-Dioxane	0.665	0.642	3.5	121	0.00
3	n-Nitrosodimethylamine	0.309	0.281	9.1	115	0.00
4 S	2-Fluorophenol	1.490	1.361	8.7	126	0.00
5 S	Phenol-d6	1.953	1.679	14.0	129	0.00
6	bis(2-Chloroethyl)ether	1.445	1.475	-2.1	134	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	133	0.00
8 S	Nitrobenzene-d5	0.348	0.346	0.6	138	0.00
9	Naphthalene	1.143	1.178	-3.1	133	0.00
10	Hexachlorobutadiene	0.234	0.241	-3.0	131	0.00
11 SURR	2-Methylnaphthalene-d10	0.665	0.687	-3.3	135	0.00
12	2-Methylnaphthalene	0.779	0.800	-2.7	132	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	132	0.00
14 S	2,4,6-Tribromophenol	0.209	0.193	7.7	125	0.00
15 S	2-Fluorobiphenyl	1.613	1.658	-2.8	133	0.00
16	Acenaphthylene	2.367	2.371	-0.2	131	0.00
17	Acenaphthene	1.385	1.411	-1.9	132	0.00
18	Fluorene	1.751	1.757	-0.3	130	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	131	0.00
20	4-Bromophenyl-phenylether	0.158	0.180	-13.9	173#	0.00
21	Hexachlorobenzene	0.268	0.281	-4.9	136	0.00
22	Pentachlorophenol	0.083	0.083	0.0	139	0.00
23	Phenanthrene	1.203	1.202	0.1	131	0.00
24	Anthracene	1.210	1.191	1.6	128	0.00
25 SURR	Fluoranthene-d10	1.288	1.346	-4.5	133	0.00
26	Fluoranthene	1.545	1.597	-3.4	132	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	130	0.00
28	Pvrene	1.609	1.662	-3.3	131	0.00
29 S	Terphenyl-d14	1.047	1.103	-5.3	132	0.00
30	Benzo(a)anthracene	1.571	1.606	-2.2	131	0.00
31	Chrysene	1.516	1.597	-5.3	131	0.00
32	Bis(2-ethylhexyl)phthalate	1.177	1.100	6.5	128	0.00
33	Indeno(1,2,3-cd)pyrene	1.827	1.875	-2.6	129	0.00
34 I	Perylene-d12	1.000	1.000	0.0	129	0.00
35	Benzo(b)fluoranthene	1.449	1.468	-1.3	128	0.00
36	Benzo(k)fluoranthene	1.432	1.536	-7.3	133	0.00
37 C	Benzo(a)pyrene	1.379	1.417	-2.8	129	0.00
38	Dibenzo(a,h)anthracene	1.376	1.401	-1.8	127	0.00
39	Benzo(a,h,i)perylene	1.371	1.400	-2.1	128	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082319\
Data File : BN007342.D
Acq On : 24 Aug 2019 00:25
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 24 07:46:18 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 24 00:38:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			