


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082418\
 Data File : BN002501.D
 Acq On : 24 Aug 2018 15:19
 Operator : JU/SJ
 Sample : DFTPP08
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP08

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN081318MA.M
 Title : SVOA CALIBRATION
 Last Update : Sat Aug 25 04:34:18 2018



AutoFind: Scans 2467, 2468, 2469; Background Corrected with Scan 2461

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.8	111138	PASS
68	69	0.00	2	1.7	1899	PASS
69	198	0.00	100	39.1	109330	PASS
70	69	0.00	2	0.4	460	PASS
127	198	10	80	47.0	131376	PASS
197	198	0.00	2	0.5	1405	PASS
198	198	100	100	100.0	279466	PASS
199	198	5	9	6.7	18811	PASS
275	198	10	60	26.4	73741	PASS
365	198	1	100	3.4	9387	PASS
441	198	0.01	100	14.6	40858	PASS
442	198	50	100	96.3	269013	PASS
443	442	15	24	19.0	51141	PASS

m/z	Abundance
37.00	538.0
38.00	1516.0
39.00	10096.0
40.00	1174.0
41.00	321.0
44.00	4557.0
49.00	1120.0
50.00	33992.0
51.00	137984.0
52.00	6736.0
54.90	747.0
56.00	4137.0
57.00	9104.0
58.00	479.0
61.00	1977.0
62.00	1969.0
63.00	5825.0
64.00	713.0
65.00	2654.0
68.00	2220.0
69.00	132288.0
70.10	639.0
72.90	1064.0
74.00	12881.0
75.00	20800.0
76.00	7401.0
77.00	145024.0
78.00	10189.0
79.00	9461.0
80.00	6904.0
81.00	10434.0
81.90	2821.0
82.90	2230.0
85.00	1889.0
85.90	2860.0
86.90	1127.0
87.90	336.0
91.00	2485.0
91.90	2724.0
93.00	16160.0
94.00	1300.0
95.00	346.0
95.90	777.0
96.80	447.0
98.00	10985.0
99.00	9814.0
100.00	684.0
100.90	5839.0
103.00	1823.0
103.90	3341.0
105.00	3771.0
106.00	1155.0
107.00	43120.0
108.00	7654.0
109.00	1226.0
110.00	78344.0
111.00	11853.0
112.00	1372.0
113.00	438.0
116.00	2283.0
117.00	28624.0
118.00	2536.0
119.00	313.0
119.90	452.0
121.90	2894.0
122.90	4229.0
123.90	1827.0
124.90	1640.0
127.00	153536.0
128.00	11920.0
129.00	57448.0
130.00	5230.0
130.90	825.0
131.90	700.0
133.90	1603.0
135.00	3974.0
136.00	1753.0
137.00	2510.0
137.90	538.0
140.00	673.0
141.00	6935.0
142.00	2222.0
143.00	1624.0
143.90	445.0
145.00	471.0
146.00	1128.0
146.90	3437.0
148.00	7919.0
148.90	1550.0
149.90	625.0
150.90	940.0

Instrument :
BNA_N
ClientSampleId :
DFTPP08

152.00	608.0
152.90	2071.0
154.00	1681.0
155.00	3958.0
156.00	6165.0
157.00	1089.0
158.00	1552.0
158.90	1314.0
159.90	2323.0
161.00	3492.0
161.90	1139.0
163.00	418.0
164.00	377.0
164.90	2771.0
165.90	1856.0
167.00	13883.0
168.00	6964.0
169.00	1382.0
170.00	450.0
170.90	692.0
171.90	1248.0
172.90	1735.0
174.00	2627.0
175.00	5934.0
175.90	1828.0
176.90	2312.0
177.90	1088.0
178.90	10901.0
179.90	7586.0
180.90	2898.0
181.90	555.0
183.00	447.0
183.90	718.0
185.00	5221.0
186.00	41400.0
187.00	12638.0
188.00	1127.0
188.90	2425.0
189.90	350.0
191.00	1289.0
191.90	3340.0
193.00	4318.0
194.00	1142.0
194.90	355.0
195.90	7821.0
197.00	1555.0
197.90	297984.0
198.90	20776.0
200.00	1718.0
201.40	1248.0
202.90	1963.0
204.00	10675.0
205.00	18232.0
206.00	77944.0
207.00	11664.0
208.00	2460.0
208.90	702.0
210.90	3295.0
214.80	769.0
216.00	1454.0
216.90	19256.0
217.90	2324.0
220.00	458.0
221.00	10297.0
223.00	4543.0
224.00	39536.0
225.00	10834.0
226.00	1293.0
227.00	16640.0
227.90	2283.0
229.00	3474.0
230.00	857.0
231.00	1595.0
233.90	1182.0
235.00	1215.0
235.90	749.0
237.00	1748.0
237.90	335.0
238.80	735.0
239.90	676.0
240.90	911.0
242.00	1920.0
243.00	2396.0
244.00	31256.0
245.00	4368.0
245.90	6050.0
246.90	1316.0
248.90	1130.0
250.90	464.0
252.00	354.0
253.00	661.0
254.10	990.0
255.00	165312.0
256.00	23968.0
257.00	1929.0

Instrument :
BNA_N
ClientSampleId :
DFTPP08

258.00	8392.0
259.00	1367.0
264.00	318.0
265.00	3678.0
265.80	541.0
270.90	521.0
272.00	505.0
273.00	4753.0
273.90	12075.0
275.00	71952.0
276.00	9794.0
277.00	5362.0
277.90	941.0
279.00	335.0
283.00	658.0
283.80	365.0
285.00	1076.0
291.90	408.0
292.90	1177.0
293.90	520.0
295.00	555.0
295.90	18280.0
297.00	2651.0
301.10	367.0
301.90	341.0
303.00	2409.0
304.00	779.0
307.90	313.0
314.00	1024.0
314.90	1783.0
316.00	1211.0
320.90	632.0
322.00	338.0
323.00	6680.0
324.10	1389.0
327.00	1200.0
327.90	777.0
332.00	320.0
333.10	646.0
334.00	4157.0
335.00	1272.0
341.00	764.0
341.90	409.0
346.00	1166.0
352.10	1807.0
353.00	1288.0
354.00	1874.0
354.90	378.0
364.90	7743.0
366.00	1176.0
371.00	341.0
372.00	3007.0
372.90	648.0
383.00	988.0
383.90	361.0
389.90	459.0
401.90	1534.0
403.00	2076.0
404.00	702.0
421.00	1530.0
422.00	1190.0
423.00	11541.0
424.00	3341.0
441.00	33488.0
442.00	223616.0
443.00	44872.0
444.00	3953.0

Instrument :
BNA_N
ClientSampleId :
DFTPP08

m/z	Abundance
37.00	611.0
38.00	1580.0
39.00	9519.0
40.00	1264.0
41.00	312.0
44.00	5095.0
49.00	1057.0
50.00	31696.0
51.00	128256.0
52.00	6640.0
55.00	556.0
55.90	4141.0
57.00	9291.0
57.90	305.0
60.90	1505.0
61.90	1713.0
63.00	5047.0
64.00	734.0
65.00	2534.0
68.00	2479.0
69.00	127672.0
70.00	741.0
73.00	1110.0
74.00	12398.0
75.00	20400.0
76.00	7205.0
77.00	139968.0
78.00	9298.0
79.00	9886.0
80.00	6816.0
81.00	10994.0
82.00	2456.0
82.90	2458.0
83.90	563.0
85.00	1906.0
86.00	2515.0
86.90	1219.0
87.90	545.0
91.00	2355.0
92.00	2048.0
93.00	15780.0
94.00	963.0
96.10	632.0
98.00	11317.0
99.00	10180.0
100.00	893.0
100.90	5389.0
101.80	471.0
102.90	1719.0
104.00	3269.0
104.90	3154.0
106.00	1069.0
107.00	43768.0
107.90	6597.0
109.00	1216.0
110.00	79264.0
111.00	11909.0
112.00	1689.0
112.90	367.0
114.70	313.0
116.00	1934.0
117.00	31624.0
117.90	2398.0
119.90	724.0
121.90	2527.0
123.00	4519.0
124.00	2098.0
125.00	1744.0
127.00	155136.0
128.00	11151.0
129.00	61000.0
130.00	5700.0
130.90	983.0
132.00	472.0
134.00	1865.0
135.00	5321.0
135.90	1487.0
137.00	2302.0
137.80	545.0
138.90	331.0
139.90	750.0
141.00	7134.0
141.90	2504.0
142.90	1985.0
143.90	423.0
144.90	552.0
145.90	1311.0
147.00	3783.0
147.90	8235.0
149.00	1794.0
149.90	611.0

Instrument :
BNA_N
ClientSampleId :
DFTPP08

151.00	1230.0
152.00	830.0
153.00	2415.0
154.00	1689.0
155.00	4369.0
156.00	6045.0
156.90	1124.0
157.90	1412.0
158.90	1256.0
159.90	2659.0
161.00	3989.0
162.00	983.0
163.80	502.0
164.90	2871.0
165.90	2132.0
167.00	13675.0
168.00	6908.0
168.90	1665.0
169.90	581.0
170.80	575.0
171.90	1355.0
172.90	2151.0
174.00	3363.0
175.00	6274.0
176.00	1529.0
176.90	2556.0
177.90	837.0
178.90	11437.0
180.00	8449.0
181.00	3736.0
181.90	578.0
182.80	313.0
183.90	1008.0
185.00	5592.0
186.00	45736.0
187.00	13343.0
187.90	1209.0
188.90	2953.0
190.00	478.0
191.00	1717.0
192.00	3856.0
193.00	4200.0
194.00	764.0
195.00	459.0
196.00	9908.0
196.60	1725.0
196.90	1630.0
197.90	342528.0
198.90	23184.0
199.90	1411.0
201.50	1428.0
202.90	2074.0
204.00	11400.0
205.00	20112.0
206.00	87856.0
207.00	12418.0
208.00	3053.0
208.90	884.0
210.00	1150.0
211.00	3034.0
211.80	653.0
215.00	874.0
216.10	2238.0
216.90	24392.0
218.00	3266.0
221.00	13558.0
221.80	3185.0
223.00	5877.0
224.00	48232.0
225.00	12740.0
226.00	1183.0
226.90	19704.0
228.00	2336.0
228.90	4834.0
230.00	454.0
231.00	1675.0
233.00	329.0
233.90	1187.0
234.90	1749.0
235.90	1127.0
236.90	1454.0
238.90	946.0
240.00	637.0
240.90	1290.0
242.00	2487.0
243.00	2740.0
244.00	36864.0
245.00	4423.0
246.00	8061.0
246.90	1642.0
248.00	736.0
249.00	1528.0
251.90	416.0
253.00	534.0
254.10	1544.0

Instrument :
BNA_N
ClientSampleId :
DFTPP08

255.00	208064.0
256.00	29224.0
257.00	2097.0
258.00	9566.0
258.90	1464.0
260.90	475.0
264.90	4181.0
266.00	652.0
270.90	548.0
272.10	437.0
273.00	5505.0
273.90	15778.0
275.00	88928.0
276.00	11498.0
277.00	7151.0
277.90	1177.0
279.00	308.0
281.00	360.0
282.90	907.0
284.00	734.0
285.00	1375.0
285.80	327.0
288.90	424.0
290.00	314.0
291.90	313.0
293.00	1515.0
293.90	770.0
294.90	585.0
296.00	24936.0
297.00	3365.0
301.10	314.0
302.00	333.0
303.00	3324.0
304.00	825.0
307.90	421.0
310.00	383.0
314.00	1012.0
314.90	2355.0
316.00	1720.0
317.00	308.0
321.00	690.0
322.00	603.0
323.00	7500.0
324.00	1436.0
326.80	1296.0
327.90	948.0
332.00	689.0
333.10	547.0
334.00	5824.0
335.00	1286.0
341.00	1036.0
345.90	1886.0
347.00	369.0
352.00	2686.0
353.00	1678.0
354.00	2711.0
355.00	610.0
365.00	11966.0
365.90	1380.0
370.90	692.0
372.00	4125.0
373.10	952.0
383.00	1253.0
389.90	855.0
390.90	301.0
391.80	337.0
402.00	1440.0
402.90	2804.0
403.90	1132.0
420.90	2361.0
421.90	2230.0
423.00	15863.0
424.10	3468.0
425.00	317.0
441.00	49784.0
442.00	327616.0
443.00	61368.0
444.00	5903.0
444.80	360.0

Instrument :
BNA_N
ClientSampleId :
DFTPP08

m/z	Abundance
37.00	310.0
38.00	775.0
39.00	4426.0
39.90	1065.0
44.00	4850.0
49.00	628.0
50.00	16225.0
51.00	67176.0
52.00	3236.0
55.00	467.0
55.90	1934.0
57.00	5028.0
57.90	317.0
61.00	819.0
62.00	972.0
63.00	2271.0
64.10	328.0
65.00	1535.0
68.00	999.0
69.00	68032.0
73.10	640.0
74.00	7646.0
75.00	10308.0
76.00	3164.0
77.00	73368.0
78.00	4314.0
79.00	4955.0
80.00	4121.0
81.00	5341.0
81.90	1431.0
83.00	1477.0
84.90	664.0
85.90	1492.0
87.00	648.0
91.00	1385.0
91.90	1275.0
93.00	9253.0
93.90	644.0
96.00	520.0
98.00	6856.0
99.00	5957.0
99.90	466.0
100.90	3167.0
102.90	887.0
104.00	1949.0
104.90	1903.0
105.90	577.0
107.00	23912.0
108.00	3287.0
109.00	766.0
110.00	44488.0
110.90	5791.0
112.00	1002.0
116.00	1282.0
117.00	16560.0
118.00	1296.0
121.90	1142.0
122.90	3158.0
124.00	1180.0
125.00	973.0
127.00	85456.0
128.00	5853.0
129.00	33960.0
130.00	2806.0
131.00	360.0
132.00	333.0
134.00	917.0
135.00	2919.0
135.90	1033.0
137.00	1324.0
138.00	399.0
139.90	519.0
140.90	3903.0
142.00	1541.0
143.00	1183.0
145.80	699.0
146.90	2149.0
148.00	4706.0
149.00	1088.0
151.10	531.0
151.70	465.0
152.90	1308.0
154.00	966.0
155.00	2841.0
156.00	4248.0
157.00	820.0
158.00	742.0
158.90	744.0
159.90	1525.0
160.90	1955.0
161.90	615.0

Instrument :
BNA_N
ClientSampleId :
DFTPP08

165.00	1866.0
165.90	1403.0
167.00	8776.0
168.00	4373.0
168.90	705.0
171.00	384.0
171.90	716.0
173.00	1075.0
173.90	1888.0
175.00	2939.0
176.00	1148.0
177.00	1605.0
178.00	537.0
178.90	7616.0
180.00	4963.0
180.90	2508.0
181.90	425.0
182.90	336.0
184.00	493.0
185.00	3437.0
186.00	26032.0
187.00	7534.0
187.90	1012.0
189.00	2040.0
190.90	855.0
192.00	2488.0
192.90	2865.0
193.90	654.0
194.80	426.0
196.00	6172.0
197.00	1030.0
197.90	197888.0
198.90	12474.0
199.90	1216.0
201.40	992.0
202.90	1175.0
204.00	7126.0
205.00	12096.0
206.00	51912.0
207.00	7864.0
207.90	1880.0
208.90	764.0
209.90	770.0
210.90	1865.0
211.90	378.0
215.00	536.0
216.00	1356.0
216.90	14644.0
218.00	1613.0
219.10	341.0
221.00	7556.0
221.80	2148.0
223.00	3456.0
224.00	29808.0
225.00	6814.0
225.90	851.0
227.00	12587.0
227.90	1894.0
229.00	2495.0
229.80	341.0
230.90	1081.0
233.80	827.0
235.00	816.0
236.00	514.0
237.00	942.0
239.00	407.0
241.00	635.0
242.00	1520.0
242.90	1426.0
244.00	23456.0
245.10	2924.0
245.90	5171.0
246.80	839.0
249.00	1012.0
253.00	484.0
254.00	833.0
255.00	127504.0
256.00	18224.0
256.90	1579.0
257.90	6260.0
258.90	1283.0
264.90	2523.0
266.00	338.0
270.90	424.0
271.90	478.0
272.90	3930.0
274.00	10631.0
275.00	60344.0
276.00	6543.0
276.90	4871.0
278.00	670.0
279.00	314.0
283.00	501.0
283.90	486.0
285.00	1182.0

Instrument :
BNA_N
ClientSampleId :
DFTPP08

292.90	1233.0
296.00	16968.0
296.90	2010.0
301.90	415.0
303.00	1854.0
303.90	629.0
313.90	722.0
314.90	1852.0
316.00	850.0
321.00	584.0
323.00	5737.0
324.00	1090.0
326.90	1106.0
328.00	601.0
331.90	481.0
333.00	760.0
334.00	3580.0
335.00	1068.0
341.00	923.0
345.90	1276.0
346.90	315.0
352.10	1920.0
353.00	1380.0
353.90	2232.0
355.10	321.0
358.90	311.0
364.90	8453.0
365.90	881.0
370.90	483.0
372.00	3535.0
373.00	907.0
382.90	1089.0
390.10	317.0
391.10	369.0
401.80	1103.0
403.00	2262.0
404.00	776.0
421.00	2182.0
422.00	1811.0
423.00	12540.0
423.90	3014.0
441.00	39304.0
442.00	255808.0
443.00	47184.0
444.00	4412.0

Instrument :
BNA_N
ClientSampleId :
DFTPP08

Scan 2461 (17.463 min): BN002501.D
DFTPP08

m/z	Abundance
40.00	1128.0
44.00	4260.0
48.90	415.0
206.90	679.0

Instrument :
BNA_N
ClientSampleId :
DFTPP08