

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN082419\
 Data File : BN007375.D
 Acq On : 24 Aug 2019 22:06
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 24 23:43:54 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 24 00:38:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	169#	0.00
2	1,4-Dioxane	0.665	0.609	8.4	149	0.00
3	n-Nitrosodimethylamine	0.309	0.285	7.8	150#	0.00
4 S	2-Fluorophenol	1.490	1.355	9.1	162#	0.00
5 S	Phenol-d6	1.953	1.687	13.6	168#	0.00
6	bis(2-Chloroethyl)ether	1.445	1.463	-1.2	173#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	171#	-0.01
8 S	Nitrobenzene-d5	0.348	0.340	2.3	174#	0.00
9	Naphthalene	1.143	1.186	-3.8	173#	0.00
10	Hexachlorobutadiene	0.234	0.243	-3.8	170#	0.00
11 SURR	2-Methylnaphthalene-d10	0.665	0.694	-4.4	176#	0.00
12	2-Methylnaphthalene	0.779	0.809	-3.9	172#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	170#	0.00
14 S	2,4,6-Tribromophenol	0.209	0.233	-11.5	193#	0.00
15 S	2-Fluorobiphenyl	1.613	1.713	-6.2	177#	0.00
16	Acenaphthylene	2.367	2.442	-3.2	173#	0.00
17	Acenaphthene	1.385	1.453	-4.9	174#	0.00
18	Fluorene	1.751	1.843	-5.3	175#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	164#	0.00
20	4-Bromophenyl-phenylether	0.158	0.150	5.1	180#	0.00
21	Hexachlorobenzene	0.268	0.299	-11.6	180#	0.00
22	Pentachlorophenol	0.083	0.112	-34.9#	236#	0.00
23	Phenanthrene	1.203	1.298	-7.9	176#	0.00
24	Anthracene	1.210	1.268	-4.8	170#	0.00
25 SURR	Fluoranthene-d10	1.288	1.331	-3.3	164#	0.00
26	Fluoranthene	1.545	1.632	-5.6	168#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	163#	0.00
28	Pvrene	1.609	1.696	-5.4	168#	0.00
29 S	Terphenyl-d14	1.047	1.080	-3.2	162#	0.00
30	Benzo(a)anthracene	1.571	1.646	-4.8	168#	0.00
31	Chrysene	1.516	1.608	-6.1	165#	-0.01
32	Bis(2-ethylhexyl)phthalate	1.177	0.999	15.1	146	0.00
33	Indeno(1,2,3-cd)pyrene	1.827	1.926	-5.4	166#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	162#	0.00
35	Benzo(b)fluoranthene	1.449	1.543	-6.5	168#	0.00
36	Benzo(k)fluoranthene	1.432	1.536	-7.3	167#	0.00
37 C	Benzo(a)pyrene	1.379	1.426	-3.4	162#	0.00
38	Dibenzo(a,h)anthracene	1.376	1.421	-3.3	162#	-0.01
39	Benzo(a,h,i)perylene	1.371	1.426	-4.0	163#	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN082419\
Data File : BN007375.D
Acq On : 24 Aug 2019 22:06
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 24 23:43:54 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 24 00:38:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			