

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN082419\
 Data File : BN007375.D
 Acq On : 24 Aug 2019 22:06
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 24 23:43:54 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 24 00:38:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	169	0.00
2	1,4-Dioxane	0.400	0.366	8.5	149	0.00
3	n-Nitrosodimethylamine	0.400	0.369	7.8	150	0.00
4 S	2-Fluorophenol	0.400	0.390	2.5	162	0.00
5 S	Phenol-d6	0.400	0.381	4.8	168	0.00
6	bis(2-Chloroethyl)ether	0.400	0.405	-1.3	173	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	171	-0.01
8 S	Nitrobenzene-d5	0.400	0.391	2.3	174	0.00
9	Naphthalene	0.400	0.415	-3.7	173	0.00
10	Hexachlorobutadiene	0.400	0.416	-4.0	170	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.418	-4.5	176	0.00
12	2-Methylnaphthalene	0.400	0.416	-4.0	172	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	170	0.00
14 S	2,4,6-Tribromophenol	0.400	0.445	-11.2	193	0.00
15 S	2-Fluorobiphenyl	0.400	0.425	-6.2	177	0.00
16	Acenaphthylene	0.400	0.413	-3.2	173	0.00
17	Acenaphthene	0.400	0.420	-5.0	174	0.00
18	Fluorene	0.400	0.421	-5.2	175	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	164	0.00
20	4-Bromophenyl-phenylether	0.400	0.382	4.5	180	0.00
21	Hexachlorobenzene	0.400	0.446	-11.5	180	0.00
22	Pentachlorophenol	0.400	0.540	-35.0#	236	0.00
23	Phenanthrene	0.400	0.431	-7.7	176	0.00
24	Anthracene	0.400	0.419	-4.7	170	0.00
25 SURR	Fluoranthene-d10	0.400	0.413	-3.2	164	0.00
26	Fluoranthene	0.400	0.423	-5.7	168	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	163	0.00
28	Pvrene	0.400	0.422	-5.5	168	0.00
29 S	Terphenyl-d14	0.400	0.413	-3.2	162	0.00
30	Benzo(a)anthracene	0.400	0.419	-4.7	168	0.00
31	Chrysene	0.400	0.424	-6.0	165	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.374	6.5	146	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.422	-5.5	166	0.00
34 I	Perylene-d12	0.400	0.400	0.0	162	0.00
35	Benzo(b)fluoranthene	0.400	0.426	-6.5	168	0.00
36	Benzo(k)fluoranthene	0.400	0.429	-7.2	167	0.00
37 C	Benzo(a)pyrene	0.400	0.414	-3.5	162	0.00
38	Dibenzo(a,h)anthracene	0.400	0.413	-3.2	162	-0.01
39	Benzo(a,h,i)perylene	0.400	0.416	-4.0	163	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN082419\
Data File : BN007375.D
Acq On : 24 Aug 2019 22:06
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 24 23:43:54 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 24 00:38:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				