

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN082419\
 Data File : BN007377.D
 Acq On : 25 Aug 2019 00:32
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 25 05:27:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 24 00:38:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	170#	0.00
2	1,4-Dioxane	0.665	0.621	6.6	153#	0.00
3	n-Nitrosodimethylamine	0.309	0.267	13.6	142	0.00
4 S	2-Fluorophenol	1.490	1.330	10.7	161#	0.00
5 S	Phenol-d6	1.953	1.620	17.1	162#	0.00
6	bis(2-Chloroethyl)ether	1.445	1.445	0.0	172#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	169#	-0.01
8 S	Nitrobenzene-d5	0.348	0.335	3.7	170#	0.00
9	Naphthalene	1.143	1.172	-2.5	169#	0.00
10	Hexachlorobutadiene	0.234	0.244	-4.3	169#	0.00
11 SURR	2-Methylnaphthalene-d10	0.665	0.673	-1.2	169#	0.00
12	2-Methylnaphthalene	0.779	0.788	-1.2	166#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	158#	0.00
14 S	2,4,6-Tribromophenol	0.209	0.206	1.4	158#	0.00
15 S	2-Fluorobiphenyl	1.613	1.749	-8.4	168#	0.00
16	Acenaphthylene	2.367	2.334	1.4	154#	0.00
17	Acenaphthene	1.385	1.430	-3.2	160#	0.00
18	Fluorene	1.751	1.762	-0.6	156#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	149	0.00
20	4-Bromophenyl-phenylether	0.158	0.162	-2.5	177#	0.00
21	Hexachlorobenzene	0.268	0.302	-12.7	165#	0.00
22	Pentachlorophenol	0.083	0.093	-12.0	178#	0.00
23	Phenanthrene	1.203	1.255	-4.3	154#	0.00
24	Anthracene	1.210	1.177	2.7	143	0.00
25 SURR	Fluoranthene-d10	1.288	1.286	0.2	144	0.00
26	Fluoranthene	1.545	1.553	-0.5	145	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	143	0.00
28	Pvrene	1.609	1.659	-3.1	144	0.00
29 S	Terphenyl-d14	1.047	1.077	-2.9	143	0.00
30	Benzo(a)anthracene	1.571	1.624	-3.4	146	0.00
31	Chrysene	1.516	1.612	-6.3	146	-0.01
32	Bis(2-ethylhexyl)phthalate	1.177	0.900	23.5	116	-0.01
33	Indeno(1,2,3-cd)pyrene	1.827	1.961	-7.3	149	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	146	0.00
35	Benzo(b)fluoranthene	1.449	1.513	-4.4	149	0.00
36	Benzo(k)fluoranthene	1.432	1.523	-6.4	150	0.00
37 C	Benzo(a)pyrene	1.379	1.411	-2.3	145	-0.01
38	Dibenzo(a,h)anthracene	1.376	1.419	-3.1	146	-0.02
39	Benzo(a,h,i)perylene	1.371	1.432	-4.4	148	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN082419\
Data File : BN007377.D
Acq On : 25 Aug 2019 00:32
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 25 05:27:42 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 24 00:38:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			