

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN082419\
 Data File : BN007388.D
 Acq On : 25 Aug 2019 07:17
 Operator : HP/JU
 Sample : SSTDC0.4
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC0.4

Quant Time: Aug 26 05:03:03 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 24 00:38:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	198#	0.00
2	1,4-Dioxane	0.665	0.453	31.9#	130	0.00
3	n-Nitrosodimethylamine	0.309	0.179	42.1#	110	0.00
4 S	2-Fluorophenol	1.490	0.909	39.0#	127	0.00
5 S	Phenol-d6	1.953	1.112	43.1#	129	0.00
6	bis(2-Chloroethyl)ether	1.445	0.992	31.3#	137	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	195#	-0.01
8 S	Nitrobenzene-d5	0.348	0.241	30.7#	141	0.00
9	Naphthalene	1.143	0.841	26.4#	140	0.00
10	Hexachlorobutadiene	0.234	0.176	24.8	141	0.00
11 SURR	2-Methylnaphthalene-d10	0.665	0.474	28.7#	137	0.00
12	2-Methylnaphthalene	0.779	0.550	29.4#	134	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	181#	-0.01
14 S	2,4,6-Tribromophenol	0.209	0.139	33.5#	122	-0.01
15 S	2-Fluorobiphenyl	1.613	1.215	24.7	133	0.00
16	Acenaphthylene	2.367	1.850	21.8	139	0.00
17	Acenaphthene	1.385	1.000	27.8#	128	0.00
18	Fluorene	1.751	1.271	27.4#	128	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	168#	0.00
20	4-Bromophenyl-phenylether	0.158	0.063	60.1#	77	0.00
21	Hexachlorobenzene	0.268	0.205	23.5	127	0.00
22	Pentachlorophenol	0.083	0.067	19.3	144	-0.01
23	Phenanthrene	1.203	0.874	27.3#	122	0.00
24	Anthracene	1.210	0.826	31.7#	114	0.00
25 SURR	Fluoranthene-d10	1.288	0.951	26.2#	120	0.00
26	Fluoranthene	1.545	1.157	25.1#	122	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	180#	0.00
28	Pvrene	1.609	1.129	29.8#	123	0.00
29 S	Terphenyl-d14	1.047	0.737	29.6#	122	0.00
30	Benzo(a)anthracene	1.571	1.132	27.9#	127	-0.01
31	Chrysene	1.516	1.114	26.5#	127	-0.01
32	Bis(2-ethylhexyl)phthalate	1.177	0.656	44.3#	106	0.00
33	Indeno(1,2,3-cd)pyrene	1.827	1.334	27.0#	127	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	181#	0.00
35	Benzo(b)fluoranthene	1.449	1.079	25.5#	131	0.00
36	Benzo(k)fluoranthene	1.432	1.028	28.2#	125	0.00
37 C	Benzo(a)pyrene	1.379	1.000	27.5#	127	-0.01
38	Dibenzo(a,h)anthracene	1.376	0.967	29.7#	123	-0.01
39	Benzo(a,h,i)perylene	1.371	1.011	26.3#	129	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN082419\
Data File : BN007388.D
Acq On : 25 Aug 2019 07:17
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 26 05:03:03 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 24 00:38:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 1			