

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082419\
 Data File : BN007384.D
 Acq On : 25 Aug 2019 04:54
 Operator : HP/JU
 Sample : K4441-12
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 S700-N-1.0-1.5-082019

Manual Integrations
 APPROVED

mohammad
 8/31/2019 2:40:32 PM

Quant Time: Aug 25 05:39:59 2019

Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Sat Aug 24 00:38:17 2019

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	7230	0.40	ng	0.00
7) Naphthalene-d8	10.15	136	28479	0.40	ng	-0.01
13) Acenaphthene-d10	14.04	164	13700	0.40	ng	-0.01
19) Phenanthrene-d10	16.80	188	28061	0.40	ng	0.00
27) Chrysene-d12	21.02	240	32615	0.40	ng	0.00
34) Perylene-d12	23.12	264	38322	0.40	ng	0.00

System Monitoring Compounds

4) 2-Fluorophenol	5.04	112	5221	0.19	ng	0.00
5) Phenol-d6	6.59	99	6769	0.20	ng	0.00
8) Nitrobenzene-d5	8.53	82	5041	0.20	ng	0.00
11) 2-Methylnaphthalene-d10	11.76	152	9743	0.21	ng	0.00
14) 2,4,6-Tribromophenol	15.55	330	1501	0.21	ng	-0.01
15) 2-Fluorobiphenyl	12.66	172	11748	0.21	ng	0.00
25) Fluoranthene-d10	18.84	212	16711	0.18	ng	0.00
29) Terphenyl-d14	19.46	244	12756	0.15	ng	0.00

Target Compounds

						Qvalue
9) Naphthalene	10.20	128	22123	0.272	ng	# 90
12) 2-Methylnaphthalene	11.83	142	10152	0.183	ng	97
16) Acenaphthylene	13.76	152	7191	0.089	ng	# 96
17) Acenaphthene	14.11	154	77265	1.629	ng	98
18) Fluorene	15.10	166	75456	1.258	ng	99
23) Phenanthrene	16.84	178	1255711	14.877	ng	99
24) Anthracene	16.93	178	233744	2.753	ng	# 93
26) Fluoranthene	18.87	202	2384766	22.005	ng	99
28) Pyrene	19.24	202	2106576	16.056	ng	# 96
30) Benzo(a)anthracene	21.01	228	1808985	14.123	ng	98
31) Chrysene	21.05	228	1533509	12.404	ng	98
32) Bis(2-ethylhexyl)phthalate	20.96	149	18681	0.206	ng	# 88
33) Indeno(1,2,3-cd)pyrene	25.18	276	1022049	6.862	ng	97
35) Benzo(b)fluoranthene	22.51	252	2404040	17.314	ng	# 94
36) Benzo(k)fluoranthene	22.54	252	859107m	6.260	ng	
37) Benzo(a)pyrene	23.03	252	1684092	12.746	ng	# 92
38) Dibenzo(a,h)anthracene	25.19	278	307280	2.331	ng	# 96
39) Benzo(a,h,i)perylene	25.81	276	1016060	7.735	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082419\
Data File : BN007384.D
Acq On : 25 Aug 2019 04:54
Operator : HP/JU
Sample : K4441-12
Misc :
ALS Vial : 25 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
S700-N-1.0-1.5-082019

Manual Integrations
APPROVED

mohammad
8/31/2019 2:40:32 PM

Quant Time: Aug 25 05:39:59 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 24 00:38:17 2019
Response via : Initial Calibration





#1


1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.40 min Scan# 568
Delta R.T. 0.00 min
Lab File: BN007384.D
Acq: 25 Aug 2019 04:54

Instrument :
BNA_N
ClientSampleId :
S700-N-1.0-1.5-082019

Tot Ion	Ratio	Lower	Upper
152	100		
150	152.4	109.1	163.7
115	67.9	59.6	89.4

Manual Integrations
APPROVED

mohammad
8/31/2019 2:40:32 PM






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E


Ion 115.00 (114.70 to 115.70): E



#4

2-Fluorophenol
Concen: 0.194 ng
RT: 5.04 min Scan# 274
Delta R.T. 0.00 min
Lab File: BN007384.D
Acq: 25 Aug 2019 04:54

Tgt Ion	Ratio	Lower	Upper
112	100		
64	53.1	42.3	63.5
63	28.3	23.6	35.4








Abundance

Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BNC

Ion 63.00 (62.70 to 63.70): BNC





#5

Phenol-d6

Concen: 0.197 ng

RT: 6.59 min Scan# 467

Delta R.T. 0.00 min

Lab File: BN007384.D

Acq: 25 Aug 2019 04:54

Tot Ion:	99	Resp:	6769
Ion	Ratio	Lower	Upper
99	100		
42	16.2	11.2	16.8
71	34.2	31.8	47.8

Instrument :





BNA_N

ClientSampleId :

S700-N-1.0-1.5-082019

Manual Integrations
APPROVED

mohammad
8/31/2019 2:40:32 PM



#7

Naphthalene-d8

Concen: 0.400 ng


RT: 10.15 min Scan# 826


Delta R.T. -0.01 min

Lab File: BN007384.D

Acq: 25 Aug 2019 04:54

Tgt Ion:	136	Resp:	28479
Ion	Ratio	Lower	Upper
136	100		
137	11.3	12.9	19.3#
54	5.3	8.6	12.8#
68	6.7	17.0	25.4#






#8
Nitrobenzene-d5
Concen: 0.203 ng
RT: 8.53 min Scan# 698
Delta R.T. 0.00 min
Lab File: BN007384.D
Acq: 25 Aug 2019 04:54

Instrument :
BNA_N
ClientSampleId :
S700-N-1.0-1.5-082019




Tot Ion	Ratio	Lower	Upper
82	100		
128	35.9	19.7	29.5#
54	36.7	25.8	38.6

Manual Integrations
APPROVED

mohammad
8/31/2019 2:40:32 PM




Abundance Ion 82.00 (81.70 to 82.70): BN0
Ion 128.00 (127.70 to 128.70): E
Ion 54.00 (53.70 to 54.70): BN0






#9
Naphthalene
Concen: 0.272 ng
RT: 10.20 min Scan# 830
Delta R.T. 0.00 min
Lab File: BN007384.D
Acq: 25 Aug 2019 04:54

Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.6	11.1	16.7
127	13.6	15.5	23.3#



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E









#11
 2-Methylnaphthalene-d10
 Concen: 0.206 ng
 RT: 11.76 min Scan# 989
 Delta R.T. -0.00 min
 Lab File: BN007384.D
 Acq: 25 Aug 2019 04:54

Instrument :
 BNA_N
 ClientSampleId :
 S700-N-1.0-1.5-082019

Tot Ion	Ratio	Lower	Upper
152	100		
151	19.2	16.2	24.2




Manual Integrations
 APPROVED


mohammad
 8/31/2019 2:40:32 PM



#12
 2-Methylnaphthalene
 Concen: 0.183 ng
 RT: 11.83 min Scan# 1006
 Delta R.T. -0.00 min
 Lab File: BN007384.D
 Acq: 25 Aug 2019 04:54

Tgt Ion	Ratio	Lower	Upper
142	100		
141	89.0	72.8	109.2
115	41.2	35.2	52.8





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.04 min Scan# 1295

Delta R.T. -0.01 min

Lab File: BN007384.D

Acq: 25 Aug 2019 04:54

Instrument :

BNA_N


ClientSampleId :

S700-N-1.0-1.5-082019

Tot Ion:	164	Resp:	13700
Ion Ratio	Lower	Upper	
164	100		
162	105.6	83.5	125.3
160	48.3	38.2	57.4

Manual Integrations
APPROVED

mohammad
8/31/2019 2:40:32 PM






Abundance

Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#14

2,4,6-Tribromophenol

Concen: 0.209 ng


RT: 15.55 min Scan# 1430

Delta R.T. -0.01 min

Lab File: BN007384.D

Acq: 25 Aug 2019 04:54

Tgt Ion:	330	Resp:	1501
Ion Ratio	Lower	Upper	
330	100		
332	0.0	0.0	0.0
141	47.1	26.9	40.3#






Abundance

Ion 330.00 (329.70 to 330.70): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#15
2-Fluorobiphenyl
Concen: 0.213 ng
RT: 12.66 min Scan# 1171
Delta R.T. 0.00 min
Lab File: BN007384.D
Acq: 25 Aug 2019 04:54

Instrument :
BNA_N
ClientSampleId :
S700-N-1.0-1.5-082019

Tot Ion	Ratio	Lower	Upper
172	100		
171	35.3	30.3	45.5
170	24.1	20.7	31.1




Manual Integrations
APPROVED


mohammad
8/31/2019 2:40:32 PM



#16
Acenaphthylene
Concen: 0.089 ng
RT: 13.76 min Scan# 1270
Delta R.T. 0.00 min
Lab File: BN007384.D
Acq: 25 Aug 2019 04:54

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.5	16.1	24.1
153	16.4	10.6	15.8#





#17

Acenaphthene

Concen: 1.629 ng

RT: 14.11 min Scan# 1301

Delta R.T. -0.00 min

Lab File: BN007384.D

Acq: 25 Aug 2019 04:54

Instrument :

BNA_N


ClientSampleId :

S700-N-1.0-1.5-082019

Tot Ion:	154	Resp:	77265
Ion Ratio	Lower	Upper	
154	100		
153	113.6	92.8	139.2
152	54.4	43.8	65.8

Manual Integrations
APPROVED

mohammad
8/31/2019 2:40:32 PM






Abundance

Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#18

Fluorene

Concen: 1.258 ng


RT: 15.10 min Scan# 1390

Delta R.T. -0.01 min

Lab File: BN007384.D

Acq: 25 Aug 2019 04:54

Tgt Ion:	166	Resp:	75456
Ion Ratio	Lower	Upper	
166	100		
165	97.9	79.0	118.4
167	14.3	10.6	15.8






Abundance

Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.80 min Scan# 1546

Delta R.T. 0.00 min

Lab File: BN007384.D

Acq: 25 Aug 2019 04:54

Instrument :

BNA_N





ClientSampleId :

S700-N-1.0-1.5-082019

Tot Ion:	188	Resp:	28061
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	7.5	11.8	17.8#

Manual Integrations
APPROVED

mohammad
8/31/2019 2:40:32 PM



#23

Phenanthrene

Concen: 14.877 ng




RT: 16.84 min Scan# 1550


Delta R.T. 0.00 min

Lab File: BN007384.D

Acq: 25 Aug 2019 04:54

Tgt Ion:	178	Resp:	1255711
Ion Ratio	Lower	Upper	
178	100		
176	18.7	15.3	22.9
179	15.5	12.6	19.0





#24

Anthracene

Concen: 2.753 ng

RT: 16.93 min Scan# 1558

Delta R.T. 0.00 min

Lab File: BN007384.D

Acq: 25 Aug 2019 04:54

Instrument :

BNA_N





ClientSampleId :

S700-N-1.0-1.5-082019

Tot Ion:	178	Resp:	233744
Ion Ratio	Lower	Upper	
178	100		
176	19.1	14.9	22.3
179	21.4	12.4	18.6

Manual Integrations
APPROVED

mohammad
8/31/2019 2:40:32 PM



#25

Fluoranthene-d10

Concen: 0.185 ng




RT: 18.84 min Scan# 1827


Delta R.T. -0.00 min

Lab File: BN007384.D

Acq: 25 Aug 2019 04:54

Tgt Ion:	212	Resp:	16711
Ion Ratio	Lower	Upper	
212	100		
106	10.9	9.0	13.4
104	7.2	5.4	8.2





#26

Fluoranthene

Concen: 22.005 ng

RT: 18.87 min Scan# 1833

Delta R.T. -0.00 min

Lab File: BN007384.D

Acq: 25 Aug 2019 04:54

Instrument :

BNA_N

ClientSampleId :

S700-N-1.0-1.5-082019

Tot Ion:202 Resp: 2384766

Ion Ratio Lower Upper


202 100

101 9.5 7.7 11.5




203 17.7 13.7 20.5

Manual Integrations
APPROVED

mohammad
8/31/2019 2:40:32 PM



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E



#27

Chrysene-d12

Concen: 0.400 ng

RT: 21.02 min Scan# 2157

Delta R.T. 0.00 min

Lab File: BN007384.D

Acq: 25 Aug 2019 04:54


Tgt Ion:240 Resp: 32615

Ion Ratio Lower Upper



240 100


120 9.1 9.6 14.4#

236 29.7 22.2 33.2



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E









#28
Pvrene
Concen: 16.056 ng
RT: 19.24 min Scan# 1907
Delta R.T. 0.00 min
Lab File: BN007384.D
Acq: 25 Aug 2019 04:54

Instrument :
BNA_N
ClientSampleId :
S700-N-1.0-1.5-082019

Tot Ion	Ratio	Lower	Upper
202	100		
200	20.8	16.6	25.0
203	22.1	14.2	21.4




Manual Integrations
APPROVED


mohammad
8/31/2019 2:40:32 PM



#29
Terphenyl-d14
Concen: 0.149 ng
RT: 19.46 min Scan# 1952
Delta R.T. -0.00 min
Lab File: BN007384.D
Acq: 25 Aug 2019 04:54

Tgt Ion	Ratio	Lower	Upper
244	100		
212	9.2	7.0	10.6
122	12.8	9.6	14.4





#30

Benzo(a)anthracene

Concen: 14.123 ng

RT: 21.01 min Scan# 2156

Delta R.T. 0.00 min

Lab File: BN007384.D

Acq: 25 Aug 2019 04:54

Instrument :

BNA_N

ClientSampleId :

S700-N-1.0-1.5-082019

Tot Ion:228 Resp: 1808985

Ion Ratio Lower Upper


228 100

226 26.3 21.3 31.9




229 21.6 16.1 24.1

Manual Integrations
APPROVED

mohammad
8/31/2019 2:40:32 PM



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E



#31

Chrysene

Concen: 12.404 ng

RT: 21.05 min Scan# 2160

Delta R.T. -0.01 min

Lab File: BN007384.D

Acq: 25 Aug 2019 04:54

Tgt Ion:228 Resp: 1533509




Ion Ratio Lower Upper


228 100

226 30.0 23.5 35.3

229 22.4 16.6 24.8

Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E






#32
 Bis(2-ethylhexyl)phthalate
 Concen: 0.206 ng
 RT: 20.96 min Scan# 2152
 Delta R.T. -0.01 min
 Lab File: BN007384.D
 Acq: 25 Aug 2019 04:54

Instrument :
 BNA_N
 ClientSampleId :
 S700-N-1.0-1.5-082019

Tot Ion	Ratio	Lower	Upper
149	100		
167	21.2	22.8	34.2#
279	3.6	3.5	5.3

Manual Integrations
 APPROVED

mohammad
 8/31/2019 2:40:32 PM






Abundance

Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#33
 Indeno(1,2,3-cd)pyrene
 Concen: 6.862 ng
 RT: 25.18 min Scan# 3099
 Delta R.T. -0.01 min
 Lab File: BN007384.D
 Acq: 25 Aug 2019 04:54




Tgt Ion	Ratio	Lower	Upper
276	100		
138	16.6	15.1	22.7
277	24.9	19.7	29.5


Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#34
 Pervlene-d12
 Concen: 0.400 ng
 RT: 23.12 min Scan# 2496
 Delta R.T. -0.01 min
 Lab File: BN007384.D
 Acq: 25 Aug 2019 04:54

Instrument :
 BNA_N
 ClientSampleId :
 S700-N-1.0-1.5-082019

Tot Ion	Ratio	Lower	Upper
264	100		
260	26.0	19.7	29.5
265	58.1	29.9	44.9




Manual Integrations
 APPROVED


mohammad
 8/31/2019 2:40:32 PM



#35
 Benzo(b)fluoranthene
 Concen: 17.314 ng
 RT: 22.51 min Scan# 2317
 Delta R.T. 0.00 min
 Lab File: BN007384.D
 Acq: 25 Aug 2019 04:54

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.7	19.2	28.8
125	8.2	10.6	16.0









#36
Benzo(k)fluoranthene
Concen: 6.260 ng m
RT: 22.54 min Scan# 2327
Delta R.T. -0.01 min
Lab File: BN007384.D
Acq: 25 Aug 2019 04:54

Instrument :
BNA_N
ClientSampleId :
S700-N-1.0-1.5-082019

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.3	19.4	29.2
125	8.2	13.1	19.7#




Manual Integrations
APPROVED


mohammad
8/31/2019 2:40:32 PM



#37
Benzo(a)pyrene
Concen: 12.746 ng
RT: 23.03 min Scan# 2470
Delta R.T. -0.01 min
Lab File: BN007384.D
Acq: 25 Aug 2019 04:54

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.5	19.5	29.3
125	8.6	12.1	18.1#









#38
Dibenzo(a,h)anthracene
Concen: 2.331 ng
RT: 25.19 min Scan# 3101
Delta R.T. -0.02 min
Lab File: BN007384.D
Acq: 25 Aug 2019 04:54

Instrument :
BNA_N
ClientSampleId :
S700-N-1.0-1.5-082019

Tot Ion	Ratio	Lower	Upper
278	100		
139	13.2	13.7	20.5#
279	25.6	20.2	30.4

Manual Integrations
APPROVED

mohammad
8/31/2019 2:40:32 PM



#39
Benzo(g,h,i)perylene
Concen: 7.735 ng
RT: 25.81 min Scan# 3283
Delta R.T. -0.00 min
Lab File: BN007384.D
Acq: 25 Aug 2019 04:54

Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.9	19.6	29.4
138	16.4	16.3	24.5

