


Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN082419\
 Data File : BN007376.D
 Acq On : 24 Aug 2019 23:19
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082319.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sat Aug 24 00:38:17 2019



AutoFind: Scans 2455, 2456, 2457; Background Corrected with Scan 2448

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	24.4	100136	PASS
68	69	0.00	2	1.3	2009	PASS
69	198	0.00	100	38.7	158953	PASS
70	69	0.00	2	0.6	985	PASS
127	198	10	80	49.5	203050	PASS
197	198	0.00	2	0.2	995	PASS
198	198	100	100	100.0	410474	PASS
199	198	5	9	6.9	28461	PASS
275	198	10	60	27.6	113160	PASS
365	198	1	100	3.5	14441	PASS
441	198	0.01	100	11.9	49032	PASS
442	198	50	100	77.8	319338	PASS
443	442	15	24	19.1	61074	PASS