

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082420\
 Data File : BN011607.D
 Acq On : 24 Aug 2020 21:11
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN082420

Quant Time: Aug 25 03:39:47 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 25 03:31:11 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.620	0.627	-1.1	99	0.00
3	n-Nitrosodimethylamine	0.653	0.646	1.1	88	0.00
4 S	2-Fluorophenol	1.279	1.260	1.5	92	0.00
5 S	Phenol-d6	1.808	1.819	-0.6	91	0.00
6	bis(2-Chloroethyl)ether	1.285	1.283	0.2	94	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
8 S	Nitrobenzene-d5	0.346	0.336	2.9	96	-0.01
9	Naphthalene	1.137	1.060	6.8	95	0.00
10	Hexachlorobutadiene	0.238	0.222	6.7	95	-0.01
11 SURR	2-Methylnaphthalene-d10	0.757	0.701	7.4	96	0.00
12	2-Methylnaphthalene	0.839	0.790	5.8	96	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
14 S	2,4,6-Tribromophenol	0.220	0.213	3.2	94	0.00
15 S	2-Fluorobiphenyl	1.557	1.470	5.6	94	-0.01
16	Acenaphthylene	2.108	1.927	8.6	95	0.00
17	Acenaphthene	1.337	1.246	6.8	95	0.00
18	Fluorene	1.730	1.621	6.3	97	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
20	4,6-Dinitro-2-methylphenol	0.046	0.042	8.7	107	-0.01
21	4-Bromophenyl-phenylether	0.233	0.214	8.2	98	0.00
22	Hexachlorobenzene	0.270	0.252	6.7	96	0.00
23	Atrazine	0.230	0.162	29.6#	98	0.00
24	Pentachlorophenol	0.119	0.109	8.4	95	0.00
25	Phenanthrene	1.243	1.143	8.0	97	0.00
26	Anthracene	1.224	1.118	8.7	95	0.00
27 SURR	Fluoranthene-d10	1.304	1.178	9.7	98	0.00
28	Fluoranthene	1.506	1.387	7.9	96	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
30	Pyrene	1.418	1.314	7.3	95	0.00
31 S	Terphenyl-d14	1.022	0.966	5.5	95	0.00
32	Benzo(a)anthracene	1.443	1.355	6.1	95	0.00
33	Chrysene	1.375	1.296	5.7	100	0.00
34	Bis(2-ethylhexyl)phthalate	1.027	0.847	17.5	89	0.00
35	Indeno(1,2,3-cd)pyrene	1.679	1.545	8.0	94	0.00
36 I	Perylene-d12	1.000	1.000	0.0	96	0.00
37	Benzo(b)fluoranthene	1.449	1.362	6.0	96	0.00
38	Benzo(k)fluoranthene	1.441	1.361	5.6	95	0.00
39 C	Benzo(a)pyrene	1.351	1.264	6.4	95	0.00
40	Dibenzo(a,h)anthracene	1.412	1.318	6.7	95	0.00
41	Benzo(g,h,i)perylene	1.370	1.276	6.9	94	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082420\
Data File : BN011607.D
Acq On : 24 Aug 2020 21:11
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN082420

Quant Time: Aug 25 03:39:47 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 25 03:31:11 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0