

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082420\
 Data File : BN011607.D
 Acq On : 24 Aug 2020 21:11
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN082420

Quant Time: Aug 25 03:39:47 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 25 03:31:11 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	95	0.00
2	1,4-Dioxane	0.400	0.404	-1.0	99	0.00
3	n-Nitrosodimethylamine	0.400	0.396	1.0	88	0.00
4 S	2-Fluorophenol	0.400	0.394	1.5	92	0.00
5 S	Phenol-d6	0.400	0.402	-0.5	91	0.00
6	bis(2-Chloroethyl)ether	0.400	0.399	0.3	94	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	96	0.00
8 S	Nitrobenzene-d5	0.400	0.388	3.0	96	-0.01
9	Naphthalene	0.400	0.373	6.8	95	0.00
10	Hexachlorobutadiene	0.400	0.373	6.8	95	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.370	7.5	96	0.00
12	2-Methylnaphthalene	0.400	0.377	5.8	96	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	95	0.00
14 S	2,4,6-Tribromophenol	0.400	0.387	3.3	94	0.00
15 S	2-Fluorobiphenyl	0.400	0.378	5.5	94	-0.01
16	Acenaphthylene	0.400	0.366	8.5	95	0.00
17	Acenaphthene	0.400	0.373	6.8	95	0.00
18	Fluorene	0.400	0.375	6.3	97	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	99	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.361	9.8	107	-0.01
21	4-Bromophenyl-phenylether	0.400	0.366	8.5	98	0.00
22	Hexachlorobenzene	0.400	0.373	6.8	96	0.00
23	Atrazine	0.400	0.281	29.8#	98	0.00
24	Pentachlorophenol	0.400	0.365	8.8	95	0.00
25	Phenanthrene	0.400	0.368	8.0	97	0.00
26	Anthracene	0.400	0.365	8.8	95	0.00
27 SURR	Fluoranthene-d10	0.400	0.361	9.8	98	0.00
28	Fluoranthene	0.400	0.368	8.0	96	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	97	0.00
30	Pyrene	0.400	0.370	7.5	95	0.00
31 S	Terphenyl-d14	0.400	0.378	5.5	95	0.00
32	Benzo(a)anthracene	0.400	0.376	6.0	95	0.00
33	Chrysene	0.400	0.377	5.8	100	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.334	16.5	89	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.368	8.0	94	0.00
36 I	Perylene-d12	0.400	0.400	0.0	96	0.00
37	Benzo(b)fluoranthene	0.400	0.376	6.0	96	0.00
38	Benzo(k)fluoranthene	0.400	0.378	5.5	95	0.00
39 C	Benzo(a)pyrene	0.400	0.374	6.5	95	0.00
40	Dibenzo(a,h)anthracene	0.400	0.373	6.8	95	0.00
41	Benzo(g,h,i)perylene	0.400	0.372	7.0	94	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082420\
Data File : BN011607.D
Acq On : 24 Aug 2020 21:11
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN082420

Quant Time: Aug 25 03:39:47 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 25 03:31:11 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0