

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN082519\
 Data File : BN007390.D
 Acq On : 25 Aug 2019 11:50
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 26 05:45:32 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 24 00:38:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
2	1,4-Dioxane	0.665	0.613	7.8	107	0.00
3	n-Nitrosodimethylamine	0.309	0.260	15.9	98	0.00
4 S	2-Fluorophenol	1.490	1.181	20.7	101	0.00
5 S	Phenol-d6	1.953	1.441	26.2#	102	0.00
6	bis(2-Chloroethyl)ether	1.445	1.317	8.9	111	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	124	-0.01
8 S	Nitrobenzene-d5	0.348	0.312	10.3	116	0.00
9	Naphthalene	1.143	1.107	3.1	117	0.00
10	Hexachlorobutadiene	0.234	0.222	5.1	113	0.00
11 SURR	2-Methylnaphthalene-d10	0.665	0.642	3.5	118	0.00
12	2-Methylnaphthalene	0.779	0.747	4.1	116	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	128	0.00
14 S	2,4,6-Tribromophenol	0.209	0.200	4.3	125	0.00
15 S	2-Fluorobiphenyl	1.613	1.559	3.3	121	0.00
16	Acenaphthylene	2.367	2.133	9.9	114	0.00
17	Acenaphthene	1.385	1.322	4.5	120	0.00
18	Fluorene	1.751	1.683	3.9	121	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	134	0.00
20	4-Bromophenyl-phenylether	0.158	0.136	13.9	134	0.00
21	Hexachlorobenzene	0.268	0.271	-1.1	134	0.00
22	Pentachlorophenol	0.083	0.077	7.2	132	0.00
23	Phenanthrene	1.203	1.157	3.8	129	0.00
24	Anthracene	1.210	1.091	9.8	120	0.00
25 SURR	Fluoranthene-d10	1.288	1.262	2.0	127	0.00
26	Fluoranthene	1.545	1.542	0.2	130	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	145	0.00
28	Pvrene	1.609	1.463	9.1	129	0.00
29 S	Terphenyl-d14	1.047	0.970	7.4	130	0.00
30	Benzo(a)anthracene	1.571	1.452	7.6	132	0.00
31	Chrysene	1.516	1.486	2.0	137	0.00
32	Bis(2-ethylhexyl)phthalate	1.177	0.831	29.4#	109	0.00
33	Indeno(1,2,3-cd)pyrene	1.827	1.829	-0.1	140	0.00
34 I	Perylene-d12	1.000	1.000	0.0	149	0.00
35	Benzo(b)fluoranthene	1.449	1.389	4.1	139	0.00
36	Benzo(k)fluoranthene	1.432	1.422	0.7	142	0.00
37 C	Benzo(a)pyrene	1.379	1.311	4.9	137	0.00
38	Dibenzo(a,h)anthracene	1.376	1.329	3.4	139	0.00
39	Benzo(a,h,i)perylene	1.371	1.322	3.6	139	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN082519\
Data File : BN007390.D
Acq On : 25 Aug 2019 11:50
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 26 05:45:32 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 24 00:38:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			