

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN082519\
 Data File : BN007402.D
 Acq On : 25 Aug 2019 19:01
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 26 05:27:41 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 24 00:38:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
2	1,4-Dioxane	0.665	0.585	12.0	102	0.00
3	n-Nitrosodimethylamine	0.309	0.279	9.7	104	-0.02
4 S	2-Fluorophenol	1.490	1.236	17.0	105	0.00
5 S	Phenol-d6	1.953	1.576	19.3	111	0.00
6	bis(2-Chloroethyl)ether	1.445	1.334	7.7	112	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	120	-0.01
8 S	Nitrobenzene-d5	0.348	0.334	4.0	120	0.00
9	Naphthalene	1.143	1.113	2.6	114	0.00
10	Hexachlorobutadiene	0.234	0.219	6.4	108	0.00
11 SURR	2-Methylnaphthalene-d10	0.665	0.630	5.3	112	0.00
12	2-Methylnaphthalene	0.779	0.741	4.9	111	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
14 S	2,4,6-Tribromophenol	0.209	0.221	-5.7	127	0.00
15 S	2-Fluorobiphenyl	1.613	1.569	2.7	113	0.00
16	Acenaphthylene	2.367	2.209	6.7	109	0.00
17	Acenaphthene	1.385	1.323	4.5	111	0.00
18	Fluorene	1.751	1.673	4.5	111	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
20	4-Bromophenyl-phenylether	0.158	0.125	20.9	107	0.00
21	Hexachlorobenzene	0.268	0.269	-0.4	117	0.00
22	Pentachlorophenol	0.083	0.098	-18.1	148	0.00
23	Phenanthrene	1.203	1.185	1.5	116	0.00
24	Anthracene	1.210	1.115	7.9	108	0.00
25 SURR	Fluoranthene-d10	1.288	1.216	5.6	108	0.00
26	Fluoranthene	1.545	1.488	3.7	110	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	120	0.00
28	Pvrene	1.609	1.513	6.0	110	0.00
29 S	Terphenyl-d14	1.047	0.963	8.0	107	0.00
30	Benzo(a)anthracene	1.571	1.504	4.3	113	0.00
31	Chrysene	1.516	1.482	2.2	113	0.00
32	Bis(2-ethylhexyl)phthalate	1.177	0.970	17.6	105	0.00
33	Indeno(1,2,3-cd)pyrene	1.827	1.842	-0.8	117	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	123	0.00
35	Benzo(b)fluoranthene	1.449	1.377	5.0	114	0.00
36	Benzo(k)fluoranthene	1.432	1.409	1.6	117	0.00
37 C	Benzo(a)pyrene	1.379	1.322	4.1	115	0.00
38	Dibenzo(a,h)anthracene	1.376	1.323	3.9	114	0.00
39	Benzo(a,h,i)perylene	1.371	1.355	1.2	118	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN082519\
Data File : BN007402.D
Acq On : 25 Aug 2019 19:01
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 26 05:27:41 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 24 00:38:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			