

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082523\
 Data File : BN027151.D
 Acq On : 26 Aug 2023 12:19
 Operator : MA/JU
 Sample : 04095-06
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 FB01-20230818

Quant Time: Aug 26 22:33:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 26 02:13:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.081	152	5978	0.400	ng	0.00
7) Naphthalene-d8	10.906	136	13185	0.400	ng	0.00
13) Acenaphthene-d10	14.705	164	8758	0.400	ng	0.00
19) Phenanthrene-d10	17.455	188	20735	0.400	ng	# 0.00
29) Chrysene-d12	21.632	240	16907	0.400	ng	# 0.00
35) Perylene-d12	24.174	264	15449	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.596	112	481	0.029	ng	0.00
5) Phenol-d6	7.221	99	881	0.042	ng	0.00
8) Nitrobenzene-d5	9.273	82	3163	0.284	ng	0.00
11) 2-Methylnaphthalene-d10	12.479	152	7996	0.382	ng	0.00
14) 2,4,6-Tribromophenol	16.202	330	20	0.004	ng	0.00
15) 2-Fluorobiphenyl	13.336	172	10269	0.316	ng	0.00
27) Fluoranthene-d10	19.477	212	23464	0.441	ng	0.00
31) Terphenyl-d14	20.062	244	17363	0.491	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.489	149	2299	0.052	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082523\
Data File : BN027151.D
Acq On : 26 Aug 2023 12:19
Operator : MA/JU
Sample : 04095-06
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
FB01-20230818

Quant Time: Aug 26 22:33:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 26 02:13:59 2023
Response via : Initial Calibration

