

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082523\
 Data File : BN027185.D
 Acq On : 27 Aug 2023 11:34
 Operator : MA/JU
 Sample : 04051-11
 Misc :
 ALS Vial : 75 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE109D3-20230816

Quant Time: Aug 28 00:43:27 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 26 02:13:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.073	152	7160	0.400	ng	-0.01
7) Naphthalene-d8	10.895	136	16014	0.400	ng	#-0.01
13) Acenaphthene-d10	14.705	164	10384	0.400	ng	0.00
19) Phenanthrene-d10	17.455	188	23895	0.400	ng	0.00
29) Chrysene-d12	21.632	240	16843	0.400	ng	# 0.00
35) Perylene-d12	24.168	264	14530	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.596	112	2025	0.100	ng	0.00
5) Phenol-d6	7.214	99	1365	0.055	ng	0.00
8) Nitrobenzene-d5	9.262	82	3460	0.256	ng	-0.01
11) 2-Methylnaphthalene-d10	12.472	152	7550	0.297	ng	-0.01
14) 2,4,6-Tribromophenol	16.201	330	786	0.142	ng	0.00
15) 2-Fluorobiphenyl	13.336	172	11621	0.301	ng	0.00
27) Fluoranthene-d10	19.472	212	20340	0.332	ng	0.00
31) Terphenyl-d14	20.057	244	17427	0.495	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.458	88	18545	2.408	ng	# 97
34) Bis(2-ethylhexyl)phtha...	21.488	149	2077	0.047	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082523\
Data File : BN027185.D
Acq On : 27 Aug 2023 11:34
Operator : MA/JU
Sample : 04051-11
Misc :
ALS Vial : 75 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE109D3-20230816

Quant Time: Aug 28 00:43:27 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 26 02:13:59 2023
Response via : Initial Calibration

