

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082523\
 Data File : BN027186.D
 Acq On : 27 Aug 2023 12:10
 Operator : MA/JU
 Sample : 04051-10
 Misc :
 ALS Vial : 76 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE125D1-20230816

Quant Time: Aug 28 00:43:36 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 26 02:13:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.073	152	7214	0.400	ng	-0.01
7) Naphthalene-d8	10.895	136	14224	0.400	ng	#-0.01
13) Acenaphthene-d10	14.705	164	10595	0.400	ng	0.00
19) Phenanthrene-d10	17.455	188	24729	0.400	ng	0.00
29) Chrysene-d12	21.623	240	18427	0.400	ng	# 0.00
35) Perylene-d12	24.162	264	15439	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.596	112	2379	0.117	ng	0.00
5) Phenol-d6	7.214	99	1619	0.064	ng	0.00
8) Nitrobenzene-d5	9.262	82	4385	0.365	ng	-0.01
11) 2-Methylnaphthalene-d10	12.476	152	7645	0.338	ng	0.00
14) 2,4,6-Tribromophenol	16.202	330	1177	0.209	ng	0.00
15) 2-Fluorobiphenyl	13.336	172	12025	0.306	ng	0.00
27) Fluoranthene-d10	19.467	212	23538	0.371	ng	0.00
31) Terphenyl-d14	20.053	244	18059	0.469	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.466	88	39312	5.067	ng	98
34) Bis(2-ethylhexyl)phtha...	21.480	149	2535	0.053	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082523\
Data File : BN027186.D
Acq On : 27 Aug 2023 12:10
Operator : MA/JU
Sample : 04051-10
Misc :
ALS Vial : 76 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE125D1-20230816

Quant Time: Aug 28 00:43:36 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 26 02:13:59 2023
Response via : Initial Calibration

