

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082523\
 Data File : BN027205.D
 Acq On : 28 Aug 2023 00:57
 Operator : MA/JU
 Sample : 04051-07
 Misc :
 ALS Vial : 96 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE139D1-20230816

Quant Time: Aug 28 01:28:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 26 02:13:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.073	152	5691	0.400	ng	-0.01
7) Naphthalene-d8	10.895	136	13248	0.400	ng	#-0.01
13) Acenaphthene-d10	14.705	164	8382	0.400	ng	0.00
19) Phenanthrene-d10	17.443	188	18785	0.400	ng	#-0.01
29) Chrysene-d12	21.623	240	14209	0.400	ng	# 0.00
35) Perylene-d12	24.156	264	12131	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.589	112	1709	0.107	ng	-0.01
5) Phenol-d6	7.214	99	1095	0.055	ng	0.00
8) Nitrobenzene-d5	9.262	82	2987	0.267	ng	-0.01
11) 2-Methylnaphthalene-d10	12.472	152	6349	0.302	ng	-0.01
14) 2,4,6-Tribromophenol	16.189	330	863	0.193	ng	-0.01
15) 2-Fluorobiphenyl	13.326	172	9968	0.320	ng	-0.01
27) Fluoranthene-d10	19.467	212	18261	0.379	ng	0.00
31) Terphenyl-d14	20.053	244	14081	0.474	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.458	88	9227	1.508	ng	# 95
25) Phenanthrene	17.492	178	1502	0.028	ng	# 93
30) Pyrene	19.862	202	1464	0.027	ng	97
34) Bis(2-ethylhexyl)phtha...	21.480	149	2930	0.079	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082523\
Data File : BN027205.D
Acq On : 28 Aug 2023 00:57
Operator : MA/JU
Sample : 04051-07
Misc :
ALS Vial : 96 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE139D1-20230816

Quant Time: Aug 28 01:28:05 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 26 02:13:59 2023
Response via : Initial Calibration

