

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/25/2023	BNA_N	<u>BN027112.D</u>
Compound Name	Response	Retention Time
DDT	4204734	20.91
DDD	95892	20.469
DDE	1544	19.957
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
97436	4302170	2.26