

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/25/2023	BNA_N	<u>BN027121.D</u>
Compound Name	Response	Retention Time
DDT	3862224	20.91
DDD	56506	20.468
DDE	1236	19.957
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
57742	3919966	1.47