

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/25/2023	BNA_N	<u>BN027139.D</u>
Compound Name	Response	Retention Time
DDT	1906133	20.904
DDD	26914	20.463
DDE	929	19.951
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
27843	1933976	1.44