

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/25/2023	BNA_N	<u>BN027157.D</u>
Compound Name	Response	Retention Time
DDT	1469122	20.904
DDD	19330	20.463
DDE	1006	19.951
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
20336	1489458	1.37