

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/25/2023	BNA_N	<u>BN027175.D</u>
Compound Name	Response	Retention Time
DDT	1657131	20.898
DDD	21046	20.457
DDE	910	19.945
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
21956	1679087	1.31