

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082523\
 Data File : BN027174.D
 Acq On : 27 Aug 2023 04:18
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 64 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 27 05:02:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 26 02:13:59 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	72	0.00
2	1,4-Dioxane	0.430	0.461	-7.2	82	0.00
3	n-Nitrosodimethylamine	0.479	0.517	-7.9	80	0.00
4 S	2-Fluorophenol	1.127	0.947	16.0	63	0.00
5 S	Phenol-d6	1.392	1.154	17.1	63	0.00
6	bis(2-Chloroethyl)ether	1.171	1.220	-4.2	76	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	72	0.00
8 S	Nitrobenzene-d5	0.338	0.286	15.4	59	-0.01
9	Naphthalene	1.052	1.061	-0.9	75	-0.01
10	Hexachlorobutadiene	0.215	0.225	-4.7	75	-0.01
11 SURR	2-Methylnaphthalene-d10	0.636	0.597	6.1	68	0.00
12	2-Methylnaphthalene	0.823	0.781	5.1	68	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	62	0.00
14 S	2,4,6-Tribromophenol	0.213	0.125	41.3#	39#	0.00
15 S	2-Fluorobiphenyl	1.485	1.504	-1.3	65	0.00
16	Acenaphthylene	1.909	1.841	3.6	60	-0.01
17	Acenaphthene	1.179	1.197	-1.5	64	0.00
18	Fluorene	1.583	1.590	-0.4	62	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	57	0.00
20	4,6-Dinitro-2-methylphenol	0.006	0.006	0.0	53	-0.01
21	4-Bromophenyl-phenylether	0.228	0.205	10.1	52	0.00
22	Hexachlorobenzene	0.244	0.265	-8.6	63	0.00
23	Atrazine	0.203	0.147	27.6#	41#	-0.01
24	Pentachlorophenol	0.123	0.073	40.7#	36#	0.00
25	Phenanthrene	1.150	1.148	0.2	58	0.00
26	Anthracene	1.073	0.937	12.7	51	0.00
27 SURR	Fluoranthene-d10	1.026	0.940	8.4	54	0.00
28	Fluoranthene	1.362	1.301	4.5	56	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	48#	0.00
30	Pyrene	1.539	1.743	-13.3	55	0.00
31 S	Terphenyl-d14	0.836	0.795	4.9	48#	0.00
32	Benzo(a)anthracene	1.444	1.300	10.0	44#	0.00
33	Chrysene	1.457	1.531	-5.1	51	0.00
34	Bis(2-ethylhexyl)phthalate	1.040	0.643	38.2#	29#	0.00
35 I	Perylene-d12	1.000	1.000	0.0	39#	0.00
36	Indeno(1,2,3-cd)pyrene	1.638	1.401	14.5	34#	-0.01
37	Benzo(b)fluoranthene	1.339	1.517	-13.3	47#	0.00
38	Benzo(k)fluoranthene	1.376	1.464	-6.4	42#	0.00
39 C	Benzo(a)pyrene	1.341	1.371	-2.2	42#	0.00
40	Dibenzo(a,h)anthracene	1.304	1.099	15.7	34#	-0.01
41	Benzo(g,h,i)perylene	1.403	1.320	5.9	37#	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4