

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082619\
 Data File : BN007410.D
 Acq On : 26 Aug 2019 15:25
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 26 16:00:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 24 00:38:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	111	0.00
2	1,4-Dioxane	0.400	0.410	-2.5	109	0.00
3	n-Nitrosodimethylamine	0.400	0.348	13.0	93	0.00
4 S	2-Fluorophenol	0.400	0.344	14.0	95	0.00
5 S	Phenol-d6	0.400	0.320	20.0	94	0.00
6	bis(2-Chloroethyl)ether	0.400	0.379	5.3	106	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	112	-0.01
8 S	Nitrobenzene-d5	0.400	0.368	8.0	108	0.00
9	Naphthalene	0.400	0.382	4.5	105	0.00
10	Hexachlorobutadiene	0.400	0.370	7.5	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.385	3.8	107	0.00
12	2-Methylnaphthalene	0.400	0.383	4.3	104	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	119	0.00
14 S	2,4,6-Tribromophenol	0.400	0.417	-4.2	126	0.00
15 S	2-Fluorobiphenyl	0.400	0.365	8.8	106	0.00
16	Acenaphthylene	0.400	0.350	12.5	103	0.00
17	Acenaphthene	0.400	0.372	7.0	108	0.00
18	Fluorene	0.400	0.391	2.3	113	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	126	0.00
20	4-Bromophenyl-phenylether	0.400	0.347	13.3	126	0.00
21	Hexachlorobenzene	0.400	0.388	3.0	121	0.00
22	Pentachlorophenol	0.400	0.384	4.0	129	0.00
23	Phenanthrene	0.400	0.388	3.0	122	0.00
24	Anthracene	0.400	0.359	10.3	112	0.00
25 SURR	Fluoranthene-d10	0.400	0.382	4.5	117	0.00
26	Fluoranthene	0.400	0.391	2.3	120	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	130	0.00
28	Pvrene	0.400	0.381	4.8	120	0.00
29 S	Terphenyl-d14	0.400	0.388	3.0	122	0.00
30	Benzo(a)anthracene	0.400	0.377	5.8	120	0.00
31	Chrysene	0.400	0.405	-1.3	126	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.343	14.2	107	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.421	-5.2	131	0.00
34 I	Perylene-d12	0.400	0.400	0.0	134	0.00
35	Benzo(b)fluoranthene	0.400	0.378	5.5	124	0.00
36	Benzo(k)fluoranthene	0.400	0.411	-2.7	132	0.00
37 C	Benzo(a)pyrene	0.400	0.388	3.0	126	0.00
38	Dibenzo(a,h)anthracene	0.400	0.400	0.0	130	0.00
39	Benzo(a,h,i)perylene	0.400	0.412	-3.0	134	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082619\
Data File : BN007410.D
Acq On : 26 Aug 2019 15:25
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 26 16:00:42 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 24 00:38:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				