

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082619\
 Data File : BN007412.D
 Acq On : 26 Aug 2019 16:37
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 27 12:35:32 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 24 00:38:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2	1,4-Dioxane	0.665	0.671	-0.9	114	0.00
3	n-Nitrosodimethylamine	0.309	0.248	19.7	91	0.00
4 S	2-Fluorophenol	1.490	1.189	20.2	99	0.00
5 S	Phenol-d6	1.953	1.410	27.8#	97	0.00
6	bis(2-Chloroethyl)ether	1.445	1.334	7.7	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	117	-0.01
8 S	Nitrobenzene-d5	0.348	0.311	10.6	109	0.00
9	Naphthalene	1.143	1.102	3.6	110	0.00
10	Hexachlorobutadiene	0.234	0.218	6.8	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.665	0.626	5.9	108	0.00
12	2-Methylnaphthalene	0.779	0.743	4.6	108	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	124	0.00
14 S	2,4,6-Tribromophenol	0.209	0.204	2.4	122	0.00
15 S	2-Fluorobiphenyl	1.613	1.467	9.1	110	0.00
16	Acenaphthylene	2.367	2.050	13.4	106	0.00
17	Acenaphthene	1.385	1.261	9.0	110	0.00
18	Fluorene	1.751	1.677	4.2	116	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	129	0.00
20	4-Bromophenyl-phenylether	0.158	0.138	12.7	130	0.00
21	Hexachlorobenzene	0.268	0.266	0.7	126	0.00
22	Pentachlorophenol	0.083	0.092	-10.8	152#	0.00
23	Phenanthrene	1.203	1.162	3.4	124	0.00
24	Anthracene	1.210	1.081	10.7	114	0.00
25 SURR	Fluoranthene-d10	1.288	1.225	4.9	118	0.00
26	Fluoranthene	1.545	1.528	1.1	124	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	133	0.00
28	Pvrene	1.609	1.532	4.8	124	0.00
29 S	Terphenyl-d14	1.047	1.026	2.0	126	0.00
30	Benzo(a)anthracene	1.571	1.471	6.4	122	0.00
31	Chrysene	1.516	1.492	1.6	125	0.00
32	Bis(2-ethylhexyl)phthalate	1.177	0.832	29.3#	99	0.00
33	Indeno(1,2,3-cd)pyrene	1.827	1.784	2.4	125	0.00
34 I	Perylene-d12	1.000	1.000	0.0	133	0.00
35	Benzo(b)fluoranthene	1.449	1.446	0.2	129	0.00
36	Benzo(k)fluoranthene	1.432	1.447	-1.0	129	0.00
37 C	Benzo(a)pyrene	1.379	1.323	4.1	123	0.00
38	Dibenzo(a,h)anthracene	1.376	1.308	4.9	122	0.00
39	Benzo(a,h,i)perylene	1.371	1.333	2.8	125	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082619\
Data File : BN007412.D
Acq On : 26 Aug 2019 16:37
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 27 12:35:32 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 24 00:38:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			