


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN082619\
 Data File : BN007403.D
 Acq On : 26 Aug 2019 08:38
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082319.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sat Aug 24 00:38:17 2019



AutoFind: Scans 2455, 2456, 2457; Background Corrected with Scan 2448

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	22.8	140210	PASS
68	69	0.00	2	1.8	4179	PASS
69	198	0.00	100	37.9	232648	PASS
70	69	0.00	2	0.4	857	PASS
127	198	10	80	49.0	300864	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	614549	PASS
199	198	5	9	6.7	41024	PASS
275	198	10	60	28.1	172736	PASS
365	198	1	100	3.4	20843	PASS
441	198	0.01	100	13.0	79858	PASS
442	198	50	100	83.9	515413	PASS
443	442	15	24	18.8	96933	PASS