

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082621\
 Data File : BN016075.D
 Acq On : 26 Aug 2021 16:54
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4259

Quant Time: Aug 26 17:27:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN082621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 26 11:35:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.867	152	1268	0.400	ng/ul	0.00
4) Naphthalene-d8	10.666	136	5360	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.502	164	3165	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.251	188	6675	0.400	ng/ul	0.00
17) Chrysene-d12	21.442	240	6720	0.400	ng/ul	0.01
23) Perylene-d12	23.833	264	5823	0.400	ng/ul	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.315	96	700	0.407	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.256	152	3239	0.401	ng/ul	0.00
18) Fluoranthene-d10	19.281	212	7318	0.404	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.353	88	682	0.393	ng/ul	94
5) Naphthalene	10.716	128	5926	0.428	ng/ul	100
7) 2-Methylnaphthalene	12.327	142	4000	0.427	ng/ul	99
8) 1-Methylnaphthalene	12.547	142	3664	0.400	ng/ul	98
10) Acenaphthylene	14.224	152	4676	0.418	ng/ul	99
11) Acenaphthene	14.566	153	4064	0.422	ng/ul	98
12) Fluorene	15.552	166	4752	0.416	ng/ul	99
14) Pentachlorophenol	16.900	266	524	0.375	ng/ul	96
15) Phenanthrene	17.289	178	7829	0.425	ng/ul	99
16) Anthracene	17.381	178	6739	0.424	ng/ul	99
19) Fluoranthene	19.308	202	8936	0.426	ng/ul	100
20) Pyrene	19.671	202	9077	0.429	ng/ul	99
21) Benzo(a)anthracene	21.425	228	8722	0.417	ng/ul	100
22) Chrysene	21.477	228	9636	0.428	ng/ul	99
24) Benzo(b)fluoranthene	23.102	252	9442	0.435	ng/ul	100
25) Benzo(k)fluoranthene	23.152	252	9498	0.434	ng/ul	100
26) Benzo(a)pyrene	23.728	252	7829	0.428	ng/ul	98
27) Indeno(1,2,3-cd)pyrene	26.300	276	10276	0.425	ng/ul#	58
28) Dibenzo(a,h)anthracene	26.324	278	8638	0.424	ng/ul	98
29) Benzo(g,h,i)perylene	27.061	276	9040	0.434	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082621\
Data File : BN016075.D
Acq On : 26 Aug 2021 16:54
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4259

Quant Time: Aug 26 17:27:14 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN082621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 26 11:35:43 2021
Response via : Initial Calibration

