

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082720\
 Data File : BN011677.D
 Acq On : 27 Aug 2020 23:32
 Operator : CG/JU
 Sample : L3820-06
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB181-GW-658-660

Quant Time: Aug 28 00:40:58 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 27 14:34:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	2027	0.40	ng	0.00
7) Naphthalene-d8	10.53	136	7714	0.40	ng	0.00
13) Acenaphthene-d10	14.38	164	4608	0.40	ng	0.00
19) Phenanthrene-d10	17.13	188	10311	0.40	ng	0.00
29) Chrysene-d12	21.32	240	12001	0.40	ng	0.00
36) Perylene-d12	23.58	264	12305	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.34	112	380	0.07	ng	0.00
5) Phenol-d6	6.91	99	339	0.04	ng	0.00
8) Nitrobenzene-d5	8.88	82	1741	0.23	ng	0.00
11) 2-Methylnaphthalene-d10	12.12	152	3065	0.22	ng	0.00
14) 2,4,6-Tribromophenol	15.87	330	349	0.16	ng	-0.01
15) 2-Fluorobiphenyl	13.00	172	4882	0.25	ng	0.00
27) Fluoranthene-d10	19.16	212	8066	0.24	ng	0.00
31) Terphenyl-d14	19.76	244	8115	0.27	ng	-0.01
Target Compounds						
30) Pyrene	19.55	202	926	0.022	ng	96
34) Bis(2-ethylhexyl)phthalate	21.25	149	5666	0.325	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082720\
Data File : BN011677.D
Acq On : 27 Aug 2020 23:32
Operator : CG/JU
Sample : L3820-06
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB181-GW-658-660

Quant Time: Aug 28 00:40:58 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 27 14:34:48 2020
Response via : Initial Calibration

