

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082720\
 Data File : BN011662.D
 Acq On : 27 Aug 2020 13:54
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN082720

Quant Time: Aug 27 14:38:52 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 27 14:34:48 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
2	1,4-Dioxane	0.643	0.627	2.5	85	0.00
3	n-Nitrosodimethylamine	0.743	0.708	4.7	76	0.00
4 S	2-Fluorophenol	1.130	1.077	4.7	80	0.00
5 S	Phenol-d6	1.535	1.436	6.4	80	0.00
6	bis(2-Chloroethyl)ether	1.249	1.225	1.9	83	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	79	0.00
8 S	Nitrobenzene-d5	0.387	0.355	8.3	79	0.01
9	Naphthalene	1.144	1.050	8.2	79	0.00
10	Hexachlorobutadiene	0.251	0.230	8.4	79	0.00
11 SURR	2-Methylnaphthalene-d10	0.721	0.652	9.6	79	0.00
12	2-Methylnaphthalene	0.814	0.732	10.1	78	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	78	0.00
14 S	2,4,6-Tribromophenol	0.194	0.160	17.5	73	0.00
15 S	2-Fluorobiphenyl	1.691	1.527	9.7	77	0.01
16	Acenaphthylene	1.948	1.679	13.8	76	0.00
17	Acenaphthene	1.345	1.200	10.8	78	0.00
18	Fluorene	1.724	1.501	12.9	76	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	78	0.00
20	4,6-Dinitro-2-methylphenol	0.062	0.050	19.4	74	0.00
21	4-Bromophenyl-phenylether	0.234	0.209	10.7	76	0.00
22	Hexachlorobenzene	0.289	0.264	8.7	78	0.01
23	Atrazine	0.206	0.178	13.6	75	0.00
24	Pentachlorophenol	0.115	0.101	12.2	75	0.01
25	Phenanthrene	1.292	1.166	9.8	78	0.01
26	Anthracene	1.148	1.006	12.4	76	0.00
27 SURR	Fluoranthene-d10	1.284	1.150	10.4	77	0.00
28	Fluoranthene	1.532	1.363	11.0	77	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	76	0.00
30	Pyrene	1.391	1.313	5.6	77	0.00
31 S	Terphenyl-d14	0.993	0.939	5.4	77	0.00
32	Benzo(a)anthracene	1.428	1.314	8.0	76	0.00
33	Chrysene	1.473	1.380	6.3	76	0.00
34	Bis(2-ethylhexyl)phthalate	0.581	0.502	13.6	73	0.00
35	Indeno(1,2,3-cd)pyrene	1.921	1.749	9.0	73	0.00
36 I	Perylene-d12	1.000	1.000	0.0	74	0.00
37	Benzo(b)fluoranthene	1.534	1.354	11.7	76	0.00
38	Benzo(k)fluoranthene	1.568	1.390	11.4	74	0.00
39 C	Benzo(a)pyrene	1.371	1.165	15.0	73	0.00
40	Dibenzo(a,h)anthracene	1.548	1.345	13.1	73	0.00
41	Benzo(g,h,i)perylene	1.516	1.308	13.7	73	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082720\
Data File : BN011662.D
Acq On : 27 Aug 2020 13:54
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN082720

Quant Time: Aug 27 14:38:52 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 27 14:34:48 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0