

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082720\
 Data File : BN011662.D
 Acq On : 27 Aug 2020 13:54
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN082720

Quant Time: Aug 27 14:38:52 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 27 14:34:48 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	79	0.00
2	1,4-Dioxane	0.400	0.390	2.5	85	0.00
3	n-Nitrosodimethylamine	0.400	0.381	4.8	76	0.00
4 S	2-Fluorophenol	0.400	0.381	4.8	80	0.00
5 S	Phenol-d6	0.400	0.374	6.5	80	0.00
6	bis(2-Chloroethyl)ether	0.400	0.392	2.0	83	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	79	0.00
8 S	Nitrobenzene-d5	0.400	0.366	8.5	79	0.01
9	Naphthalene	0.400	0.367	8.3	79	0.00
10	Hexachlorobutadiene	0.400	0.367	8.3	79	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.362	9.5	79	0.00
12	2-Methylnaphthalene	0.400	0.360	10.0	78	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	78	0.00
14 S	2,4,6-Tribromophenol	0.400	0.330	17.5	73	0.00
15 S	2-Fluorobiphenyl	0.400	0.361	9.8	77	0.01
16	Acenaphthylene	0.400	0.345	13.8	76	0.00
17	Acenaphthene	0.400	0.357	10.8	78	0.00
18	Fluorene	0.400	0.348	13.0	76	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	78	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.326	18.5	74	0.00
21	4-Bromophenyl-phenylether	0.400	0.357	10.8	76	0.00
22	Hexachlorobenzene	0.400	0.366	8.5	78	0.01
23	Atrazine	0.400	0.347	13.3	75	0.00
24	Pentachlorophenol	0.400	0.351	12.3	75	0.01
25	Phenanthrene	0.400	0.361	9.8	78	0.01
26	Anthracene	0.400	0.350	12.5	76	0.00
27 SURR	Fluoranthene-d10	0.400	0.358	10.5	77	0.00
28	Fluoranthene	0.400	0.356	11.0	77	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	76	0.00
30	Pyrene	0.400	0.378	5.5	77	0.00
31 S	Terphenyl-d14	0.400	0.378	5.5	77	0.00
32	Benzo(a)anthracene	0.400	0.368	8.0	76	0.00
33	Chrysene	0.400	0.375	6.3	76	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.346	13.5	73	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.364	9.0	73	0.00
36 I	Perylene-d12	0.400	0.400	0.0	74	0.00
37	Benzo(b)fluoranthene	0.400	0.353	11.8	76	0.00
38	Benzo(k)fluoranthene	0.400	0.355	11.3	74	0.00
39 C	Benzo(a)pyrene	0.400	0.340	15.0	73	0.00
40	Dibenzo(a,h)anthracene	0.400	0.348	13.0	73	0.00
41	Benzo(g,h,i)perylene	0.400	0.345	13.8	73	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082720\
Data File : BN011662.D
Acq On : 27 Aug 2020 13:54
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN082720

Quant Time: Aug 27 14:38:52 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 27 14:34:48 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)
----------	--------	-------	------	------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0