

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082720\
 Data File : BN011674.D
 Acq On : 27 Aug 2020 21:43
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 28 00:40:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 27 14:34:48 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
2	1,4-Dioxane	0.643	0.616	4.2	88	0.00
3	n-Nitrosodimethylamine	0.743	0.737	0.8	83	0.00
4 S	2-Fluorophenol	1.130	1.045	7.5	82	0.00
5 S	Phenol-d6	1.535	1.409	8.2	83	0.00
6	bis(2-Chloroethyl)ether	1.249	1.161	7.0	83	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	83	0.00
8 S	Nitrobenzene-d5	0.387	0.356	8.0	83	0.00
9	Naphthalene	1.144	1.047	8.5	82	0.00
10	Hexachlorobutadiene	0.251	0.238	5.2	86	0.00
11 SURR	2-Methylnaphthalene-d10	0.721	0.650	9.8	83	0.00
12	2-Methylnaphthalene	0.814	0.736	9.6	83	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	84	0.00
14 S	2,4,6-Tribromophenol	0.194	0.168	13.4	83	-0.01
15 S	2-Fluorobiphenyl	1.691	1.531	9.5	84	0.00
16	Acenaphthylene	1.948	1.701	12.7	83	0.00
17	Acenaphthene	1.345	1.208	10.2	85	0.00
18	Fluorene	1.724	1.519	11.9	83	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
20	4,6-Dinitro-2-methylphenol	0.062	0.056	9.7	90	-0.01
21	4-Bromophenyl-phenylether	0.234	0.209	10.7	85	0.00
22	Hexachlorobenzene	0.289	0.263	9.0	86	0.00
23	Atrazine	0.206	0.176	14.6	82	0.00
24	Pentachlorophenol	0.115	0.100	13.0	82	0.00
25	Phenanthrene	1.292	1.154	10.7	86	0.00
26	Anthracene	1.148	0.989	13.9	83	0.00
27 SURR	Fluoranthene-d10	1.284	1.119	12.9	84	0.00
28	Fluoranthene	1.532	1.352	11.7	85	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	83	0.00
30	Pyrene	1.391	1.331	4.3	85	0.00
31 S	Terphenyl-d14	0.993	0.956	3.7	85	0.00
32	Benzo(a)anthracene	1.428	1.307	8.5	82	-0.01
33	Chrysene	1.473	1.383	6.1	83	-0.01
34	Bis(2-ethylhexyl)phthalate	0.581	0.512	11.9	81	-0.01
35	Indeno(1,2,3-cd)pyrene	1.921	1.765	8.1	79	-0.01
36 I	Perylene-d12	1.000	1.000	0.0	80	-0.01
37	Benzo(b)fluoranthene	1.534	1.330	13.3	81	0.00
38	Benzo(k)fluoranthene	1.568	1.392	11.2	80	-0.01
39 C	Benzo(a)pyrene	1.371	1.168	14.8	80	-0.01
40	Dibenzo(a,h)anthracene	1.548	1.346	13.0	79	-0.01
41	Benzo(g,h,i)perylene	1.516	1.298	14.4	78	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082720\
Data File : BN011674.D
Acq On : 27 Aug 2020 21:43
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 28 00:40:39 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 27 14:34:48 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0