

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082819\
 Data File : BN007472.D
 Acq On : 28 Aug 2019 17:25
 Operator : HP/JU
 Sample : K4373-12DL 5X
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AB4DL

Manual Integrations
 APPROVED

mohammad
 8/29/2019 7:50:45 AM

Quant Time: Aug 28 18:02:56 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN082819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 27 20:30:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	4557	0.40	ng/ul	0.00
2) Naphthalene-d8	10.15	136	19705	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.05	164	10195	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.80	188	23733m	0.40	ng/ul	0.00
16) Chrysene-d12	21.02	240	18957m	0.40	ng/ul	0.00
20) Perylene-d12	23.12	264	20498	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.76	152	2328	0.08	ng/ul	0.00
14) Fluoranthene-d10	18.84	212	6240m	0.08	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.20	128	35015	0.629	ng/ul	99
5) 2-Methylnaphthalene	11.83	142	4048	0.108	ng/ul	97
7) Acenaphthylene	13.76	152	4038	0.071	ng/ul#	80
8) Acenaphthene	14.11	153	422448	10.364	ng/ul	97
9) Fluorene	15.11	166	32014	0.726	ng/ul	96
12) Phenanthrene	16.84	178	24349m	0.329	ng/ul	
13) Anthracene	16.93	178	38189m	0.547	ng/ul	
15) Fluoranthene	18.87	202	52954m	0.534	ng/ul	
17) Pyrene	19.24	202	38265	0.413	ng/ul	98
18) Benzo(a)anthracene	21.00	228	10930m	0.129	ng/ul	
19) Chrysene	21.05	228	8844m	0.097	ng/ul	
21) Benzo(b)fluoranthene	22.50	252	10219m	0.114	ng/ul	
22) Benzo(k)fluoranthene	22.54	252	3870m	0.041	ng/ul	
23) Benzo(a)pyrene	23.03	252	6198	0.069	ng/ul#	74
24) Indeno(1,2,3-cd)pyrene	25.18	276	3256m	0.033	ng/ul	
26) Benzo(a,h,i)perylene	25.81	276	2988	0.035	ng/ul#	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082819\
Data File : BN007472.D
Acq On : 28 Aug 2019 17:25
Operator : HP/JU
Sample : K4373-12DL 5X
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AB4DL

Manual Integrations
APPROVED

mohammad
8/29/2019 7:50:45 AM

Quant Time: Aug 28 18:02:56 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN082819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 27 20:30:31 2019
Response via : Initial Calibration

