

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082819\
 Data File : BN007474.D
 Acq On : 28 Aug 2019 18:36
 Operator : HP/JU
 Sample : K4373-16DL 5X
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AD6DL

Quant Time: Aug 28 23:08:33 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN082819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 27 20:30:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	3316	0.40	ng/ul	0.00
2) Naphthalene-d8	10.15	136	14237	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.05	164	7628	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.80	188	17939	0.40	ng/ul	0.00
16) Chrysene-d12	21.02	240	14624	0.40	ng/ul	0.00
20) Perylene-d12	23.12	264	17395	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.76	152	1634	0.07	ng/ul	0.00
14) Fluoranthene-d10	18.84	212	4275	0.07	ng/ul	0.00
Target Compounds						
3) Naphthalene	10.20	128	16277	0.405	ng/ul	99
5) 2-Methylnaphthalene	11.83	142	1175	0.044	ng/ul	98
8) Acenaphthene	14.11	153	838	0.027	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082819\
Data File : BN007474.D
Acq On : 28 Aug 2019 18:36
Operator : HP/JU
Sample : K4373-16DL 5X
Misc :
ALS Vial : 47 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AD6DL

Quant Time: Aug 28 23:08:33 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN082819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 27 20:30:31 2019
Response via : Initial Calibration

