

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082819\
 Data File : BN007485.D
 Acq On : 29 Aug 2019 01:47
 Operator : HP/JU
 Sample : K4395-03
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AE0

Manual Integrations
 APPROVED

mohammad
 8/29/2019 7:51:11 AM

Quant Time: Aug 29 03:41:11 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN082819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 29 03:32:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	5158	0.40	ng/ul	0.00
2) Naphthalene-d8	10.15	136	23007	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.05	164	12087	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.80	188	28384m	0.40	ng/ul	0.00
16) Chrysene-d12	21.01	240	21943m	0.40	ng/ul	0.00
20) Perylene-d12	23.11	264	20121m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.75	152	8596	0.24	ng/ul	0.00
14) Fluoranthene-d10	18.84	212	26175m	0.28	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.20	128	4122	0.063	ng/ul#	85
7) Acenaphthylene	13.76	152	4739	0.070	ng/ul#	71
8) Acenaphthene	14.11	153	259841	5.377	ng/ul	98
9) Fluorene	15.10	166	7947	0.152	ng/ul#	19
12) Phenanthrene	16.84	178	4039m	0.046	ng/ul	
13) Anthracene	16.93	178	29224m	0.350	ng/ul	
15) Fluoranthene	18.87	202	166009m	1.400	ng/ul	
17) Pyrene	19.24	202	125709	1.171	ng/ul	97
18) Benzo(a)anthracene	21.00	228	3509m	0.036	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082819\
Data File : BN007485.D
Acq On : 29 Aug 2019 01:47
Operator : HP/JU
Sample : K4395-03
Misc :
ALS Vial : 61 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AE0

Manual Integrations
APPROVED

mohammad
8/29/2019 7:51:11 AM

Quant Time: Aug 29 03:41:11 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN082819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 03:32:53 2019
Response via : Initial Calibration

