

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082819\
 Data File : BN007467.D
 Acq On : 28 Aug 2019 14:25
 Operator : HP/JU
 Sample : K4373-08
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AH6

Manual Integrations
 APPROVED

mohammad
 8/29/2019 7:50:18 AM

Quant Time: Aug 28 15:03:50 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN082819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 27 20:30:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.40	152	4703	0.40	ng/ul	0.00
2) Naphthalene-d8	10.15	136	20866	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.05	164	10727	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.80	188	24043m	0.40	ng/ul	0.00
16) Chrysene-d12	21.02	240	19606m	0.40	ng/ul	0.00
20) Perylene-d12	23.12	264	17110m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.76	152	9722	0.30	ng/ul	0.00
14) Fluoranthene-d10	18.84	212	22848m	0.29	ng/ul	0.00
Target Compounds						
3) Naphthalene	10.20	128	3165	0.054	ng/ul#	Ovalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082819\
Data File : BN007467.D
Acq On : 28 Aug 2019 14:25
Operator : HP/JU
Sample : K4373-08
Misc :
ALS Vial : 40 Sample Multiplier: 1




Instrument :
BNA_N
ClientSampleId :
C0AH6

Manual Integrations
APPROVED

mohammad
8/29/2019 7:50:18 AM

Quant Time: Aug 28 15:03:50 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN082819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 27 20:30:31 2019
Response via : Initial Calibration





#3

Naphthalene

Concen: 0.054 ng/ul

RT: 10.20 min Scan# 2287

Delta R.T. -0.01 min

Lab File: BN007467.D

Acq: 28 Aug 2019 14:25

Instrument :

BNA_N

ClientSampleId :

C0AH6

Tot Ion	Ratio	Lower	Upper
128	100		
129	17.7	9.2	13.8#
127	16.6	10.4	15.6#

Manual Integrations
APPROVED

mohammad
8/29/2019 7:50:18 AM

