

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN082823\  
 Data File : BN027209.D  
 Acq On : 28 Aug 2023 10:02  
 Operator : MA/JU  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 LabSampleId :  
 SSTDCCC0.4

Quant Time: Aug 29 03:22:47 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN081123.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Aug 29 03:21:06 2023  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|         | Compound                   | Amount | Calc. | %Dev  | Area% | Dev(min) |
|---------|----------------------------|--------|-------|-------|-------|----------|
| 1 I     | 1,4-Dichlorobenzene-d4     | 0.400  | 0.400 | 0.0   | 70    | 0.00     |
| 2       | 1,4-Dioxane                | 0.400  | 0.452 | -13.0 | 85    | 0.00     |
| 3       | n-Nitrosodimethylamine     | 0.400  | 0.445 | -11.2 | 81    | 0.00     |
| 4 S     | 2-Fluorophenol             | 0.400  | 0.339 | 15.3  | 62    | 0.00     |
| 5 S     | Phenol-d6                  | 0.400  | 0.334 | 16.5  | 62    | 0.00     |
| 6       | bis(2-Chloroethyl)ether    | 0.400  | 0.422 | -5.5  | 75    | 0.00     |
| 7 I     | Naphthalene-d8             | 0.400  | 0.400 | 0.0   | 67    | 0.00     |
| 8 S     | Nitrobenzene-d5            | 0.400  | 0.390 | 2.5   | 64    | 0.00     |
| 9       | Naphthalene                | 0.400  | 0.409 | -2.2  | 72    | 0.00     |
| 10      | Hexachlorobutadiene        | 0.400  | 0.429 | -7.2  | 72    | 0.00     |
| 11 SURR | 2-Methylnaphthalene-d10    | 0.400  | 0.392 | 2.0   | 67    | 0.00     |
| 12      | 2-Methylnaphthalene        | 0.400  | 0.400 | 0.0   | 68    | 0.00     |
| 13 I    | Acenaphthene-d10           | 0.400  | 0.400 | 0.0   | 60    | 0.00     |
| 14 S    | 2,4,6-Tribromophenol       | 0.400  | 0.238 | 40.5# | 38    | 0.00     |
| 15 S    | 2-Fluorobiphenyl           | 0.400  | 0.407 | -1.7  | 63    | 0.00     |
| 16      | Acenaphthylene             | 0.400  | 0.399 | 0.3   | 60    | 0.00     |
| 17      | Acenaphthene               | 0.400  | 0.421 | -5.2  | 64    | 0.00     |
| 18      | Fluorene                   | 0.400  | 0.417 | -4.2  | 62    | 0.00     |
| 19 I    | Phenanthrene-d10           | 0.400  | 0.400 | 0.0   | 57    | 0.00     |
| 20      | 4,6-Dinitro-2-methylphenol | 0.400  | 0.348 | 13.0  | 48    | 0.00     |
| 21      | 4-Bromophenyl-phenylether  | 0.400  | 0.361 | 9.8   | 52    | 0.00     |
| 22      | Hexachlorobenzene          | 0.400  | 0.426 | -6.5  | 62    | 0.00     |
| 23      | Atrazine                   | 0.400  | 0.283 | 29.3# | 40    | 0.00     |
| 24      | Pentachlorophenol          | 0.400  | 0.259 | 35.3# | 39    | 0.00     |
| 25      | Phenanthrene               | 0.400  | 0.400 | 0.0   | 58    | 0.00     |
| 26      | Anthracene                 | 0.400  | 0.350 | 12.5  | 51    | 0.00     |
| 27 SURR | Fluoranthene-d10           | 0.400  | 0.365 | 8.8   | 53    | 0.00     |
| 28      | Fluoranthene               | 0.400  | 0.382 | 4.5   | 56    | 0.00     |
| 29 I    | Chrysene-d12               | 0.400  | 0.400 | 0.0   | 46    | 0.00     |
| 30      | Pyrene                     | 0.400  | 0.480 | -20.0 | 55    | 0.00     |
| 31 S    | Terphenyl-d14              | 0.400  | 0.390 | 2.5   | 47    | 0.00     |
| 32      | Benzo(a)anthracene         | 0.400  | 0.354 | 11.5  | 41    | 0.00     |
| 33      | Chrysene                   | 0.400  | 0.423 | -5.7  | 48    | 0.00     |
| 34      | Bis(2-ethylhexyl)phthalate | 0.400  | 0.233 | 41.8# | 26    | 0.00     |
| 35 I    | Perylene-d12               | 0.400  | 0.400 | 0.0   | 33    | 0.00     |
| 36      | Indeno(1,2,3-cd)pyrene     | 0.400  | 0.321 | 19.8  | 28    | 0.00     |
| 37      | Benzo(b)fluoranthene       | 0.400  | 0.462 | -15.5 | 41    | 0.00     |
| 38      | Benzo(k)fluoranthene       | 0.400  | 0.451 | -12.7 | 39    | 0.00     |
| 39 C    | Benzo(a)pyrene             | 0.400  | 0.395 | 1.3   | 35    | 0.00     |
| 40      | Dibenzo(a,h)anthracene     | 0.400  | 0.312 | 22.0  | 27    | 0.00     |
| 41      | Benzo(g,h,i)perylene       | 0.400  | 0.371 | 7.3   | 32    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

**Instrument :**  
BNA\_N  
**LabSampleId :**  
SSTDCCC0.4