

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082823\
 Data File : BN027219.D
 Acq On : 28 Aug 2023 16:12
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 29 03:25:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 29 03:21:06 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	64	0.00
2	1,4-Dioxane	0.430	0.486	-13.0	77	0.00
3	n-Nitrosodimethylamine	0.479	0.559	-16.7	77	0.00
4 S	2-Fluorophenol	1.127	0.979	13.1	58	0.00
5 S	Phenol-d6	1.392	1.186	14.8	57	0.00
6	bis(2-Chloroethyl)ether	1.171	1.226	-4.7	68	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	66	0.00
8 S	Nitrobenzene-d5	0.338	0.271	19.8	51	0.00
9	Naphthalene	1.052	1.093	-3.9	71	0.00
10	Hexachlorobutadiene	0.215	0.216	-0.5	66	0.00
11 SURR	2-Methylnaphthalene-d10	0.636	0.579	9.0	61	0.00
12	2-Methylnaphthalene	0.823	0.762	7.4	61	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	54	0.01
14 S	2,4,6-Tribromophenol	0.213	0.127	40.4#	35#	0.00
15 S	2-Fluorobiphenyl	1.485	1.489	-0.3	56	0.00
16	Acenaphthylene	1.909	1.885	1.3	54	0.00
17	Acenaphthene	1.179	1.220	-3.5	57	0.00
18	Fluorene	1.583	1.613	-1.9	55	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	51	0.00
20	4,6-Dinitro-2-methylphenol	0.006	0.007	-16.7	52	0.00
21	4-Bromophenyl-phenylether	0.228	0.199	12.7	45#	0.00
22	Hexachlorobenzene	0.244	0.258	-5.7	54	0.00
23	Atrazine	0.203	0.144	29.1#	36#	0.00
24	Pentachlorophenol	0.123	0.076	38.2#	34#	0.00
25	Phenanthrene	1.150	1.127	2.0	51	0.00
26	Anthracene	1.073	0.930	13.3	45#	0.00
27 SURR	Fluoranthene-d10	1.026	0.917	10.6	47#	0.00
28	Fluoranthene	1.362	1.280	6.0	49#	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	40#	0.00
30	Pyrene	1.539	1.821	-18.3	48#	0.00
31 S	Terphenyl-d14	0.836	0.812	2.9	41#	0.00
32	Benzo(a)anthracene	1.444	1.285	11.0	36#	0.00
33	Chrysene	1.457	1.571	-7.8	43#	0.00
34	Bis(2-ethylhexyl)phthalate	1.040	0.657	36.8#	25#	0.00
35 I	Perylene-d12	1.000	1.000	0.0	30#	0.00
36	Indeno(1,2,3-cd)pyrene	1.638	1.304	20.4	24#	0.00
37	Benzo(b)fluoranthene	1.339	1.541	-15.1	36#	0.00
38	Benzo(k)fluoranthene	1.376	1.651	-20.0	36#	0.00
39 C	Benzo(a)pyrene	1.341	1.340	0.1	31#	0.00
40	Dibenzo(a,h)anthracene	1.304	1.021	21.7	24#	0.00
41	Benzo(g,h,i)perylene	1.403	1.272	9.3	27#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4