

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082824\
 Data File : BN033639.D
 Acq On : 28 Aug 2024 11:08
 Operator : MA/JU
 Sample : PB162963BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK963

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 08/29/2024
 Supervised By :mohammad ahmed 08/30/2024

Quant Time: Aug 28 12:23:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN082724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 27 23:48:09 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.536	152	7157	0.400	ng/u1	0.00
4) Naphthalene-d8	10.299	136	20132	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.174	164	10589	0.400	ng/u1	0.00
13) Phenanthrene-d10	16.927	188	21728m	0.400	ng/u1	0.00
17) Chrysene-d12	21.138	240	16983m	0.400	ng/u1	0.00
23) Perylene-d12	23.301	264	18002m	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.148	96	48265	6.560	ng/u1	0.00
6) 2-Methylnaphthalene-d10	11.894	152	9276	0.321	ng/u1	0.00
18) Fluoranthene-d10	18.966	212	17859	0.353	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082824\
Data File : BN033639.D
Acq On : 28 Aug 2024 11:08
Operator : MA/JU
Sample : PB162963BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK963

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/29/2024
Supervised By :mohammad ahmed 08/30/2024

Quant Time: Aug 28 12:23:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN082724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 27 23:48:09 2024
Response via : Initial Calibration

