


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082918\
Data File : BN002537.D
Acq On : 29 Aug 2018 08:48
Operator : JU/SJ
Sample : DFTPP10
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP10

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SOM-EPA-BN081318MA.M
Title : SVOA CALIBRATION
Last Update : Sat Aug 25 04:34:18 2018



AutoFind: Scans 2466, 2467, 2468; Background Corrected with Scan 2461

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	47.3	122061	PASS
68	69	0.00	2	1.9	2249	PASS
69	198	0.00	100	45.3	116949	PASS
70	69	0.00	2	0.6	662	PASS
127	198	10	80	51.6	133421	PASS
197	198	0.00	2	0.2	489	PASS
198	198	100	100	100.0	258325	PASS
199	198	5	9	6.8	17449	PASS
275	198	10	60	26.2	67576	PASS
365	198	1	100	3.4	8742	PASS
441	198	0.01	100	13.2	34085	PASS
442	198	50	100	88.6	228842	PASS
443	442	15	24	19.2	43944	PASS

m/z	Abundance
37.00	504.0
38.00	1418.0
39.10	9448.0
39.90	1194.0
40.90	541.0
44.00	8705.0
45.00	319.0
49.00	872.0
50.00	29400.0
51.00	115712.0
52.00	5454.0
54.90	819.0
56.00	3283.0
57.00	7493.0
57.90	367.0
61.00	1564.0
62.00	1662.0
63.00	4460.0
63.90	598.0
65.00	2287.0
67.90	1919.0
69.00	109360.0
69.90	642.0
73.00	934.0
74.00	10499.0
75.00	17896.0
76.00	5731.0
77.00	116376.0
78.00	8708.0
79.00	8284.0
80.00	6548.0
81.00	9111.0
82.00	2360.0
83.00	1988.0
85.00	1668.0
86.00	2460.0
87.00	1183.0
87.90	359.0
90.90	2469.0
91.90	2325.0
93.00	14465.0
94.00	1011.0
96.10	762.0
98.00	10534.0
98.90	11487.0
99.90	645.0
101.00	4448.0
102.90	1324.0
104.00	3172.0
104.90	3042.0
105.90	1049.0
107.00	36232.0
108.00	5361.0
109.00	1210.0
110.00	58768.0
110.90	9817.0
112.00	991.0
115.90	2183.0
116.90	29608.0
117.90	2525.0
120.00	387.0
121.90	2548.0
122.90	3183.0
123.90	1689.0
125.00	1312.0
126.10	479.0
127.00	113760.0
128.00	8759.0
129.00	45632.0
130.00	3862.0
131.00	771.0
131.90	419.0
134.00	1218.0
135.00	3823.0
135.90	1620.0
136.90	2250.0
137.90	370.0
140.00	625.0
141.00	6030.0
141.90	1841.0
142.90	1494.0
143.90	334.0
145.00	329.0
145.90	964.0
147.00	3195.0
147.90	7450.0
149.00	1945.0
149.80	362.0
151.10	691.0
151.90	342.0
152.90	1874.0

Instrument :
BNA_N
ClientSampleId :
DFTPP10

153.90	1272.0
155.00	3022.0
156.00	5041.0
156.90	658.0
158.00	879.0
158.90	694.0
160.00	2107.0
161.00	3291.0
161.90	706.0
164.00	424.0
164.90	2322.0
166.00	1638.0
167.00	10730.0
168.00	8630.0
168.90	1368.0
169.90	344.0
170.80	522.0
171.90	1149.0
173.00	1511.0
173.90	2118.0
175.00	4982.0
176.00	1427.0
176.90	2025.0
177.90	882.0
178.90	9063.0
180.00	6379.0
181.00	2948.0
182.00	459.0
184.10	600.0
185.00	4343.0
186.00	31728.0
187.00	8158.0
188.00	943.0
188.90	2459.0
190.00	535.0
191.00	1273.0
192.00	2749.0
193.00	3377.0
194.00	540.0
194.90	444.0
196.00	5328.0
197.90	211840.0
198.90	14633.0
199.90	1158.0
201.20	916.0
201.50	925.0
202.90	1683.0
204.00	7195.0
205.00	13547.0
206.00	58464.0
207.00	7999.0
208.00	2300.0
209.00	820.0
210.00	593.0
210.90	2254.0
214.80	541.0
215.90	1367.0
216.90	15491.0
217.90	2446.0
221.00	6970.0
223.00	3135.0
224.00	29888.0
225.00	8054.0
226.00	885.0
227.00	14129.0
227.90	1900.0
229.00	2748.0
230.00	506.0
231.00	1275.0
233.90	905.0
234.90	1062.0
236.00	713.0
237.00	1372.0
238.90	533.0
239.90	554.0
240.90	876.0
242.00	1745.0
243.00	1442.0
244.00	23184.0
245.00	3542.0
246.00	4584.0
246.80	842.0
248.90	746.0
252.00	335.0
252.80	432.0
255.00	116968.0
256.00	18424.0
257.00	1474.0
258.00	7762.0
258.90	1173.0
264.90	2545.0
265.90	552.0
271.00	1398.0
273.00	3610.0
273.90	8140.0

Instrument :
BNA_N
ClientSampleId :
DFTPP10

275.00	48320.0
276.00	7194.0
277.00	4185.0
277.90	830.0
283.20	403.0
284.00	362.0
284.90	765.0
292.00	301.0
292.90	1038.0
296.00	14509.0
296.90	1915.0
298.00	357.0
302.90	1640.0
303.90	606.0
310.00	440.0
314.00	751.0
314.90	1854.0
316.00	739.0
321.00	408.0
323.00	4059.0
324.00	826.0
326.90	789.0
327.80	361.0
332.00	362.0
333.00	468.0
334.00	3063.0
334.90	626.0
340.90	435.0
345.90	951.0
352.00	1251.0
352.90	1157.0
354.00	1282.0
364.90	5855.0
366.00	893.0
370.90	328.0
372.00	2155.0
373.00	563.0
382.80	479.0
401.80	640.0
403.00	1259.0
403.90	523.0
421.00	867.0
422.00	1033.0
423.00	6448.0
424.00	1636.0
441.00	19528.0
442.00	131264.0
443.00	26080.0
444.00	2506.0

Instrument :
BNA_N
ClientSampleId :
DFTPP10

m/z	Abundance
37.00	647.0
38.00	2091.0
39.00	11533.0
40.00	1386.0
41.00	439.0
44.00	8454.0
45.00	429.0
48.90	774.0
50.00	38728.0
51.00	150336.0
52.00	7817.0
53.00	351.0
55.00	1025.0
56.00	4696.0
57.00	10155.0
58.00	369.0
61.00	2389.0
62.00	2372.0
63.00	6001.0
64.00	833.0
65.00	2602.0
66.00	372.0
68.00	2766.0
69.00	143488.0
70.00	702.0
73.00	1139.0
74.00	15082.0
75.00	23392.0
76.00	7187.0
77.00	157248.0
78.00	9982.0
79.00	10363.0
80.00	7523.0
81.00	11448.0
82.00	2929.0
82.90	2321.0
85.00	2149.0
85.90	3554.0
86.90	1224.0
88.00	376.0
88.90	325.0
91.00	2199.0
92.00	2807.0
93.00	19960.0
94.00	1415.0
96.10	890.0
96.90	460.0
97.90	14881.0
99.00	14799.0
99.90	1060.0
100.90	6410.0
103.00	1972.0
103.90	4072.0
104.90	3367.0
106.00	1131.0
107.00	47200.0
107.90	7820.0
109.00	1270.0
109.90	84352.0
111.00	13361.0
112.00	1532.0
113.00	589.0
116.00	2679.0
116.90	39208.0
118.00	2862.0
119.00	565.0
119.90	470.0
122.00	3077.0
123.00	4963.0
123.90	2517.0
125.00	2154.0
126.10	925.0
127.00	166144.0
128.00	12771.0
128.90	65288.0
130.00	5759.0
130.90	1116.0
132.00	720.0
133.90	1740.0
134.90	5241.0
135.90	2360.0
136.90	3316.0
137.90	513.0
138.90	313.0
139.90	1113.0
141.00	9529.0
142.00	2440.0
142.90	1869.0
143.90	460.0
145.00	331.0
145.90	1576.0

Instrument :
BNA_N
ClientSampleId :
DFTPP10

146.90	4020.0
147.90	10269.0
149.00	2603.0
150.00	632.0
151.00	1062.0
151.70	842.0
152.90	2975.0
153.90	2117.0
154.90	5034.0
156.00	6937.0
157.00	1418.0
157.90	1720.0
158.90	1325.0
159.90	2826.0
161.00	3714.0
161.90	1626.0
163.00	311.0
164.00	353.0
165.00	3687.0
166.00	2513.0
166.90	15482.0
168.00	12095.0
169.00	1638.0
170.00	582.0
171.00	636.0
171.90	1438.0
172.90	2417.0
173.90	3865.0
175.00	7001.0
176.00	1677.0
176.90	3081.0
177.90	1134.0
178.90	13636.0
180.00	8454.0
180.90	4296.0
182.00	857.0
182.80	309.0
184.00	1064.0
185.00	6714.0
186.00	46352.0
187.00	13413.0
188.00	1567.0
188.90	2802.0
189.90	686.0
191.00	1592.0
191.90	4227.0
193.00	4886.0
194.00	778.0
195.00	593.0
196.00	7964.0
196.90	1468.0
197.90	316608.0
198.90	21680.0
200.00	1830.0
201.50	1280.0
202.90	2769.0
204.00	13003.0
205.00	21320.0
206.00	88520.0
207.00	12511.0
207.90	3164.0
209.00	1320.0
210.20	1347.0
211.00	3873.0
211.60	636.0
212.90	321.0
214.90	907.0
215.90	2070.0
216.90	23896.0
218.00	3174.0
218.90	351.0
220.10	549.0
221.00	10423.0
221.80	3612.0
223.00	5404.0
224.00	50456.0
225.00	11647.0
226.00	1095.0
227.00	21392.0
227.90	2651.0
228.90	4381.0
230.00	774.0
231.00	1901.0
232.10	353.0
233.00	438.0
233.90	1380.0
234.90	1519.0
235.90	992.0
237.00	1532.0
238.90	886.0
239.90	587.0
240.90	1396.0
241.90	2420.0
243.00	2572.0
244.00	34776.0

Instrument :
BNA_N
ClientSampleId :
DFTPP10

245.00	4866.0
246.00	8049.0
246.90	1224.0
248.00	318.0
249.00	1307.0
251.00	372.0
251.90	494.0
252.90	763.0
254.10	1270.0
255.00	193280.0
256.00	28872.0
257.00	2305.0
257.90	10921.0
259.00	1594.0
260.00	372.0
264.00	387.0
264.90	4418.0
265.90	756.0
270.90	1400.0
272.00	694.0
273.00	5541.0
274.00	13959.0
275.00	82736.0
276.00	10894.0
277.00	7966.0
278.00	1291.0
283.00	842.0
284.00	497.0
285.00	1559.0
290.10	312.0
290.90	463.0
291.90	391.0
293.00	1691.0
294.00	434.0
295.00	645.0
295.90	24312.0
297.00	3246.0
300.90	327.0
302.00	524.0
303.00	2985.0
303.90	859.0
308.00	499.0
312.90	317.0
314.00	1027.0
314.90	2839.0
315.90	1400.0
321.00	891.0
321.80	438.0
323.00	7363.0
323.90	1195.0
326.90	1437.0
327.90	607.0
331.90	491.0
332.90	750.0
334.00	4836.0
335.00	1270.0
341.00	643.0
345.90	1356.0
346.90	319.0
352.00	2147.0
353.00	1554.0
354.00	2132.0
365.00	10376.0
365.90	1540.0
370.80	465.0
372.00	4325.0
373.00	1119.0
382.90	1105.0
389.90	412.0
391.00	353.0
400.80	371.0
402.00	1722.0
402.90	2054.0
404.00	1066.0
421.00	1817.0
422.00	2051.0
423.00	13095.0
424.00	3074.0
437.80	548.0
441.00	39952.0
442.00	264448.0
443.00	49736.0
444.00	5087.0

Instrument :
BNA_N
ClientSampleId :
DFTPP10

m/z	Abundance
37.00	483.0
38.10	1197.0
39.00	7627.0
40.10	1224.0
41.20	347.0
44.00	8633.0
49.00	613.0
50.00	25512.0
51.00	100136.0
52.00	5118.0
55.00	728.0
56.00	3351.0
57.00	6999.0
57.90	312.0
60.90	1620.0
62.00	1591.0
63.00	4100.0
64.10	681.0
65.00	1673.0
66.90	442.0
67.90	2062.0
69.00	98000.0
70.00	643.0
73.00	798.0
74.00	10202.0
75.00	15753.0
76.00	4689.0
77.00	104944.0
78.00	6773.0
79.00	7886.0
79.90	5910.0
81.00	8296.0
82.00	1796.0
83.00	1963.0
85.00	1227.0
86.00	2365.0
86.90	928.0
91.00	1853.0
91.90	2012.0
93.00	13342.0
93.90	810.0
95.90	524.0
98.00	9888.0
99.00	10095.0
99.90	889.0
100.90	4829.0
103.00	1565.0
103.90	3133.0
105.00	2569.0
106.00	1062.0
107.00	33720.0
107.90	5709.0
109.00	782.0
109.90	60640.0
111.00	9048.0
112.00	932.0
112.90	392.0
115.90	1724.0
116.90	29536.0
117.90	2467.0
119.00	312.0
119.80	350.0
121.90	2063.0
123.00	3813.0
123.90	1692.0
124.90	1759.0
126.10	450.0
127.00	120360.0
128.00	9289.0
129.00	49504.0
130.00	4670.0
130.90	842.0
132.00	410.0
133.90	1278.0
134.90	3689.0
135.90	1350.0
136.90	2128.0
137.70	604.0
140.00	568.0
140.90	6716.0
141.90	2112.0
142.90	1637.0
143.90	416.0
145.00	403.0
146.00	1130.0
146.90	2993.0
147.90	7107.0
149.00	1824.0
150.00	332.0
151.00	887.0
151.70	434.0

Instrument :
BNA_N
ClientSampleId :
DFTPP10

152.90	2185.0
153.90	1181.0
154.90	3260.0
156.00	5201.0
157.00	1550.0
157.90	1181.0
159.00	796.0
159.90	2253.0
160.90	3278.0
162.00	977.0
162.90	398.0
163.90	353.0
165.00	2242.0
166.00	2071.0
166.90	12478.0
167.90	8649.0
168.90	1535.0
170.10	412.0
170.90	549.0
171.90	1287.0
172.90	1538.0
174.00	2466.0
175.00	4845.0
176.00	1522.0
177.00	2527.0
177.90	746.0
178.90	10003.0
179.90	6712.0
181.00	3435.0
182.10	550.0
183.00	462.0
183.90	871.0
185.00	4543.0
186.00	38056.0
187.00	10277.0
188.00	1101.0
188.90	2577.0
189.90	454.0
191.00	1212.0
191.90	2980.0
193.00	3718.0
194.00	912.0
194.90	575.0
195.90	6820.0
197.90	246528.0
198.90	16034.0
200.00	1435.0
201.40	1191.0
202.90	2185.0
204.00	11197.0
205.00	17856.0
206.00	70144.0
207.00	9960.0
207.90	2427.0
209.00	860.0
209.90	874.0
211.00	3095.0
211.90	454.0
212.90	332.0
214.90	1131.0
216.00	1962.0
216.90	19224.0
217.90	2511.0
219.00	308.0
221.00	7793.0
221.80	2847.0
223.00	4466.0
224.00	39128.0
225.00	9386.0
226.00	934.0
227.00	17432.0
228.00	2284.0
228.90	3249.0
229.90	696.0
231.00	1522.0
231.90	304.0
233.00	481.0
233.90	932.0
234.90	1132.0
236.00	926.0
236.90	1382.0
238.90	653.0
239.90	608.0
241.00	889.0
242.00	2094.0
242.90	2030.0
244.00	31208.0
245.00	3513.0
245.90	6024.0
246.90	1499.0
248.90	1010.0
250.90	310.0
251.90	390.0
252.90	916.0
254.00	1370.0

Instrument :
BNA_N
ClientSampleId :
DFTPP10

255.00	170880.0
256.00	23448.0
256.90	1614.0
257.90	10282.0
258.90	1597.0
264.90	3969.0
265.90	767.0
271.00	1788.0
272.00	598.0
273.00	5048.0
274.00	11701.0
275.00	71672.0
276.00	8857.0
276.90	6548.0
278.00	1027.0
282.90	746.0
283.90	527.0
285.00	1266.0
291.10	365.0
292.00	375.0
292.90	1623.0
294.00	546.0
295.90	21960.0
296.90	2903.0
300.90	429.0
301.90	316.0
303.00	2631.0
303.90	711.0
312.90	368.0
314.00	1321.0
314.90	2780.0
315.90	1089.0
320.90	589.0
322.00	398.0
323.00	6829.0
324.00	1389.0
326.90	1359.0
328.00	665.0
332.00	668.0
332.90	469.0
334.00	4462.0
335.00	1048.0
341.00	726.0
345.90	1677.0
352.00	2238.0
352.90	1585.0
354.00	2038.0
355.00	381.0
365.00	9997.0
365.90	1312.0
370.90	573.0
372.00	3730.0
373.00	1142.0
382.90	1274.0
383.90	331.0
390.00	528.0
392.00	333.0
401.90	1547.0
402.80	1788.0
404.10	715.0
420.90	1966.0
422.00	2248.0
423.00	13864.0
424.00	3209.0
438.00	381.0
441.00	42776.0
442.00	290816.0
443.00	56016.0
443.90	5705.0

Instrument :
BNA_N
ClientSampleId :
DFTPP10

m/z	Abundance
39.90	971.0
44.00	8040.0
207.00	434.0

Instrument :
BNA_N
ClientSampleId :
DFTPP10