

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN083019\
 Data File : BN007524.D
 Acq On : 30 Aug 2019 15:26
 Operator : HP/JU
 Sample : K4553-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-TT-RE139D1-GW-20190826

Manual Integrations
 APPROVED

mohammad
 8/31/2019 3:50:24 PM

Quant Time: Aug 31 02:15:51 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 29 12:18:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.39	152	6131	0.40	ng	-0.02
7) Naphthalene-d8	10.15	136	25697	0.40	ng	-0.01
13) Acenaphthene-d10	14.04	164	16555	0.40	ng	-0.02
19) Phenanthrene-d10	16.79	188	39910	0.40	ng	0.00
27) Chrysene-d12	21.00	240	39208	0.40	ng	-0.01
34) Perylene-d12	23.11	264	38048	0.40	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.03	112	3044	0.12	ng	0.00
5) Phenol-d6	6.59	99	2361	0.06	ng	0.00
8) Nitrobenzene-d5	8.53	82	6986	0.31	ng	0.00
11) 2-Methylnaphthalene-d10	11.75	152	16401	0.38	ng	-0.01
14) 2,4,6-Tribromophenol	15.54	330	3544	0.41	ng	-0.02
15) 2-Fluorobiphenyl	12.65	172	35139	0.53	ng	0.00
25) Fluoranthene-d10	18.84	212	48758	0.38	ng	-0.01
29) Terphenyl-d14	19.45	244	43857	0.43	ng	-0.02
Target Compounds						
2) 1,4-Dioxane	2.99	88	8951	0.879	ng	Qvalue # 87
22) Pentachlorophenol	16.46	266	341m	0.041	ng	
32) Bis(2-ethylhexyl)phthalate	20.96	149	7545	0.057	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN083019\
Data File : BN007524.D
Acq On : 30 Aug 2019 15:26
Operator : HP/JU
Sample : K4553-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-TT-RE139D1-GW-20190826

Manual Integrations
APPROVED

mohammad
8/31/2019 3:50:24 PM

Quant Time: Aug 31 02:15:51 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 12:18:45 2019
Response via : Initial Calibration

