

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN083019\
 Data File : BN007520.D
 Acq On : 30 Aug 2019 13:03
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 31 01:58:32 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 29 12:18:45 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	184#	-0.02
2	1,4-Dioxane	0.665	0.728	-9.5	194#	0.00
3	n-Nitrosodimethylamine	0.309	0.253	18.1	145	0.00
4 S	2-Fluorophenol	1.490	1.260	15.4	164#	-0.02
5 S	Phenol-d6	1.953	1.587	18.7	171#	0.00
6	bis(2-Chloroethyl)ether	1.445	1.519	-5.1	195#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	202#	-0.03
8 S	Nitrobenzene-d5	0.348	0.331	4.9	201#	-0.01
9	Naphthalene	1.143	1.122	1.8	193#	-0.01
10	Hexachlorobutadiene	0.234	0.210	10.3	174#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.665	0.683	-2.7	204#	-0.01
12	2-Methylnaphthalene	0.779	0.816	-4.7	205#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	234#	-0.02
14 S	2,4,6-Tribromophenol	0.209	0.240	-14.8	274#	-0.02
15 S	2-Fluorobiphenyl	1.613	1.558	3.4	221#	0.00
16	Acenaphthylene	2.367	2.199	7.1	215#	-0.02
17	Acenaphthene	1.385	1.345	2.9	223#	0.00
18	Fluorene	1.751	1.774	-1.3	232#	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	244#	0.00
20	4-Bromophenyl-phenylether	0.158	0.044	72.2#	79	0.00
21	Hexachlorobenzene	0.268	0.262	2.2	236#	-0.02
22	Pentachlorophenol	0.083	0.126	-51.8#	394#	-0.02
23	Phenanthrene	1.203	1.234	-2.6	250#	-0.02
24	Anthracene	1.210	1.194	1.3	239#	0.00
25 SURR	Fluoranthene-d10	1.288	1.244	3.4	228#	-0.01
26	Fluoranthene	1.545	1.519	1.7	234#	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	217#	0.00
28	Pvrene	1.609	1.762	-9.5	232#	0.00
29 S	Terphenyl-d14	1.047	1.133	-8.2	227#	-0.01
30	Benzo(a)anthracene	1.571	1.578	-0.4	214#	-0.01
31	Chrysene	1.516	1.498	1.2	205#	-0.01
32	Bis(2-ethylhexyl)phthalate	1.177	1.338	-13.7	261#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.827	1.436	21.4	165#	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	185#	-0.01
35	Benzo(b)fluoranthene	1.449	1.516	-4.6	189#	0.00
36	Benzo(k)fluoranthene	1.432	1.543	-7.8	192#	0.00
37 C	Benzo(a)pyrene	1.379	1.396	-1.2	182#	-0.01
38	Dibenzo(a,h)anthracene	1.376	1.219	11.4	159#	-0.02
39	Benzo(a,h,i)perylene	1.371	1.243	9.3	163#	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN083019\
Data File : BN007520.D
Acq On : 30 Aug 2019 13:03
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 31 01:58:32 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 12:18:45 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			