

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083023\
 Data File : BN027283.D
 Acq On : 30 Aug 2023 21:07
 Operator : MA/JU
 Sample : PB155057BS
 Misc :
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB155057BS

Quant Time: Aug 31 04:29:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 29 05:37:30 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.066	152	4355	0.400	ng	0.00
7) Naphthalene-d8	10.895	136	12511	0.400	ng	0.00
13) Acenaphthene-d10	14.694	164	8240	0.400	ng	0.00
19) Phenanthrene-d10	17.542	188	5520	0.400	ng	0.10
29) Chrysene-d12	21.614	240	35370	0.400	ng	0.00
35) Perylene-d12	24.145	264	28836	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.589	112	4038	0.356	ng	0.00
5) Phenol-d6	7.207	99	4936	0.358	ng	0.00
8) Nitrobenzene-d5	9.251	82	3307	0.363	ng	-0.01
11) 2-Methylnaphthalene-d10	12.464	152	6600	0.359	ng	0.00
14) 2,4,6-Tribromophenol	16.189	330	1444	0.473	ng	0.00
15) 2-Fluorobiphenyl	13.325	172	10502	0.335	ng	0.00
27) Fluoranthene-d10	0.000	212	0	0.000	ng	
31) Terphenyl-d14	20.048	244	20147	0.284	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.458	88	1958	0.368	ng	# 87
3) n-Nitrosodimethylamine	3.805	42	2351	0.417	ng	90
6) bis(2-Chloroethyl)ether	7.495	93	5057	0.379	ng	97
9) Naphthalene	10.938	128	12793	0.375	ng	99
10) Hexachlorobutadiene	11.173	225	2400	0.377	ng	# 99
12) 2-Methylnaphthalene	12.540	142	8790	0.362	ng	99
16) Acenaphthylene	14.427	152	13204	0.352	ng	100
17) Acenaphthene	14.758	154	9846	0.393	ng	97
18) Fluorene	15.730	166	15146	0.454	ng	99
20) 4,6-Dinitro-2-methylph...	16.983	198	2	0.025	ng	# 35
21) 4-Bromophenyl-phenylether	16.710	248	198	0.068	ng	# 14
23) Atrazine	17.070	200	696	0.319	ng	# 56
25) Phenanthrene	17.579	178	27133	1.672	ng	99
28) Fluoranthene	19.495	202	51332	2.688	ng	99
30) Pyrene	19.857	202	68588	0.488	ng	100
32) Benzo(a)anthracene	21.605	228	41098	0.339	ng	99
33) Chrysene	21.659	228	44862	0.337	ng	99
34) Bis(2-ethylhexyl)phtha...	21.479	149	28036	0.479	ng	# 96
36) Indeno(1,2,3-cd)pyrene	26.823	276	35041	0.299	ng	99
37) Benzo(b)fluoranthene	23.358	252	40895	0.366	ng	96
38) Benzo(k)fluoranthene	23.411	252	37242	0.325	ng	96
39) Benzo(a)pyrene	24.031	252	33640	0.322	ng	94
40) Dibenzo(a,h)anthracene	26.843	278	27127	0.296	ng	98
41) Benzo(g,h,i)perylene	27.650	276	32542	0.306	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083023\
Data File : BN027283.D
Acq On : 30 Aug 2023 21:07
Operator : MA/JU
Sample : PB155057BS
Misc :
ALS Vial : 62 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB155057BS

Quant Time: Aug 31 04:29:57 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 29 05:37:30 2023
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 8.066 min Scan# 71
 Delta R.T. -0.007 min
 Lab File: BN027283.D
 Acq: 30 Aug 2023 21:07




Instrument :
 BNA_N
 ClientSampleId :
 PB155057BS


Tgt Ion:152 Resp: 4355
 Ion Ratio Lower Upper
 152 100
 150 155.8 122.7 184.1
 115 65.4 48.6 73.0



#2
 1,4-Dioxane
 Concen: 0.368 ng
 RT: 3.458 min Scan# 80
 Delta R.T. -0.007 min
 Lab File: BN027283.D
 Acq: 30 Aug 2023 21:07

Tgt Ion: 88 Resp: 1958
 Ion Ratio Lower Upper
 88 100
 43 31.2 19.1 28.7#
 58 74.6 52.2 78.2









#3
n-Nitrosodimethylamine
Concen: 0.417 ng
RT: 3.805 min Scan# 11
Delta R.T. -0.007 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07




Instrument :
BNA_N
ClientSampleId :
PB155057BS


Tgt Ion: 42 Resp: 2351
Ion Ratio Lower Upper
42 100
74 129.3 113.8 170.8
44 8.6 7.3 10.9



#4
2-Fluorophenol
Concen: 0.356 ng
RT: 5.589 min Scan# 375
Delta R.T. -0.000 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07

Tgt Ion: 112 Resp: 4038
Ion Ratio Lower Upper
112 100
64 56.5 44.4 66.6
63 30.6 23.1 34.7









#5
Phenol-d6
Concen: 0.358 ng
RT: 7.207 min Scan# 599
Delta R.T. -0.000 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07




Instrument :
BNA_N
ClientSampleId :
PB155057BS


Tgt Ion:	99	Resp:	4936
Ion	Ratio	Lower	Upper
99	100		
42	19.5	14.5	21.7
71	33.8	25.8	38.6



#6
bis(2-Chloroethyl)ether
Concen: 0.379 ng
RT: 7.495 min Scan# 639
Delta R.T. -0.007 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07

Tgt Ion:	93	Resp:	5057
Ion	Ratio	Lower	Upper
93	100		
63	76.8	58.5	87.7
95	33.0	26.1	39.1









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.895 min Scan# 994
Delta R.T. -0.000 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07




Instrument :
BNA_N
ClientSampleId :
PB155057BS


Tgt Ion:	136	Resp:	12511
Ion	Ratio	Lower	Upper
136	100		
137	11.8	9.3	13.9
54	6.9	6.7	10.1
68	6.2	6.1	9.1



#8
Nitrobenzene-d5
Concen: 0.363 ng
RT: 9.251 min Scan# 840
Delta R.T. -0.011 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07

Tgt Ion:	82	Resp:	3307
Ion	Ratio	Lower	Upper
82	100		
128	36.7	33.4	50.0
54	53.8	37.7	56.5









#9
Naphthalene
Concen: 0.375 ng
RT: 10.938 min Scan# 999
Delta R.T. -0.011 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07




Instrument :
BNA_N
ClientSampleId :
PB155057BS


Tgt Ion:128 Resp: 12793
Ion Ratio Lower Upper
128 100
129 11.9 9.4 14.0
127 14.0 10.6 15.8



#10
Hexachlorobutadiene
Concen: 0.377 ng
RT: 11.173 min Scan# 1020
Delta R.T. -0.011 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07

Tgt Ion:225 Resp: 2400
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 50.6 75.8









#11
2-Methylnaphthalene-d10
Concen: 0.359 ng
RT: 12.464 min Scan# 11
Delta R.T. -0.008 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07




Instrument :
BNA_N
ClientSampleId :
PB155057BS


Tgt Ion:152 Resp: 6600
Ion Ratio Lower Upper
152 100
151 20.4 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.362 ng
RT: 12.540 min Scan# 1325
Delta R.T. -0.004 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07

Tgt Ion:142 Resp: 8790
Ion Ratio Lower Upper
142 100
141 87.8 71.1 106.7
115 38.0 29.6 44.4





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN027283.D

Acq: 30 Aug 2023 21:07

Instrument :

BNA_N

ClientSampleId :

PB155057BS





Tgt Ion:164 Resp: 8240

Ion Ratio Lower Upper

164 100

162 103.1 85.8 128.8

160 53.2 44.2 66.2



#14

2,4,6-Tribromophenol

Concen: 0.473 ng

RT: 16.189 min Scan# 1737

Delta R.T. -0.000 min

Lab File: BN027283.D

Acq: 30 Aug 2023 21:07




Tgt Ion:330 Resp: 1444


Ion Ratio Lower Upper

330 100

332 96.1 75.2 112.8

141 52.0 36.2 54.4









#15
2-Fluorobiphenyl
Concen: 0.335 ng
RT: 13.325 min Scan# 1478
Delta R.T. -0.000 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07




Instrument :
BNA_N
ClientSampleId :
PB155057BS


Tgt Ion:172 Resp: 10502
Ion Ratio Lower Upper
172 100
171 34.5 27.8 41.8
170 23.4 19.3 28.9



#16
Acenaphthylene
Concen: 0.352 ng
RT: 14.427 min Scan# 1581
Delta R.T. -0.000 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07

Tgt Ion:152 Resp: 13204
Ion Ratio Lower Upper
152 100
151 19.6 15.5 23.3
153 13.1 10.3 15.5









#17
Acenaphthene
Concen: 0.393 ng
RT: 14.758 min Scan# 1612
Delta R.T. -0.000 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07




Instrument :
BNA_N
ClientSampleId :
PB155057BS


Tgt Ion:154 Resp: 9846
Ion Ratio Lower Upper
154 100
153 119.0 93.1 139.7
152 63.9 48.6 73.0



#18
Fluorene
Concen: 0.454 ng
RT: 15.730 min Scan# 1700
Delta R.T. -0.012 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07

Tgt Ion:166 Resp: 15146
Ion Ratio Lower Upper
166 100
165 100.3 79.0 118.4
167 13.5 10.8 16.2





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.542 min Scan# 1801

Delta R.T. 0.099 min

Lab File: BN027283.D

Acq: 30 Aug 2023 21:07

Instrument :

BNA_N

ClientSampleId :

PB155057BS





Tgt Ion:188 Resp: 5520

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.8 11.0 16.6



#20

4,6-Dinitro-2-methylphenol

Concen: 0.025 ng

RT: 16.983 min Scan# 1801

Delta R.T. 0.087 min

Lab File: BN027283.D

Acq: 30 Aug 2023 21:07




Tgt Ion:198 Resp: 2


Ion Ratio Lower Upper

198 100

51 554.7 282.2 423.2#

105 239.6 168.7 253.1









#21
4-Bromophenyl-phenylether
Concen: 0.068 ng
RT: 16.710 min Scan# 11
Delta R.T. 0.074 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07




Instrument :
BNA_N
ClientSampleId :
PB155057BS


Tgt Ion:248 Resp: 198
Ion Ratio Lower Upper
248 100
250 142.9 79.7 119.5#
141 166.8 39.8 59.8#



#23
Atrazine
Concen: 0.319 ng
RT: 17.070 min Scan# 1808
Delta R.T. 0.174 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07

Tgt Ion:200 Resp: 696
Ion Ratio Lower Upper
200 100
173 12.7 22.5 33.7#
215 12.7 39.0 58.6#





#25

Phenanthrene

Concen: 1.672 ng

RT: 17.579 min Scan# 11

Delta R.T. 0.087 min

Lab File: BN027283.D

Acq: 30 Aug 2023 21:07

Instrument :

BNA_N

ClientSampleId :

PB155057BS





Tgt Ion:178 Resp: 27133

Ion Ratio Lower Upper

178 100

176 18.7 15.6 23.4

179 15.7 12.3 18.5



#27

Fluoranthene-d10

Concen: 0.000 ng

Expected RT: 19.47 min

Lab File: BN027283.D

Acq: 30 Aug 2023 21:07


Tgt Ion: 212

Sig Exp Ratio

212 100


106 15.1

104 8.6



Ion 106.00 (105.70 to 106.70): BN027283.D\data.ms





Ion 104.00 (103.70 to 104.70): BN027283.D\data.ms



#28
Fluoranthene
Concen: 2.688 ng
RT: 19.495 min Scan# 2137
Delta R.T. -0.005 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07




Instrument :
BNA_N
ClientSampleId :
PB155057BS


Tgt Ion:202 Resp: 51332
Ion Ratio Lower Upper
202 100
101 12.6 9.6 14.4
203 17.1 13.6 20.4



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.614 min Scan# 2529
Delta R.T. -0.009 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07

Tgt Ion:240 Resp: 35370
Ion Ratio Lower Upper
240 100
120 18.5 14.2 21.4
236 28.8 22.9 34.3









#30
Pyrene
Concen: 0.488 ng
RT: 19.857 min Scan# 211
Delta R.T. -0.005 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07




Instrument :
BNA_N
ClientSampleId :
PB155057BS


Tgt Ion:202 Resp: 68588
Ion Ratio Lower Upper
202 100
200 21.0 16.6 25.0
203 18.1 14.2 21.4



#31
Terphenyl-d14
Concen: 0.284 ng
RT: 20.048 min Scan# 2255
Delta R.T. -0.005 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07

Tgt Ion:244 Resp: 20147
Ion Ratio Lower Upper
244 100
212 9.5 7.5 11.3
122 18.3 14.0 21.0





#32

Benzo(a)anthracene

Concen: 0.339 ng

RT: 21.605 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN027283.D

Acq: 30 Aug 2023 21:07

Instrument :

BNA_N

ClientSampleId :

PB155057BS





Tgt Ion:228 Resp: 41098

Ion Ratio Lower Upper

228 100

226 27.0 21.9 32.9

229 20.1 15.7 23.5



#33

Chrysene

Concen: 0.337 ng

RT: 21.659 min Scan# 2534

Delta R.T. -0.000 min

Lab File: BN027283.D

Acq: 30 Aug 2023 21:07




Tgt Ion:228 Resp: 44862


Ion Ratio Lower Upper

228 100

226 30.1 24.3 36.5

229 19.8 15.4 23.2









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.479 ng
 RT: 21.479 min Scan# 2149
 Delta R.T. -0.000 min
 Lab File: BN027283.D
 Acq: 30 Aug 2023 21:07




Instrument :
 BNA_N
 ClientSampleId :
 PB155057BS


Tgt Ion:149 Resp: 28036
 Ion Ratio Lower Upper
 149 100
 167 25.7 22.1 33.1
 279 2.7 2.7 4.1#



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 24.145 min Scan# 3211
 Delta R.T. -0.012 min
 Lab File: BN027283.D
 Acq: 30 Aug 2023 21:07

Tgt Ion:264 Resp: 28836
 Ion Ratio Lower Upper
 264 100
 260 26.5 20.7 31.1
 265 40.4 53.3 79.9#









#36
Indeno(1,2,3-cd)pyrene
Concen: 0.299 ng
RT: 26.823 min Scan# 4130
Delta R.T. -0.009 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07




Instrument :
BNA_N
ClientSampleId :
PB155057BS


Tgt Ion:276 Resp: 35041
Ion Ratio Lower Upper
276 100
138 30.6 23.4 35.2
277 24.7 19.8 29.8



#37
Benzo(b)fluoranthene
Concen: 0.366 ng
RT: 23.358 min Scan# 2942
Delta R.T. -0.009 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07

Tgt Ion:252 Resp: 40895
Ion Ratio Lower Upper
252 100
253 23.3 19.9 29.9
125 14.5 13.8 20.6





#38

Benzo(k)fluoranthene

Concen: 0.325 ng

RT: 23.411 min Scan# 2960

Delta R.T. -0.009 min

Lab File: BN027283.D


Acq: 30 Aug 2023 21:07

Instrument :

BNA_N

ClientSampleId :

PB155057BS





Tgt Ion:252 Resp: 37242

Ion Ratio Lower Upper

252 100

253 23.8 20.2 30.2

125 15.2 14.1 21.1



#39

Benzo(a)pyrene

Concen: 0.322 ng

RT: 24.031 min Scan# 3172

Delta R.T. -0.012 min

Lab File: BN027283.D

Acq: 30 Aug 2023 21:07




Tgt Ion:252 Resp: 33640


Ion Ratio Lower Upper

252 100

253 24.0 21.4 32.0

125 17.2 16.5 24.7









#40
Dibenzo(a,h)anthracene
Concen: 0.296 ng
RT: 26.843 min Scan# 4139
Delta R.T. -0.015 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07

Instrument :
BNA_N
ClientSampleId :
PB155057BS

Tgt Ion:278 Resp: 27127
Ion Ratio Lower Upper
278 100
139 21.6 17.9 26.9
279 25.7 21.8 32.6



#41
Benzo(g,h,i)perylene
Concen: 0.306 ng
RT: 27.650 min Scan# 4410
Delta R.T. -0.012 min
Lab File: BN027283.D
Acq: 30 Aug 2023 21:07

Tgt Ion:276 Resp: 32542
Ion Ratio Lower Upper
276 100
277 24.8 19.6 29.4
138 27.8 22.1 33.1

