

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083023\
 Data File : BN027305.D
 Acq On : 31 Aug 2023 12:01
 Operator : MA/JU
 Sample : 04126-21DL 5X
 Misc :
 ALS Vial : 85 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE132D4-20230821DL

Quant Time: Sep 01 03:47:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 29 05:37:30 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/01/2023
 Supervised By :mohammad ahmed 09/01/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.066	152	3802	0.400	ng	0.00
7) Naphthalene-d8	10.885	136	9405	0.400	ng	#-0.01
13) Acenaphthene-d10	14.694	164	5878	0.400	ng	0.00
19) Phenanthrene-d10	17.443	188	13867	0.400	ng	0.00
29) Chrysene-d12	21.623	240	9716	0.400	ng	0.00
35) Perylene-d12	24.156	264	7838	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.582	112	410	0.041	ng	0.00
5) Phenol-d6	7.207	99	260	0.022	ng	0.00
8) Nitrobenzene-d5	9.251	82	659	0.096	ng	-0.01
11) 2-Methylnaphthalene-d10	12.472	152	1206	0.087	ng	0.00
14) 2,4,6-Tribromophenol	16.202	330	202	0.093	ng	0.01
15) 2-Fluorobiphenyl	13.336	172	2013m	0.090	ng	0.01
27) Fluoranthene-d10	19.467	212	4051	0.118	ng	0.00
31) Terphenyl-d14	20.053	244	2793	0.143	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.458	88	6296	1.356	ng	# 92
34) Bis(2-ethylhexyl)phtha...	21.471	149	1483	0.092	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083023\
Data File : BN027305.D
Acq On : 31 Aug 2023 12:01
Operator : MA/JU
Sample : 04126-21DL 5X
Misc :
ALS Vial : 85 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE132D4-20230821DL

Quant Time: Sep 01 03:47:22 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 29 05:37:30 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/01/2023
Supervised By :mohammad ahmed 09/01/2023

