

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083023\
 Data File : BN027312.D
 Acq On : 31 Aug 2023 16:21
 Operator : MA/JU
 Sample : PB155217BL
 Misc :
 ALS Vial : 92 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB155217BL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/01/2023
 Supervised By :mohammad ahmed 09/01/2023

Quant Time: Aug 31 18:13:34 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 29 05:37:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.066	152	4453	0.400	ng	0.00
7) Naphthalene-d8	10.885	136	11642	0.400	ng	#-0.01
13) Acenaphthene-d10	14.694	164	7136	0.400	ng	0.00
19) Phenanthrene-d10	17.455	188	16860	0.400	ng	0.01
29) Chrysene-d12	21.632	240	12539	0.400	ng	0.00
35) Perylene-d12	24.162	264	10732	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.582	112	6172	0.532	ng	0.00
5) Phenol-d6	7.199	99	6928	0.491	ng	0.00
8) Nitrobenzene-d5	9.251	82	4190	0.494	ng	-0.01
11) 2-Methylnaphthalene-d10	12.464	152	8244	0.482	ng	0.00
14) 2,4,6-Tribromophenol	16.202	330	1019	0.386	ng	0.01
15) 2-Fluorobiphenyl	13.326	172	13773m	0.508	ng	0.00
27) Fluoranthene-d10	19.472	212	21459	0.512	ng	0.00
31) Terphenyl-d14	20.053	244	14980	0.595	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083023\
Data File : BN027312.D
Acq On : 31 Aug 2023 16:21
Operator : MA/JU
Sample : PB155217BL
Misc :
ALS Vial : 92 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB155217BL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/01/2023
Supervised By :mohammad ahmed 09/01/2023

Quant Time: Aug 31 18:13:34 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN082923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 29 05:37:30 2023
Response via : Initial Calibration

